



PERCENT SOLID

Supervisor: mohammad

Analyst: jignesh

Date: 5/11/2022

OVENTEMP IN Celsius(°C): 107

Time IN: 17:00

In Date: 05/10/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103

Time OUT: 08:25

Out Date: 05/11/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

BalanceID: M SC-4

Thermometer ID: %SOLIDS-OVEN

QC:LB120077

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
N2762-01	SB02	1	0.91	8.84	9.75	8.89	90.3	
N2762-02	SB03	2	0.94	9.02	9.96	8.75	86.6	
N2762-03	SB04	3	0.94	9.02	9.96	9.44	94.2	
N2762-04	SB05	4	0.92	9.03	9.95	8.59	84.9	
N2762-05	SB15	5	0.92	8.89	9.81	9.03	91.2	
N2762-06	SB16	6	0.91	8.69	9.6	8.44	86.7	
N2762-07	SB17	7	0.89	9.09	9.98	9.26	92.1	
N2762-08	SB18	8	0.95	8.99	9.94	8.06	79.1	
N2762-09	SB19	9	0.91	8.63	9.54	8.92	92.8	
N2762-09D	SB19DUP	10	0.93	8.63	9.56	8.95	92.9	
N2764-01	NYE-CON-0510	22	1.00	1.00	2.00	2.00	100.0	STONE SAMPLE,100%SOLIDS
N2769-01	CL-02-051022	11	0.91	8.90	9.81	8.87	89.4	
N2769-02	CL-02-051022-EPH-2	12	0.91	8.90	9.81	8.87	89.4	
N2771-01	CLI-2022-58	13	1.00	1.00	2.00	2.00	100.0	PILC REEL
N2771-02	CLI-2022-BC226656-1	14	1.00	1.00	2.00	2.00	100.0	PILC REEL
N2771-03	CLI-2022-BC226656-2	15	1.00	1.00	2.00	2.00	100.0	PILC REEL
N2771-04	CLI-2022-BC226656-3	16	1.00	1.00	2.00	2.00	100.0	PILC REEL
N2771-05	CLI-2022-BC226656-4	17	1.00	1.00	2.00	2.00	100.0	PILC REEL
N2771-06	CLI-2022-BC226656-5	18	1.00	1.00	2.00	2.00	100.0	PILC REEL
N2771-07	CLI-2022-BC157517-1	19	1.00	1.00	2.00	2.00	100.0	PILC REEL
N2771-08	CLI-2022-BC157517-2	20	1.00	1.00	2.00	2.00	100.0	PILC REEL
N2777-01	P-PARK-1-A	21	1.00	1.00	2.00	2.00	100.0	STONE SAMPLE,100% SOLIDS

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$