

LB120078


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

4107E



A= 0.0000e+000

B= 2.4439e-004 *510/x*C= -4.7066e-002 *y-intercept*

Rho= 0.9999715

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.008	-0.008	160	0.000	160					
0.05	0.050					185	0				-10
0.20	0.200	0.180	-0.020	930	0.0 %	930					0
2.50	2.500	2.509	0.009	10460	0.0 %	10460					0
5.00	5.000	5.056	0.056	20883	0.0 %	20883					0
7.50	7.500	7.471	-0.029	30762	0.0 %	30762					0
10.0	10.000	9.991	-0.009	41076	0.0 %	41076					

LB120078
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0	160	-		0 SFAM01.1	PPB	5/10/2022 12:29	Std
	0.2	930	-		0.2 SFAM01.1	PPB	5/10/2022 12:32	Std
	2.5	10460	-		2.5 SFAM01.1	PPB	5/10/2022 12:34	Std
	5	20883	-		5 SFAM01.1	PPB	5/10/2022 12:36	Std
	7.5	30762	-		7.5 SFAM01.1	PPB	5/10/2022 12:38	Std
	10	41076	-		10 SFAM01.1	PPB	5/10/2022 12:41	Std
ICV016	ICV016	16181	3.9073 -		SFAM01.1	PPB	5/10/2022 12:44	SMPL
ICB016	ICB016	81	-0.0273 -		SFAM01.1	PPB	5/10/2022 12:46	SMPL
CCV015	CCV015	21108	5.1114 -		SFAM01.1	PPB	5/10/2022 12:48	SMPL
CCB015	CCB015	90	-0.0251 -		SFAM01.1	PPB	5/10/2022 12:51	SMPL
PB144713BL	PBW713	162	-0.0075 -		SFAM01.1	PPB	5/10/2022 12:53	SMPL
N2613-01	BH096	141	-0.0126 -		SFAM01.1	PPB	5/10/2022 12:55	SMPL
N2613-02	BH097	234	0.0101 -		SFAM01.1	PPB	5/10/2022 12:57	SMPL
N2613-03	BH098	255	0.0153 -		SFAM01.1	PPB	5/10/2022 13:00	SMPL
N2613-04	BH099	489	0.0724 -		SFAM01.1	PPB	5/10/2022 13:02	SMPL
N2613-05	BH0A0	1152	0.2345 -		SFAM01.1	PPB	5/10/2022 13:04	SMPL
N2613-06	BH0A1	736	0.1328 -		SFAM01.1	PPB	5/10/2022 13:06	SMPL
N2613-07	BH0A5	843	0.159 -		SFAM01.1	PPB	5/10/2022 13:09	SMPL
N2613-08	BH0B2	191	-0.0004 -		SFAM01.1	PPB	5/10/2022 13:11	SMPL
N2613-09	BH0B3	186	-0.0016 -		SFAM01.1	PPB	5/10/2022 13:13	SMPL
N2613-10	BH0B4	516	0.079 -		SFAM01.1	PPB	5/10/2022 13:15	SMPL
N2613-11	BH0B5	542	0.0854 -		SFAM01.1	PPB	5/10/2022 13:18	SMPL
N2613-12	BH0B6	398	0.0502 -		SFAM01.1	PPB	5/10/2022 13:20	SMPL
N2613-13	BH0C3	269	0.0187 -		SFAM01.1	PPB	5/10/2022 13:22	SMPL
N2613-14	BH0C4	370	0.0434 -		SFAM01.1	PPB	5/10/2022 13:24	SMPL
N2613-15	BH0C5	500	0.0751 -		SFAM01.1	PPB	5/10/2022 13:27	SMPL
N2613-16	BH0C6	493	0.0734 -		SFAM01.1	PPB	5/10/2022 13:29	SMPL
N2613-17	BH0C7	358	0.0404 -		SFAM01.1	PPB	5/10/2022 13:31	SMPL
N2613-18	BH0D1	432	0.0585 -		SFAM01.1	PPB	5/10/2022 13:33	SMPL
N2613-19	BH0D1D	431	0.0583 -		SFAM01.1	PPB	5/10/2022 13:36	SMPL
CCV016	CCV016	20714	5.0152 -		SFAM01.1	PPB	5/10/2022 13:40	SMPL
CCB016	CCB016	103	-0.0219 -		SFAM01.1	PPB	5/10/2022 13:42	SMPL
N2613-20	BH0D1S	4346	1.015 -		SFAM01.1	PPB	5/10/2022 13:44	SMPL
N2613-22	BH0D3	325	0.0324 -		SFAM01.1	PPB	5/10/2022 13:46	SMPL
PB144714BL	PBW714	143	-0.0121 -		SFAM01.1	PPB	5/10/2022 13:49	SMPL
N2614-01	BH0D5	281	0.0216 -		SFAM01.1	PPB	5/10/2022 13:51	SMPL
N2614-02	BH0D6	373	0.0441 -		SFAM01.1	PPB	5/10/2022 13:53	SMPL
N2614-03	BH0D2	445	0.0617 -		SFAM01.1	PPB	5/10/2022 13:55	SMPL
N2614-04	BH0D2D	413	0.0539 -		SFAM01.1	PPB	5/10/2022 13:58	SMPL
N2614-05	BH0D2S	4350	1.016 -		SFAM01.1	PPB	5/10/2022 14:00	SMPL
N2614-06	BH0A3	505	0.0763 -		SFAM01.1	PPB	5/10/2022 14:02	SMPL
N2614-07	BH0A4	455	0.0641 -		SFAM01.1	PPB	5/10/2022 14:05	SMPL
N2614-08	BH0A6	461	0.0656 -		SFAM01.1	PPB	5/10/2022 14:07	SMPL
N2614-09	BH0A7	212	0.0047 -		SFAM01.1	PPB	5/10/2022 14:09	SMPL
N2614-10	BH0A8	181	-0.0028 -		SFAM01.1	PPB	5/10/2022 14:11	SMPL
N2614-11	BH0A9	196	0.0008 -		SFAM01.1	PPB	5/10/2022 14:14	SMPL

LB120078

INSTRUMENT ID : CV1

N2614-12	BH0B0	411	0.0534 -	SFAM01.1	PPB	5/10/2022 14:16 SMPL
N2614-13	BH0B1	508	0.0771 -	SFAM01.1	PPB	5/10/2022 14:18 SMPL
N2614-14	BH0C8	412	0.0536 -	SFAM01.1	PPB	5/10/2022 14:20 SMPL
N2614-15	BH0D4	430	0.058 -	SFAM01.1	PPB	5/10/2022 14:23 SMPL
N2614-16	BH0D8	160	-0.008 -	SFAM01.1	PPB	5/10/2022 14:25 SMPL
N2614-17	BH0D9	335	0.0348 -	SFAM01.1	PPB	5/10/2022 14:27 SMPL
CCV017	CCV017	20661	5.0022 -	SFAM01.1	PPB	5/10/2022 14:30 SMPL
CCB017	CCB017	101	-0.0224 -	SFAM01.1	PPB	5/10/2022 14:32 SMPL
PB144715BL	PBW715	151	-0.0102 -	SFAM01.1	PPB	5/10/2022 14:34 SMPL
N2679-01	BH0A2	231	0.0094 -	SFAM01.1	PPB	5/10/2022 14:36 SMPL
N2679-02	BH0B7	169	-0.0058 -	SFAM01.1	PPB	5/10/2022 14:39 SMPL
N2679-03	BH0B8	184	-0.0021 -	SFAM01.1	PPB	5/10/2022 14:41 SMPL
N2679-04	BH0B9	470	0.0678 -	SFAM01.1	PPB	5/10/2022 14:43 SMPL
N2679-05	BH0C0	311	0.0289 -	SFAM01.1	PPB	5/10/2022 14:45 SMPL
N2679-06	BH0C1	244	0.0126 -	SFAM01.1	PPB	5/10/2022 14:48 SMPL
N2679-07	BH0C2	154	-0.0094 -	SFAM01.1	PPB	5/10/2022 14:50 SMPL
N2679-08	BH0C9	367	0.0426 -	SFAM01.1	PPB	5/10/2022 14:52 SMPL
N2679-09	BH0D0	452	0.0634 -	SFAM01.1	PPB	5/10/2022 14:54 SMPL
N2679-10	BH0E1	293	0.0245 -	SFAM01.1	PPB	5/10/2022 14:57 SMPL
N2679-11	BH0D0S	4442	1.0385 -	SFAM01.1	PPB	5/10/2022 14:59 SMPL
N2679-12	BH0D0D	432	0.0585 -	SFAM01.1	PPB	5/10/2022 15:01 SMPL
CCV018	CCV018	20975	5.0789 -	SFAM01.1	PPB	5/10/2022 15:04 SMPL
CCB018	CCB018	99	-0.0229 -	SFAM01.1	PPB	5/10/2022 15:06 SMPL