

LB 120086


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

52382



A= 0.0000e+000

B= 1.9158e-004

C= -1.1957e-002

Rho= 0.9999766

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	9.01)
0.0	0.000	0.021	0.021	173	0.000	173					7
0.2	0.200	0.214	0.014	1180	0.0 %	1180					-2
2.5	2.500	2.455	-0.045	12879	0.0 %	12879					0
5.0	5.000	5.009	0.009	26210	0.0 %	26210					0
7.5	7.500	7.476	-0.024	39087	0.0 %	39087					0
10.0	10.000	10.023	0.023	52382	0.0 %	52382					0

LB120086 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 10 May 2022 15:24:48

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	173	-	0.0000	7471B	PPB	10 May 2022 15:26:29	S	Std
0.2 - 1	50.2	1180	-	0.2000	7471B	PPB	10 May 2022 15:28:46	S	Std
2.5 - 1	52.5	12879	-	2.5000	7471B	PPB	10 May 2022 15:31:02	S	Std
5.0 - 1	55	26210	-	5.0000	7471B	PPB	10 May 2022 15:33:19	S	Std
7.5 - 1	57.5	39087	-	7.5000	7471B	PPB	10 May 2022 15:35:35	S	Std
10.0 - 1	511	52382	-	10.0000	7471B	PPB	10 May 2022 15:41:12	S	Std
ICV17 - 1	ICV17	20234	3.8645	-	7471B	PPB	10 May 2022 15:43:49	U	SMPL
ICB17 - 1	ICB17	33	-0.0056	-	7471B	PPB	10 May 2022 15:48:18	U	SMPL
CCV19 - 1	CCV19	26325	5.0314	-	7471B	PPB	10 May 2022 15:50:32	U	SMPL
CCB19 - 1	CCB19	-58	-0.0231	-	7471B	PPB	10 May 2022 15:52:47	U	SMPL
CRA - 1	CRA	1157	0.2097	-	7471B	PPB	10 May 2022 15:55:03	U	SMPL
HighStd - 1	HighStd	52610	10.0672	-	7471B	PPB	10 May 2022 15:57:18	U	SMPL
ChkStd - 1	ChkStd	36742	7.0271	-	7471B	PPB	10 May 2022 15:59:33	U	SMPL
PB144716BL - 1	PBS	40	-0.0043	-	7471B	PPB	10 May 2022 16:01:50	U	SMPL
PB144716BS - 1	LCSS	22315	4.2632	-	7471B	PPB	10 May 2022 16:04:07	U	SMPL
N2738-03 - 1	2474	298	0.0451	-	7471B	PPB	10 May 2022 16:06:23	U	SMPL
N2739-01 - 1	124030	4901	0.9270	-	7471B	PPB	10 May 2022 16:08:40	U	SMPL
N2739-03 - 1	124055	17664	3.3721	-	7471B	PPB	10 May 2022 16:10:56	U	SMPL
N2739-05 - 1	124035	6169	1.1699	-	7471B	PPB	10 May 2022 16:13:13	U	SMPL
N2739-07 - 1	124042	8679	1.6508	-	7471B	PPB	10 May 2022 16:15:29	U	SMPL
CCV20 - 1	CCV20	26416	5.0489	-	7471B	PPB	10 May 2022 16:17:45	U	SMPL
CCB20 - 1	CCB20	-39	-0.0194	-	7471B	PPB	10 May 2022 16:20:00	U	SMPL
N2745-01 - 1	MH-31S	2029	0.3768	-	7471B	PPB	10 May 2022 16:22:15	U	SMPL
N2745-01DUP - 1	MH-31SDUP	2005	0.3722	-	7471B	PPB	10 May 2022 16:24:30	U	SMPL
N2745-01MS - 1	MH-31SMS	25141	4.8046	-	7471B	PPB	10 May 2022 16:26:45	U	SMPL
N2745-01MSD - 1	MH-31MSD	23425	4.4758	-	7471B	PPB	10 May 2022 16:29:01	U	SMPL
N2745-04 - 1	MH-31D	1355	0.2476	-	7471B	PPB	10 May 2022 16:31:17	U	SMPL
N2749-01 - 1	TP-1	7202	1.3678	-	7471B	PPB	10 May 2022 16:33:33	U	SMPL
N2749-04 - 1	TP-2	7081	1.3446	-	7471B	PPB	10 May 2022 16:35:50	U	SMPL
N2749-07 - 1	TP-3	17555	3.3513	-	7471B	PPB	10 May 2022 16:38:06	U	SMPL
N2749-10 - 1	TP-4	5725	1.0848	-	7471B	PPB	10 May 2022 16:40:24	U	SMPL
N2749-13 - 1	TP-5	11924	2.2725	-	7471B	PPB	10 May 2022 16:42:39	U	SMPL
CCV21 - 1	CCV21	26636	5.0910	-	7471B	PPB	10 May 2022 16:44:54	U	SMPL
CCB21 - 1	CCB21	-54	-0.0223	-	7471B	PPB	10 May 2022 16:47:10	U	SMPL
N2749-16 - 1	TP-6	4459	0.8423	-	7471B	PPB	10 May 2022 16:49:26	U	SMPL
N2749-19 - 1	TP-7	1164	0.2110	-	7471B	PPB	10 May 2022 16:51:41	U	SMPL
N2755-01 - 1	VNJ-210	1406	0.2574	-	7471B	PPB	10 May 2022 16:53:56	U	SMPL
N2755-04 - 1	VNJ-221	2645	0.4948	-	7471B	PPB	10 May 2022 16:56:12	U	SMPL
N2759-01 - 1	PL-02-050922	2176	0.4049	-	7471B	PPB	10 May 2022 16:58:27	U	SMPL
N2760-01 - 1	TP-34	356	0.0562	-	7471B	PPB	10 May 2022 17:00:43	U	SMPL
N2745-01LX5 - 1		372	0.0593	-	7471B	PPB	10 May 2022 17:02:59	U	SMPL
N2745-01A - 1		23933	4.5732	-	7471B	PPB	10 May 2022 17:05:17	U	SMPL
CCV22 - 1	CCV22	26658	5.0952	-	7471B	PPB	10 May 2022 17:07:34	U	SMPL
CCB22 - 1	CCB22	-56	-0.0227	-	7471B	PPB	10 May 2022 17:09:50	U	SMPL