

LB120099


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

39523



A= 0.0000e+000

B= 2.4806e-004

C= 1.0567e-002

Rho= 0.9988414

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.035	0.035	100	0.000	100					
0.2	0.200	0.225	0.025	863	0.0 %	863					13
2.5	2.500	2.233	-0.267	8960	0.0 %	8960					-11
5.0	5.000	5.246	0.246	21105	0.0 %	21105					5
7.5	7.500	7.646	0.146	30781	0.0 %	30781					2
10.0	10.000	9.815	-0.185	39523	0.0 %	39523					-2

LB120099 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 11 May 2022 07:31:09					
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	100	-	0.0000	7470A	PPB	11 May 2022 07:44:16	S	Std
0.2 - 1	50.2	863	-	0.2000	7470A	PPB	11 May 2022 07:46:33	S	Std
2.5 - 1	52.5	8960	-	2.5000	7470A	PPB	11 May 2022 07:48:50	S	Std
5.0 - 1	55	21105	-	5.0000	7470A	PPB	11 May 2022 07:51:07	S	Std
7.5 - 1	57.5	30781	-	7.5000	7470A	PPB	11 May 2022 07:53:23	S	Std
10.0 - 1	50	39523	-	10.0000	7470A	PPB	11 May 2022 07:58:14	S	Std
ICV18 - 1	ICV18	15799	3.9297	-	7470A	PPB	11 May 2022 08:01:02	U	SMPL
ICB18 - 1	ICB18	22	0.0160	-	7470A	PPB	11 May 2022 08:03:17	U	SMPL
CCV23 - 1	CCV23	21046	5.2313	-	7470A	PPB	11 May 2022 08:05:32	U	SMPL
CCB23 - 1	CCB23	11	0.0133	-	7470A	PPB	11 May 2022 08:07:47	U	SMPL
CRA - 1	CRA	730	0.1917	-	7470A	PPB	11 May 2022 08:10:02	U	SMPL
HighStd - 1	HighStd	38942	9.6706	-	7470A	PPB	11 May 2022 08:12:17	U	SMPL
ChkStd - 1	ChkStd	28767	7.1466	-	7470A	PPB	11 May 2022 08:14:32	U	SMPL
PB144724BL - 1	PBW	-72	-0.0073	-	7470A	PPB	11 May 2022 08:16:48	U	SMPL
PB144724BS - 1	LCSW	15093	3.7546	-	7470A	PPB	11 May 2022 08:19:04	U	SMPL
N2748-01 - 1	72-11940	-3	0.0098	-	7470A	PPB	11 May 2022 08:21:20	U	SMPL
N2748-02 - 1	72-11991	76	0.0294	-	7470A	PPB	11 May 2022 08:23:36	U	SMPL
N2749-02 - 1	TP-1	68	0.0274	-	7470A	PPB	11 May 2022 08:25:52	U	SMPL
N2749-05 - 1	TP-2	68	0.0274	-	7470A	PPB	11 May 2022 08:28:10	U	SMPL
N2749-08 - 1	TP-3	96	0.0344	-	7470A	PPB	11 May 2022 08:30:24	U	SMPL
CCV24 - 1	CCV24	20957	5.2092	-	7470A	PPB	11 May 2022 08:32:39	U	SMPL
CCB24 - 1	CCB24	24	0.0165	-	7470A	PPB	11 May 2022 08:34:55	U	SMPL
N2749-11 - 1	TP-4	47	0.0222	-	7470A	PPB	11 May 2022 08:37:11	U	SMPL
N2749-14 - 1	TP-5	134	0.0438	-	7470A	PPB	11 May 2022 08:39:25	U	SMPL
N2749-17 - 1	TP-6	47	0.0222	-	7470A	PPB	11 May 2022 08:41:41	U	SMPL
N2749-20 - 1	TP-7	61	0.0257	-	7470A	PPB	11 May 2022 08:43:56	U	SMPL
N2755-02 - 1	VNJ-210	59	0.0252	-	7470A	PPB	11 May 2022 08:46:12	U	SMPL
N2755-05 - 1	VNJ-221	78	0.0299	-	7470A	PPB	11 May 2022 08:48:29	U	SMPL
N2760-02 - 1	TP-34	82	0.0309	-	7470A	PPB	11 May 2022 08:50:44	U	SMPL
N2760-02DUP - 1	TP-34DUP	78	0.0299	-	7470A	PPB	11 May 2022 08:53:01	U	SMPL
N2760-02MS - 1	TP-34MS	13283	3.3056	-	7470A	PPB	11 May 2022 08:55:18	U	SMPL
N2760-02MSD - 1	TP-34MSD	13500	3.3594	-	7470A	PPB	11 May 2022 08:57:33	U	SMPL
CCV25 - 1	CCV25	20889	5.1923	-	7470A	PPB	11 May 2022 08:59:49	U	SMPL
CCB25 - 1	CCB25	10	0.0130	-	7470A	PPB	11 May 2022 09:02:04	U	SMPL
PB144725BL - 1	PBW	62	0.0259	-	7470A	PPB	11 May 2022 09:04:19	U	SMPL
PB144725BS - 1	LCSW	15209	3.7833	-	7470A	PPB	11 May 2022 09:06:34	U	SMPL
N2738-01 - 1	2463	6	0.0121	-	7470A	PPB	11 May 2022 09:08:50	U	SMPL
N2738-02 - 1	2468	70	0.0279	-	7470A	PPB	11 May 2022 09:11:05	U	SMPL
N2739-02 - 1	124030	96	0.0344	-	7470A	PPB	11 May 2022 09:13:21	U	SMPL
N2739-04 - 1	124055	77	0.0297	-	7470A	PPB	11 May 2022 09:15:38	U	SMPL
N2739-06 - 1	124035	65	0.0267	-	7470A	PPB	11 May 2022 09:17:54	U	SMPL
N2739-08 - 1	124042	87	0.0321	-	7470A	PPB	11 May 2022 09:20:10	U	SMPL
N2742-01 - 1	COMP-1	74	0.0289	-	7470A	PPB	11 May 2022 09:22:27	U	SMPL
N2742-02 - 1	COMP-2	74	0.0289	-	7470A	PPB	11 May 2022 09:24:42	U	SMPL
CCV26 - 1	CCV26	19734	4.9058	-	7470A	PPB	11 May 2022 09:26:58	U	SMPL
CCB26 - 1	CCB26	8	0.0126	-	7470A	PPB	11 May 2022 09:29:13	U	SMPL
N2745-02 - 1	MH-31S	45	0.0217	-	7470A	PPB	11 May 2022 09:31:28	U	SMPL
N2745-05 - 1	MH-31D	86	0.0319	-	7470A	PPB	11 May 2022 09:33:44	U	SMPL
N2745-05DUP - 1	MH-31DDUP	70	0.0279	-	7470A	PPB	11 May 2022 09:35:59	U	SMPL
N2745-05MS - 1	MH-31DMS	14371	3.5755	-	7470A	PPB	11 May 2022 09:38:14	U	SMPL
N2745-05MSD - 1	MH-31DMSD	14180	3.5281	-	7470A	PPB	11 May 2022 09:40:30	U	SMPL
PB144726BL - 1	PBW	9	0.0128	-	7470A	PPB	11 May 2022 09:42:46	U	SMPL
PB144726BS - 1	LCSW	16213	4.0324	-	7470A	PPB	11 May 2022 09:45:02	U	SMPL
N2742-03 - 1	17149	67	0.0272	-	7470A	PPB	11 May 2022 09:47:19	U	SMPL
N2742-04 - 1	17152	43	0.0212	-	7470A	PPB	11 May 2022 09:49:35	U	SMPL
N2742-04DUP - 1	17152DUP	53	0.0237	-	7470A	PPB	11 May 2022 09:51:51	U	SMPL
CCV27 - 1	CCV27	19591	4.8703	-	7470A	PPB	11 May 2022 09:54:06	U	SMPL
CCB27 - 1	CCB27	0	0.0106	-	7470A	PPB	11 May 2022 09:56:22	U	SMPL
N2742-04MS - 1	17152MS	14649	3.6444	-	7470A	PPB	11 May 2022 09:58:38	U	SMPL
N2742-04MSD - 1	17152MSD	14521	3.6127	-	7470A	PPB	11 May 2022 10:00:54	U	SMPL
PB144727BL - 1	PBW	-6	0.0091	-	7470A	PPB	11 May 2022 10:03:09	U	SMPL
PB144727BS - 1	LCSW	13636	3.3931	-	7470A	PPB	11 May 2022 10:05:25	U	SMPL
N2734-03 - 1	RL-TOTE-COMP	69	0.0277	-	7470A	PPB	11 May 2022 10:07:41	U	SMPL
N2734-05 - 1	FMI-120	91	0.0331	-	7470A	PPB	11 May 2022 10:09:56	U	SMPL
N2754-01 - 1	A748	766	0.2006	-	7470A	PPB	11 May 2022 10:17:30	U	SMPL
N2754-01DUP - 1	A748DUP	721	0.1894	-	7470A	PPB	11 May 2022 10:19:47	U	SMPL
N2754-01MS - 1	A748MS	15349	3.8181	-	7470A	PPB	11 May 2022 10:22:03	U	SMPL
N2754-01MSD - 1	A748MSD	14631	3.6400	-	7470A	PPB	11 May 2022 10:24:19	U	SMPL
CCV28 - 1	CCV28	20045	4.9830	-	7470A	PPB	11 May 2022 10:26:35	U	SMPL
CCB28 - 1	CCB28	19	0.0153	-	7470A	PPB	11 May 2022 10:31:41	U	SMPL
N2760-02LX5 - 1		3	0.0113	-	7470A	PPB	11 May 2022 10:33:56	U	SMPL
N2760-02A - 1		15599	3.8801	-	7470A	PPB	11 May 2022 10:36:12	U	SMPL
N2745-05LX5 - 1		-36	0.0016	-	7470A	PPB	11 May 2022 10:38:28	U	SMPL

LB120099 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 11 May 2022 07:31:09

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
N2745-05A - 1		15617	3.8845		-7470A	PPB	11 May 2022 10:40:44	U	SMPL
N2742-04LX5 - 1		-54	-0.0028		-7470A	PPB	11 May 2022 10:43:00	U	SMPL
N2742-04A - 1		16085	4.0006		-7470A	PPB	11 May 2022 10:45:16	U	SMPL
N2754-01LX5 - 1		67	0.0272		-7470A	PPB	11 May 2022 10:47:32	U	SMPL
N2754-01A - 1		15449	3.8429		-7470A	PPB	11 May 2022 10:49:49	U	SMPL
CCV29 - 1	CCV29	20124	5.0026		-7470A	PPB	11 May 2022 10:52:06	U	SMPL
CCB29 - 1	CCB29	10	0.0130		-7470A	PPB	11 May 2022 10:56:24	U	SMPL
PB144701TB - 1	PB144701TB	61	0.0257		-7470A	PPB	11 May 2022 10:58:39	U	SMPL
PB144653TB - 1	PB144653TB	71	0.0282		-7470A	PPB	11 May 2022 11:00:55	U	SMPL
PB144654TB - 1	PB144654TB	59	0.0252		-7470A	PPB	11 May 2022 11:03:11	U	SMPL
CCV30 - 1	CCV30	18533	4.6079		-7470A	PPB	11 May 2022 11:05:27	U	SMPL
CCB30 - 1	CCB30	-109	-0.0165		-7470A	PPB	11 May 2022 11:07:43	U	SMPL