



## Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB120101

Analyst By : jignesh

Supervisor Review By : mohammad

Slope : 98.7

pH Meter ID : WC PH METER-1

| Calibration Standards          | Chemtech Log# |
|--------------------------------|---------------|
| PH 4 BUFFER SOLUTION           | W2886         |
| buffer solution pH 7 yellow    | W2813         |
| PH 10.01 BUFFER,COLOR CD 475ML | W2809         |
| buffer solution pH 7 yellow    | W2887         |
| Buffer Solution, PH2 (500ml)   | W2806         |
| Buffer Solution, PH12 (500ml)  | W2807         |

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

| Seq | LabID       | DF | Matrix | Weight (gm) | Volume (ml) | Temperature (°C) | Result (pH) | Anal Date  | Anal Time |
|-----|-------------|----|--------|-------------|-------------|------------------|-------------|------------|-----------|
| 1   | CAL1        | 1  | Water  | NA          | NA          | 20.3             | 4.01        | 05/11/2022 | 11:40     |
| 2   | CAL2        | 1  | Water  | NA          | NA          | 20.3             | 7.00        | 05/11/2022 | 11:41     |
| 3   | CAL3        | 1  | Water  | NA          | NA          | 20.3             | 10.03       | 05/11/2022 | 11:42     |
| 4   | ICV         | 1  | Water  | NA          | NA          | 20.4             | 6.98        | 05/11/2022 | 11:45     |
| 5   | CCV1        | 1  | Water  | NA          | NA          | 20.4             | 2.01        | 05/11/2022 | 11:47     |
| 6   | N2778-02    | 1  | Water  | NA          | NA          | 20.3             | 9.73        | 05/11/2022 | 11:50     |
| 7   | N2778-02DUP | 1  | Water  | NA          | NA          | 20.4             | 9.74        | 05/11/2022 | 11:51     |
| 8   | CCV2        | 1  | Water  | NA          | NA          | 20.4             | 12.03       | 05/11/2022 | 12:00     |

# WORKLIST(Hardcopy Internal Chain)

VB 120101

WorkList Name : PH N2785      WorkList ID : 159506      Department : Wet-Chemistry      Date : 05-11-2022 11:32:32

| Due Date   | Matrix | Sample   | Test | Preservative | Customer | Raw Sample Storage Location | Customer Sample       | Collect Date | Method |
|------------|--------|----------|------|--------------|----------|-----------------------------|-----------------------|--------------|--------|
| 05/09/2022 | Water  | N2778-02 | pH   | Cool 4 deg C | TETR06   | K33                         | TT-043-IDWGW-20220509 | 05/09/2022   | 9040C  |

Date/Time 05.11.22 11:35  
 Raw Sample Received by: JD (CPC)  
 Raw Sample Relinquished by: Cap Sm

Date/Time 05.11.22 12:40  
 Raw Sample Received by: JD (CPC)  
 Raw Sample Relinquished by: JD (CPC)