

245.1


LB120105

INSTRUMENT ID: CV1

Linear

 μ Abs.:

37583



A= 0.0000e+000

B= 2.6733e-004

C= -3.2360e-002

Rho= 0.9999830

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.011	-0.011	79	0.000	79					
0.05	0.050	0.042	-0.008	279	0.0 %	279	0				-16
0.2	0.200	0.202	0.002	877	0.0 %	877					1
2.5	2.500	2.544	0.044	9639	0.0 %	9639					2
5.0	5.000	4.978	-0.022	18742	0.0 %	18742					0
7.5	7.500	7.480	-0.020	28101	0.0 %	28101					0
10.0	10.000	10.015	0.015	37583	0.0 %	37583					0

LB120105 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 11 May 2022 12:17:08

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	79	-	0.0000	245.1	PPB	11 May 2022 12:25:38	S	Std
0.05 - 1	50.05	279	-	0.0500	245.1	PPB	11 May 2022 12:27:55	S	Std
0.2 - 1	50.2	877	-	0.2000	245.1	PPB	11 May 2022 12:30:12	S	Std
2.5 - 1	52.5	9639	-	2.5000	245.1	PPB	11 May 2022 12:32:28	S	Std
5.0 - 1	55	18742	-	5.0000	245.1	PPB	11 May 2022 12:34:44	S	Std
7.5 - 1	57.5	28101	-	7.5000	245.1	PPB	11 May 2022 12:37:00	S	Std
10.0 - 1	51.0	37583	-	10.0000	245.1	PPB	11 May 2022 12:44:27	S	Std
ICV19 - 1	ICV19	14638	3.8808	-245.1		PPB	11 May 2022 12:47:07	U	SMPL
ICB19 - 1	ICB19	38	-0.0222	-245.1		PPB	11 May 2022 12:49:22	U	SMPL
CCV31 - 1	CCV31	19230	5.1084	-245.1		PPB	11 May 2022 12:51:37	U	SMPL
CCB31 - 1	CCB31	41	-0.0214	-245.1		PPB	11 May 2022 12:53:52	U	SMPL
CRA - 1	CRA	253	0.0353	-245.1		PPB	11 May 2022 12:56:07	U	SMPL
HighStd - 1	HighStd	37128	9.8931	-245.1		PPB	11 May 2022 12:58:23	U	SMPL
ChkStd - 1	ChkStd	26587	7.0751	-245.1		PPB	11 May 2022 13:00:37	U	SMPL
PB144751BL - 1	PBW	-144	-0.0709	-245.1		PPB	11 May 2022 13:02:55	U	SMPL
PB144751BS - 1	LCSW	13905	3.6849	-245.1		PPB	11 May 2022 13:05:11	U	SMPL
N2741-01 - 1	TWP01	1518	0.3734	-245.1		PPB	11 May 2022 13:07:27	U	SMPL
N2756-02 - 1	HP-640	42	-0.0211	-245.1		PPB	11 May 2022 13:09:43	U	SMPL
N2757-02 - 1	TI-658	146	0.0067	-245.1		PPB	11 May 2022 13:12:00	U	SMPL
N2761-01 - 1	TWP04	369	0.0663	-245.1		PPB	11 May 2022 13:14:16	U	SMPL
CCV32 - 1	CCV32	18679	4.9611	-245.1		PPB	11 May 2022 13:16:31	U	SMPL
CCB32 - 1	CCB32	18	-0.0275	-245.1		PPB	11 May 2022 13:18:46	U	SMPL
N2761-02 - 1	TWP05	363	0.0647	-245.1		PPB	11 May 2022 13:25:47	U	SMPL
N2761-02DUP - 1	TWP05DUP	364	0.0649	-245.1		PPB	11 May 2022 13:28:02	U	SMPL
N2761-02MS - 1	TWP05MS	14842	3.9353	-245.1		PPB	11 May 2022 13:30:17	U	SMPL
N2761-02MSD - 1	TWP05MSD	13977	3.7041	-245.1		PPB	11 May 2022 13:32:32	U	SMPL
N2761-02LX5 - 1		52	-0.0185	-245.1		PPB	11 May 2022 13:34:48	U	SMPL
N2761-02A - 1		14521	3.8495	-245.1		PPB	11 May 2022 13:37:03	U	SMPL
CCV33 - 1	CCV33	18802	4.9940	-245.1		PPB	11 May 2022 13:39:19	U	SMPL
CCB33 - 1	CCB33	14	-0.0286	-245.1		PPB	11 May 2022 13:41:35	U	SMPL