

LB120142


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

37397



A= 0.0000e+000

B= 2.6578e-004

C= -6.6954e-002

Rho= 0.9994828

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.062	-0.062	17	0.000	17					
0.2	0.200	0.220	0.020	1079	0.0 %	1079					10
2.5	2.500	2.403	-0.097	9292	0.0 %	9292					-4
5.0	5.000	5.230	0.230	19931	0.0 %	19931					5
7.5	7.500	7.537	0.037	28612	0.0 %	28612					0
10.0	10.000	9.872	-0.128	37397	0.0 %	37397					-1

LB120142 INSTRUMENT ID : CV1

Method: 7470A	Operator: Admin	Date of Analysis: 13 May 2022 09:18:42					Date	Type	Type
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units			
0.0 - 1	50	17	-	0.0000	7470A	PPB	13 May 2022 09:24:10	S	Std
0.2 - 1	50.2	1079	-	0.2000	7470A	PPB	13 May 2022 09:37:35	S	Std
2.5 - 1	50.5	9292	-	2.5000	7470A	PPB	13 May 2022 09:39:55	S	Std
5.0 - 1	55	19931	-	5.0000	7470A	PPB	13 May 2022 09:42:57	S	Std
7.5 - 1	57.5	28612	-	7.5000	7470A	PPB	13 May 2022 09:45:13	S	Std
10.0 - 1	57.0	37397	-	10.0000	7470A	PPB	13 May 2022 09:48:59	S	Std
ICV22 - 1	ICV22	15180	3.9675	-	7470A	PPB	13 May 2022 09:51:59	U	SMPL
ICB22 - 1	ICB22	-39	-0.0773	-	7470A	PPB	13 May 2022 09:56:00	U	SMPL
CCV39 - 1	CCV39	17772	4.6564	-	7470A	PPB	13 May 2022 09:58:47	U	SMPL
CCB39 - 1	CCB39	106	-0.0388	-	7470A	PPB	13 May 2022 10:01:41	U	SMPL
CRA - 1	CRA	1091	0.2230	-	7470A	PPB	13 May 2022 10:07:43	U	SMPL
HighStd - 1	HighStd	36041	9.5119	-	7470A	PPB	13 May 2022 10:10:36	U	SMPL
ChkStd - 1	ChkStd	25059	6.5931	-	7470A	PPB	13 May 2022 10:15:42	U	SMPL
PB144765BL - 1	PBW	32	-0.0584	-	7470A	PPB	13 May 2022 10:22:11	U	SMPL
PB144765BS - 1	LCSW	15165	3.9635	-	7470A	PPB	13 May 2022 10:27:43	U	SMPL
N2762-01 - 1	SB02	199	-0.0141	-	7470A	PPB	13 May 2022 10:31:25	U	SMPL
N2762-02 - 1	SB03	280	0.0075	-	7470A	PPB	13 May 2022 10:33:46	U	SMPL
N2762-03 - 1	SB04	-133	-0.1023	-	7470A	PPB	13 May 2022 10:36:48	U	SMPL
N2762-04 - 1	SB05	553	0.0800	-	7470A	PPB	13 May 2022 10:39:09	U	SMPL
N2762-05 - 1	SB15	63	-0.0502	-	7470A	PPB	13 May 2022 10:41:24	U	SMPL
CCV40 - 1	CCV40	18701	4.9033	-	7470A	PPB	13 May 2022 10:44:27	U	SMPL
CCB40 - 1	CCB40	33	-0.0582	-	7470A	PPB	13 May 2022 10:47:11	U	SMPL
N2762-06 - 1	SB16	148	-0.0276	-	7470A	PPB	13 May 2022 10:49:37	U	SMPL
N2762-07 - 1	SB17	277	0.0067	-	7470A	PPB	13 May 2022 10:51:59	U	SMPL
N2762-08 - 1	SB18	-6	-0.0685	-	7470A	PPB	13 May 2022 10:55:37	U	SMPL
N2762-09 - 1	SB19	-15	-0.0709	-	7470A	PPB	13 May 2022 10:58:08	U	SMPL
N2775-02 - 1	MH-35S	83	-0.0449	-	7470A	PPB	13 May 2022 11:00:27	U	SMPL
N2775-05 - 1	MH-35D	3	-0.0662	-	7470A	PPB	13 May 2022 11:02:54	U	SMPL
N2775-05DUP - 1	MH-35DDUP	244	-0.0021	-	7470A	PPB	13 May 2022 11:05:17	U	SMPL
N2775-05MS - 1	MH-35DMS	15103	3.9471	-	7470A	PPB	13 May 2022 11:08:43	U	SMPL
N2775-05MSD - 1	MH-35DMSD	14811	3.8695	-	7470A	PPB	13 May 2022 11:14:51	U	SMPL
PB144767BL - 1	PBW	388	0.0362	-	7470A	PPB	13 May 2022 11:17:27	U	SMPL
CCV41 - 1	CCV41	20667	5.4258	-	7470A	PPB	13 May 2022 11:20:59	U	SMPL
CCB41 - 1	CCB41	248	-0.0010	-	7470A	PPB	13 May 2022 11:24:41	U	SMPL
PB144767BS - 1	LCSW	15210	3.9755	-	7470A	PPB	13 May 2022 11:28:23	U	SMPL
N2770-01 - 1	MH-15820	59	-0.0513	-	7470A	PPB	13 May 2022 11:33:09	U	SMPL
N2770-01DUP - 1	MH-15820DUP	412	0.0425	-	7470A	PPB	13 May 2022 11:36:31	U	SMPL
N2770-01MS - 1	MH-15820MS	12724	3.3148	-	7470A	PPB	13 May 2022 11:39:05	U	SMPL
N2770-01MSD - 1	MH-15820MSD	13791	3.5984	-	7470A	PPB	13 May 2022 11:45:53	U	SMPL
N2770-02 - 1	MH-15834	277	0.0067	-	7470A	PPB	13 May 2022 11:49:24	U	SMPL
N2770-03 - 1	MH-15842	384	0.0351	-	7470A	PPB	13 May 2022 11:51:59	U	SMPL
N2770-04 - 1	MH-15844	131	-0.0321	-	7470A	PPB	13 May 2022 11:54:24	U	SMPL
N2770-06 - 1	M-9747	-69	-0.0853	-	7470A	PPB	13 May 2022 11:56:45	U	SMPL
N2770-07 - 1	M-21518	158	-0.0250	-	7470A	PPB	13 May 2022 11:59:13	U	SMPL
CCV42 - 1	CCV42	18141	4.7545	-	7470A	PPB	13 May 2022 12:02:04	U	SMPL
CCB42 - 1	CCB42	134	-0.0313	-	7470A	PPB	13 May 2022 12:04:59	U	SMPL
N2770-09 - 1	M-21519	306	0.0144	-	7470A	PPB	13 May 2022 12:07:58	U	SMPL
N2778-02 - 1	TT-043-IDWGW-20220509	90	-0.0430	-	7470A	PPB	13 May 2022 12:10:23	U	SMPL
PB144723TB - 1		124	-0.0340	-	7470A	PPB	13 May 2022 12:12:39	U	SMPL
N2775-05LX5 - 1		-104	-0.0946	-	7470A	PPB	13 May 2022 12:14:59	U	SMPL
N2775-05A - 1		15008	3.9218	-	7470A	PPB	13 May 2022 12:18:15	U	SMPL
N2770-01LX5 - 1		-9	-0.0693	-	7470A	PPB	13 May 2022 12:20:40	U	SMPL
N2770-01A - 1		15415	4.0300	-	7470A	PPB	13 May 2022 12:22:59	U	SMPL
CCV43 - 1	CCV43	19492	5.1136	-	7470A	PPB	13 May 2022 12:27:54	U	SMPL
CCB43 - 1	CCB43	62	-0.0505	-	7470A	PPB	13 May 2022 12:30:12	U	SMPL