



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB120147

Analyst By : jignesh

Supervisor Review By : mohammad

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	05/13/2022	10:15
2	CAL2	1	Water	NA	NA	20.6	7.00	05/13/2022	10:16
3	CAL3	1	Water	NA	NA	20.7	10.03	05/13/2022	10:18
4	ICV	1	Water	NA	NA	20.7	7.01	05/13/2022	10:20
5	CCV1	1	Water	NA	NA	20.8	2.02	05/13/2022	10:40
6	N2765-01	1	Water	NA	NA	21.4	4.64	05/13/2022	10:42
7	N2765-02	1	Water	NA	NA	21.3	4.93	05/13/2022	10:45
8	N2765-03	1	Water	NA	NA	20.8	5.15	05/13/2022	10:47
9	N2813-02	1	Water	NA	NA	23.2	6.90	05/13/2022	10:49
10	N2813-03	1	Water	NA	NA	23.2	7.65	05/13/2022	10:52
11	N2813-05	1	Water	NA	NA	22.9	7.09	05/13/2022	11:00
12	N2813-06	1	Water	NA	NA	22.8	5.28	05/13/2022	11:03
13	N2830-04	1	Water	NA	NA	23.0	7.10	05/13/2022	11:05
14	N2835-01	1	Water	NA	NA	20.3	4.65	05/13/2022	11:09
15	N2835-02	1	Water	NA	NA	20.4	4.66	05/13/2022	11:11
16	CCV2	1	Water	NA	NA	20.6	12.02	05/13/2022	11:15
17	N2835-03	1	Water	NA	NA	21.1	4.70	05/13/2022	11:19
18	N2837-02	1	Water	NA	NA	20.1	5.67	05/13/2022	11:20
19	N2838-02	1	Water	NA	NA	20.2	6.24	05/13/2022	11:23
20	N2838-02DUP	1	Water	NA	NA	20.3	6.25	05/13/2022	11:25

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	CCV3	1	Water	NA	NA	20.9	2.01	05/13/2022	11:30

WORKLIST(Hardcopy Internal Chain)

13 1201147

WorkList Name : ph n2814 WorkList ID : 159569 Department : Wet-Chemistry Date : 05-13-2022 10:05:02

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/09/2022	Water	N2765-01	pH	Cool 4 deg C	STER07	K31	CELL-A	05/09/2022	9040C
05/09/2022	Water	N2765-02	pH	Cool 4 deg C	STER07	K31	CELL-B/C	05/09/2022	9040C
05/09/2022	Water	N2765-03	pH	Cool 4 deg C	STER07	K31	CELL-D	05/09/2022	9040C
05/12/2022	Water	N2813-02	pH	Cool 4 deg C	PSEG03	K33	BUR-1148	05/12/2022	9040C
05/12/2022	Water	N2813-03	pH	Cool 4 deg C	PSEG03	K33	BUR-1012	05/12/2022	9040C
05/12/2022	Water	N2813-05	pH	Cool 4 deg C	PSEG03	K33	BUR-1149	05/12/2022	9040C
05/12/2022	Water	N2813-06	pH	Cool 4 deg C	PSEG03	K33	BUR-1144	05/12/2022	9040C
05/12/2022	Water	N2830-04	pH	Cool 4 deg C	PSEG03	K33	AUD-1594	05/12/2022	9040C
05/12/2022	Water	N2835-01	pH	Cool 4 deg C	STER13	L11	CELL-A-METALS-ANAYLSIS	05/12/2022	9040C
05/12/2022	Water	N2835-02	pH	Cool 4 deg C	STER13	L11	CELL-B-METALS-ANAYLSIS	05/12/2022	9040C
05/12/2022	Water	N2835-03	pH	Cool 4 deg C	STER13	L11	CELL-C-METALS-ANAYLSIS	05/12/2022	9040C
05/12/2022	Water	N2837-02	pH	Cool 4 deg C	TETR06	L11	RW8-IDWGW-20220512	05/12/2022	9040C
05/11/2022	Water	N2838-02	pH	Cool 4 deg C	TETR06	L11	TT-044-IDWGW-20220511	05/11/2022	9040C

Date/Time 05/13/2022 10:10
 Raw Sample Received by: JPC/SM
 Raw Sample Relinquished by: JPC/SM

Date/Time 05/13/22 13:00
 Raw Sample Received by: JPC/SM
 Raw Sample Relinquished by: JPC/SM