

LB120153


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

50352



A= 0.0000e+000

B= 1.9900e-004

C= -4.8684e-002

Rho= 0.9999747

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.005	-0.005	218	0.000	218					
0.2	0.200	0.196	-0.004	1230	0.0 %	1230					-2
2.5	2.500	2.475	-0.025	12683	0.0 %	12683					-1
5.0	5.000	5.050	0.050	25621	0.0 %	25621					1
7.5	7.500	7.513	0.013	37998	0.0 %	37998					0
10.0	10.000	9.971	-0.029	50352	0.0 %	50352					0

LB120153 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 13 May 2022 14:05:32

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	218	-	0.0000	7471B	PPB	13 May 2022 14:08:01	S	Std
0.2 - 1	502	1230	-	0.2000	7471B	PPB	13 May 2022 14:10:29	S	Std
2.5 - 1	525	12683	-	2.5000	7471B	PPB	13 May 2022 14:17:22	S	Std
5.0 - 1	55	25621	-	5.0000	7471B	PPB	13 May 2022 14:19:53	S	Std
7.5 - 1	575	37998	-	7.5000	7471B	PPB	13 May 2022 14:22:25	S	Std
10.0 - 1	510	50352	-	10.0000	7471B	PPB	13 May 2022 14:25:02	S	Std
ICV23 - 1	ICV23	20058	3.9428	-	7471B	PPB	13 May 2022 14:28:00	U	SMPL
ICB23 - 1	ICB23	120	-0.0248	-	7471B	PPB	13 May 2022 14:32:49	U	SMPL
CCV44 - 1	CCV44	25245	4.9750	-	7471B	PPB	13 May 2022 14:35:19	U	SMPL
CCB44 - 1	CCB44	88	-0.0312	-	7471B	PPB	13 May 2022 14:42:38	U	SMPL
CRA - 1	CRA	1159	0.1820	-	7471B	PPB	13 May 2022 14:45:07	U	SMPL
HighStd - 1	HighStd	49963	9.8939	-	7471B	PPB	13 May 2022 14:51:04	U	SMPL
ChkStd - 1	ChkStd	35174	6.9509	-	7471B	PPB	13 May 2022 14:53:25	U	SMPL
PB144809BL - 1	PBS	49	-0.0389	-	7471B	PPB	13 May 2022 14:55:56	U	SMPL
PB144809BS - 1	LCSS	19946	3.9205	-	7471B	PPB	13 May 2022 15:03:54	U	SMPL
N2770-10 - 1	M-15848	225	-0.0039	-	7471B	PPB	13 May 2022 15:06:28	U	SMPL
N2792-01 - 1	RO-371	859	0.1223	-	7471B	PPB	13 May 2022 15:09:06	U	SMPL
N2795-02 - 1	VNJ-239	9028	1.7479	-	7471B	PPB	13 May 2022 15:13:55	U	SMPL
N2802-01 - 1	SHALLOW-WC-1	5447	1.0353	-	7471B	PPB	13 May 2022 15:16:45	U	SMPL
N2802-01DUP - 1	SHALLOW-WC-1DUP	5678	1.0812	-	7471B	PPB	13 May 2022 15:19:09	U	SMPL
CCV45 - 1	CCV45	24742	4.8749	-	7471B	PPB	13 May 2022 15:21:42	U	SMPL
CCB45 - 1	CCB45	-149	-0.0783	-	7471B	PPB	13 May 2022 15:24:36	U	SMPL
N2802-03 - 1	DEEP-WC-1	1862	0.3219	-	7471B	PPB	13 May 2022 15:33:17	U	SMPL
N2809-01 - 1	WC-14A-TP-5R	27894	5.5022	-	7471B	PPB	13 May 2022 15:35:36	U	SMPL
N2823-01 - 1	SC-1	2589	0.4665	-	7471B	PPB	13 May 2022 15:38:09	U	SMPL
N2823-02 - 1	SC-2	648	0.0803	-	7471B	PPB	13 May 2022 15:40:37	U	SMPL
N2823-03 - 1	SC-3	2767	0.5019	-	7471B	PPB	13 May 2022 15:43:31	U	SMPL
N2823-04 - 1	SC-4	1414	0.2327	-	7471B	PPB	13 May 2022 15:45:49	U	SMPL
N2823-05 - 1	SC-5	755	0.1016	-	7471B	PPB	13 May 2022 15:48:21	U	SMPL
N2823-06 - 1	SC-6	1225	0.1951	-	7471B	PPB	13 May 2022 15:50:53	U	SMPL
N2802-01MS - 1	SHALLOW-WC-1MS	27622	5.4481	-	7471B	PPB	13 May 2022 15:53:24	U	SMPL
N2802-01MSD - 1	SHALLOW-WC-1MSD	27697	5.4630	-	7471B	PPB	13 May 2022 15:55:55	U	SMPL
CCV46 - 1	CCV46	25038	4.9338	-	7471B	PPB	13 May 2022 15:58:35	U	SMPL
CCB46 - 1	CCB46	18	-0.0451	-	7471B	PPB	13 May 2022 16:00:52	U	SMPL
N2823-07 - 1	SC-7	1490	0.2478	-	7471B	PPB	13 May 2022 16:03:08	U	SMPL
N2823-08 - 1	SC-8	1246	0.1993	-	7471B	PPB	13 May 2022 16:05:26	U	SMPL
N2823-09 - 1	SC-9	1918	0.3330	-	7471B	PPB	13 May 2022 16:07:51	U	SMPL
N2823-10 - 1	SC-10	1139	0.1780	-	7471B	PPB	13 May 2022 16:10:13	U	SMPL
N2823-11 - 1	SC-11	2360	0.4210	-	7471B	PPB	13 May 2022 16:12:35	U	SMPL
N2823-12 - 1	SC-12	217	-0.0055	-	7471B	PPB	13 May 2022 16:15:32	U	SMPL
N2823-13 - 1	SC-13	2467	0.4422	-	7471B	PPB	13 May 2022 16:18:02	U	SMPL
N2823-14 - 1	SC-14	539	0.0586	-	7471B	PPB	13 May 2022 16:20:38	U	SMPL
N2802-01LX5 - 1		1072	0.1646	-	7471B	PPB	13 May 2022 16:22:55	U	SMPL
N2802-01A - 1		25696	5.0648	-	7471B	PPB	13 May 2022 16:25:16	U	SMPL
CCV47 - 1	CCV47	25041	4.9344	-	7471B	PPB	13 May 2022 16:27:45	U	SMPL
CCB47 - 1	CCB47	297	0.0104	-	7471B	PPB	13 May 2022 16:30:10	U	SMPL