



PERCENT SOLID

Supervisor: Sohil
Analyst: jignesh
Date: 6/2/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 14:30
In Date: 05/14/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:22
Out Date: 05/15/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB120175

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N2840-01	YB6X2	1	0.92	8.79	9.71	9.62	99.0	
N2840-02	YB6X3	2	0.93	8.87	9.8	9.66	98.4	
N2840-03	YB6X4	3	0.93	8.82	9.75	9.53	97.5	
N2840-04	YB6X5	4	0.91	8.61	9.52	9.31	97.6	
N2840-05	YB6X6	5	0.91	8.59	9.5	9.44	99.3	
N2840-06	YB6X7	6	0.9	9.07	9.97	9.73	97.4	
N2840-07	YB6X8	7	0.91	8.80	9.71	9.59	98.6	
N2840-08	YB6X9	8	0.91	8.72	9.63	9.34	96.7	
N2840-09	YB6Y0	9	0.93	8.88	9.81	9.61	97.7	
N2840-10	YB6Y1	10	0.93	8.79	9.72	9.22	94.3	
N2840-11	YB6Y2	11	0.91	9.03	9.94	8.99	89.5	
N2840-12	YB6Y3	12	0.94	8.62	9.56	8.69	89.9	
N2840-13	YB6Y4	13	0.92	8.88	9.8	9.03	91.3	
N2840-14	YB6Y5	14	0.93	8.58	9.51	9.28	97.3	
N2840-15	YB6Y6	15	0.93	8.60	9.53	9.31	97.4	
N2840-16	YB6Y7	16	0.93	8.87	9.8	9.3	94.4	
N2840-17	YB6Y8	17	0.95	8.70	9.65	9.15	94.3	
N2840-18	YB6Y9	18	0.91	8.82	9.73	9.16	93.5	
N2840-19	YB6Y9D	19	0.91	8.82	9.73	9.16	93.5	
N2840-20	YB6Y9S	20	0.91	8.82	9.73	9.16	93.5	
N2840-21	YB6Z0	21	0.9	8.80	9.7	9.17	94.0	
N2840-22	YB6Z1	22	0.88	8.63	9.51	8.7	90.6	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$