



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB120176

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.3	05/14/2022	09:15
2	N2754-01		1	15	100.00	>212.0	>212.0	05/14/2022	09:35
3	N2754-01DUP		1	15	100.00	>212.0	>212.0	05/14/2022	09:55
4	N2761-01		1	12	100.00	>212.0	>212.0	05/14/2022	10:20
5	N2761-02		1	14	100.00	>212.0	>212.0	05/14/2022	10:45
6	N2813-02		1	15	100.00	>212.0	>212.0	05/14/2022	11:10
7	N2813-03		1	13	100.00	>212.0	>212.0	05/14/2022	11:32
8	N2813-05		1	20	100.00	>212.0	>212.0	05/14/2022	11:55
9	N2813-06		1	18	100.00	>212.0	>212.0	05/14/2022	12:10
10	N2814-13		1	17	100.00	>212.0	>212.0	05/14/2022	12:28
11	N2830-04		1	20	100.00	>212.0	>212.0	05/14/2022	12:40
12	N2833-01		1	18	100.00	>212.0	>212.0	05/14/2022	13:05
13	N2833-02		1	17	100.00	>212.0	>212.0	05/14/2022	13:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB120176

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp51422 WorkList ID : 159625 Department : Wet-Chemistry Date : 05-14-2022 08:51:29

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/16/2022	Water	N2761-01	Flash Point	Cool 4 deg C	loui01	K32	TWP04	05/09/2022	1010B
05/16/2022	Water	N2761-02	Flash Point	Cool 4 deg C	loui01	K32	TWP05	05/09/2022	1010B
05/19/2022	Water	N2833-01	Flash Point	Cool 4 deg C	loui01	K33	TWP02	05/11/2022	1010B
05/19/2022	Water	N2833-02	Flash Point	Cool 4 deg C	loui01	K33	TWP03	05/12/2022	1010B
05/16/2022	Water	N2754-01	Flash Point	Cool 4 deg C	PSEG03	K31	40828	05/09/2022	1010B
05/19/2022	Water	N2813-02	Flash Point	Cool 4 deg C	PSEG03	K33	BUR-1148	05/12/2022	1010B
05/19/2022	Water	N2813-03	Flash Point	Cool 4 deg C	PSEG03	K33	BUR-1012	05/12/2022	1010B
05/19/2022	Water	N2813-05	Flash Point	Cool 4 deg C	PSEG03	K33	BUR-1149	05/12/2022	1010B
05/19/2022	Water	N2813-06	Flash Point	Cool 4 deg C	PSEG03	K33	BUR-1144	05/12/2022	1010B
05/17/2022	Water	N2814-13	Flash Point	Cool 4 deg C	PSEG03	K33	HEATER-WATER	05/12/2022	1010B
05/19/2022	Water	N2830-04	Flash Point	Cool 4 deg C	PSEG03	K33	AUD-1594	05/12/2022	1010B

Date/Time 05/14/22 08:55
Raw Sample Received by: KS(wc)
Raw Sample Relinquished by: JH 4027

Date/Time 05/14/22 13:50
Raw Sample Received by: JH 4027
Raw Sample Relinquished by: KS(wc)