

7470A


LB120181

INSTRUMENT ID: CV1

Linear

 μ Abs.:

35692



A= 0.0000e+000

B= 2.8129e-004

C= -9.4615e-002

Rho= 0.9987196

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.111	-0.111	-59	0.000	-59					-
0.2	0.200	0.224	0.024	1133	0.0 %	1133					12
2.5	2.500	2.452	-0.048	9052	0.0 %	9052					-2
5.0	5.000	5.389	0.389	19494	0.0 %	19494					8
7.5	7.500	7.301	-0.199	26292	0.0 %	26292					-3
10.0	10.000	9.945	-0.055	35692	0.0 %	35692					-1

LB120181 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 14 May 2022 13:28:20

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-59	-	0.0000	7470A	PPB	14 May 2022 13:30:00	S	Std
0.2 - 1	50.2	1133	-	0.2000	7470A	PPB	14 May 2022 13:37:20	S	Std
2.5 - 1	52.5	9052	-	2.5000	7470A	PPB	14 May 2022 13:42:32	S	Std
5.0 - 1	55	19494	-	5.0000	7470A	PPB	14 May 2022 13:46:03	S	Std
7.5 - 1	57.5	26292	-	7.5000	7470A	PPB	14 May 2022 13:49:30	S	Std
10.0 - 1	510	35692	-	10.0000	7470A	PPB	14 May 2022 14:13:15	S	Std
ICV24 - 1	ICV24	14599	4.0120	-	7470A	PPB	14 May 2022 14:16:09	U	SMPL
ICB24 - 1	ICB24	27	-0.0870	-	7470A	PPB	14 May 2022 14:18:50	U	SMPL
CCV48 - 1	CCV48	19679	5.4410	-	7470A	PPB	14 May 2022 14:22:19	U	SMPL
CCB48 - 1	CCB48	70	-0.0749	-	7470A	PPB	14 May 2022 14:25:23	U	SMPL
CRA - 1	CRA	1170	0.2345	-	7470A	PPB	14 May 2022 14:50:27	U	SMPL
HighStd - 1	HighStd	35299	9.8348	-	7470A	PPB	14 May 2022 14:54:32	U	SMPL
ChkStd - 1	ChkStd	25557	7.0944	-	7470A	PPB	14 May 2022 14:57:35	U	SMPL
PB144833BL - 1	PBW	580	0.0685	-	7470A	PPB	14 May 2022 15:08:29	U	SMPL
PB144833BS - 1	LCSW	13101	3.5906	-	7470A	PPB	14 May 2022 15:10:48	U	SMPL
N2793-01 - 1	RW9-MW01S-20220510	-5	-0.0960	-	7470A	PPB	14 May 2022 15:17:23	U	SMPL
N2793-01DUP - 1	RW9-MW01S-20220510DUP	544	0.0584	-	7470A	PPB	14 May 2022 15:24:43	U	SMPL
N2793-01MS - 1	RW9-MW01S-20220510MS	13349	3.6604	-	7470A	PPB	14 May 2022 15:27:41	U	SMPL
N2793-01MSD - 1	RW9-MW01S-20220510MSD	13942	3.8272	-	7470A	PPB	14 May 2022 15:34:17	U	SMPL
N2793-02 - 1	RW9-MW01D1-20220510	-309	-0.1815	-	7470A	PPB	14 May 2022 15:41:36	U	SMPL
CCV49 - 1	CCV49	18972	5.2421	-	7470A	PPB	14 May 2022 15:48:54	U	SMPL
CCB49 - 1	CCB49	517	0.0508	-	7470A	PPB	14 May 2022 15:51:17	U	SMPL
N2793-03 - 1	RW9-MW-01D2-20220510	222	-0.0322	-	7470A	PPB	14 May 2022 15:58:35	U	SMPL
N2793-04 - 1	RW9-MW01D3-20220510	135	-0.0566	-	7470A	PPB	14 May 2022 16:05:54	U	SMPL
N2793-05 - 1	RW9-MW01D2-F-20220510	248	-0.0249	-	7470A	PPB	14 May 2022 16:08:14	U	SMPL
N2793-06 - 1	RW9-MW01D3-F-20220510	22	-0.0884	-	7470A	PPB	14 May 2022 16:15:32	U	SMPL
N2830-04 - 1	AUD-1594	10877	2.9650	-	7470A	PPB	14 May 2022 16:22:50	U	SMPL
N2837-02 - 1	RW8-IDWGW-20220512	119	-0.0611	-	7470A	PPB	14 May 2022 16:30:09	U	SMPL
N2838-02 - 1	TT-044-IDWGW-20220511	9	-0.0921	-	7470A	PPB	14 May 2022 16:34:16	U	SMPL
N2793-01LX5 - 1		-70	-0.1143	-	7470A	PPB	14 May 2022 16:36:43	U	SMPL
N2793-01A - 1		13869	3.8067	-	7470A	PPB	14 May 2022 16:42:30	U	SMPL
CCV50 - 1	CCV50	19239	5.3172	-	7470A	PPB	14 May 2022 16:45:04	U	SMPL
CCB50 - 1	CCB50	192	-0.0406	-	7470A	PPB	14 May 2022 16:47:56	U	SMPL