


LB120182

7471B

INSTRUMENT 11): CV1

Linear



A= 0.0000e+000

B= 1.7918e-004

C= -1.2639e-001

Rho= 0.9998743

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.065	-0.065	344	0.000	344					3
0.2	0.200	0.206	0.006	1853	0.0 %	1853					1
2.5	2.500	2.518	0.018	14759	0.0 %	14759					1
5.0	5.000	5.062	0.062	28956	0.0 %	28956					1
7.5	7.500	7.567	0.067	42934	0.0 %	42934					1
10.0	10.000	9.912	-0.088	56025	0.0 %	56025					1

LB120182 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 14 May 2022 16:57:13

Sample ID	Extended ID	µ Abs.	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	S0	344	-	0.00007471B	PPB	14 May 2022 16:57:25	S	Std
0.2 - 1	S0.2	1853	-	0.20007471B	PPB	14 May 2022 16:59:46	S	Std
2.5 - 1	S2.5	14759	-	2.50007471B	PPB	14 May 2022 17:03:26	S	Std
5.0 - 1	S5	28956	-	5.00007471B	PPB	14 May 2022 17:07:07	S	Std
7.5 - 1	S7.5	42934	-	7.50007471B	PPB	14 May 2022 17:10:05	S	Std
10.0 - 1	S10	56025	-	10.00007471B	PPB	14 May 2022 17:15:12	S	Std
ICV25 - 1	ICV25	22835	3.9652	-7471B	PPB	14 May 2022 17:22:47	U	SMPL
ICB25 - 1	ICB25	226	-0.0859	-7471B	PPB	14 May 2022 17:25:06	U	SMPL
CCV51 - 1	CCV51	28121	4.9124	-7471B	PPB	14 May 2022 17:27:52	U	SMPL
CCB51 - 1	CCB51	305	-0.0717	-7471B	PPB	14 May 2022 17:35:09	U	SMPL
CRA - 1	CRA	1650	0.1693	-7471B	PPB	14 May 2022 17:45:35	U	SMPL
HighStd - 1	HighStd	53734	9.5018	-7471B	PPB	14 May 2022 17:48:25	U	SMPL
ChkStd - 1	ChkStd	37465	6.5867	-7471B	PPB	14 May 2022 17:51:04	U	SMPL
PB144834BL - 1	PBS	335	-0.0664	-7471B	PPB	14 May 2022 17:53:26	U	SMPL
PB144834BS - 1	LCSS	19915	3.4420	-7471B	PPB	14 May 2022 17:57:06	U	SMPL
N2813-01 - 1	BUR-1103	757	0.0092	-7471B	PPB	14 May 2022 17:59:58	U	SMPL
N2820-01 - 1	STONE-PILE	9280	1.5364	-7471B	PPB	14 May 2022 18:02:36	U	SMPL
N2826-01 - 1	PL-05-051222	1839	0.2031	-7471B	PPB	14 May 2022 18:05:25	U	SMPL
N2827-01 - 1	40579	15872	2.7176	-7471B	PPB	14 May 2022 18:08:33	U	SMPL
N2829-01 - 1	OK-02-051222	1132	0.0764	-7471B	PPB	14 May 2022 18:13:53	U	SMPL
CCV52 - 1	CCV52	27757	4.8472	-7471B	PPB	14 May 2022 18:16:17	U	SMPL
CCB52 - 1	CCB52	69	-0.1140	-7471B	PPB	14 May 2022 18:18:37	U	SMPL
N2830-01 - 1	AUD-1275	483	-0.0398	-7471B	PPB	14 May 2022 18:21:04	U	SMPL
N2834-01 - 1	TR-06-051222	1976	0.2277	-7471B	PPB	14 May 2022 18:25:26	U	SMPL
N2847-01 - 1	PSC-124019	3555	0.5106	-7471B	PPB	14 May 2022 18:28:50	U	SMPL
N2847-03 - 1	REDBOX-164718	2819	0.3787	-7471B	PPB	14 May 2022 18:31:09	U	SMPL
N2851-01 - 1	OR-02-051322	4241	0.6335	-7471B	PPB	14 May 2022 18:34:49	U	SMPL
N2855-01 - 1	NB-08-051322	637	-0.0123	-7471B	PPB	14 May 2022 18:37:30	U	SMPL
N2856-01 - 1	PL-02-051322	3496	0.5000	-7471B	PPB	14 May 2022 18:39:48	U	SMPL
N2856-01DUP - 1	PL-02-051322DUP	3697	0.5360	-7471B	PPB	14 May 2022 18:42:08	U	SMPL
N2856-01MS - 1	PL-02-051322MS	22357	3.8796	-7471B	PPB	14 May 2022 18:44:58	U	SMPL
N2856-01MSD - 1	PL-02-051322MSD	22148	3.8421	-7471B	PPB	14 May 2022 18:47:26	U	SMPL
CCV53 - 1	CCV53	26889	4.6916	-7471B	PPB	14 May 2022 18:50:44	U	SMPL
CCB53 - 1	CCB53	330	-0.0673	-7471B	PPB	14 May 2022 18:53:02	U	SMPL
N2856-01LX5 - 1		86	-0.1110	-7471B	PPB	14 May 2022 18:55:48	U	SMPL
N2856-01A - 1		21835	3.7861	-7471B	PPB	14 May 2022 19:00:17	U	SMPL
CCV54 - 1	CCV54	27022	4.7155	-7471B	PPB	14 May 2022 19:06:59	U	SMPL
CCB54 - 1	CCB54	344	-0.0648	-7471B	PPB	14 May 2022 19:09:41	U	SMPL