



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB120188

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	05/16/2022	10:35
2	CAL2	1	Water	NA	NA	20.4	7.00	05/16/2022	10:36
3	CAL3	1	Water	NA	NA	20.4	10.03	05/16/2022	10:37
4	ICV	1	Water	NA	NA	20.5	6.98	05/16/2022	10:40
5	CCV1	1	Water	NA	NA	20.6	2.01	05/16/2022	10:42
6	N2808-01	1	Solid	20.03	20	24.3	8.70	05/16/2022	10:45
7	N2808-01DUP	1	Solid	20.04	20	24.4	8.71	05/16/2022	10:46
8	CCV2	1	Water	NA	NA	20.3	12.03	05/16/2022	10:49

WORKLIST(Hardcopy Internal Chain)

VP 120188

WorkList Name : PH N2808 WorkList ID : 159657 Department : Wet-Chemistry Date : 05-16-2022 10:25:26

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/12/2022	Solid	N2808-01	pH	Cool 4 deg C	ROUX02	L11	WC-14B5-TP	05/12/2022	9045D

Date/Time 05-16-22 10:30
 Raw Sample Received by: JDC
 Raw Sample Relinquished by: JDC

Date/Time 05-16-22 12:30
 Raw Sample Received by: JDC
 Raw Sample Relinquished by: JDC