

LB 120189

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : 12 Instrument ID : Konelab

5/16/2022 16:02

Test: Phenolics-

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.000	0.0	0.107	
ICB1	0.002	0.0	0.005	
CCV1	0.981	0.0	0.105	
CCB1	0.001	0.0	0.005	
PB144832BL	0.002	0.0	0.005	
PB144832BS	1.019	0.0	0.109	
N2833-01	-0.000	0.0	0.005	
N2833-01DUP	-0.002	0.0	0.005	
N2833-01MS	1.070	0.0	0.114	
N2833-01MSD	1.067	0.0	0.114	
N2833-02	0.122	0.0	0.017	
N2852-03	0.003	0.0	0.005	
N2852-04	0.004	0.0	0.005	
CCV2	0.980	0.0	0.105	
CCB2	0.001	0.0	0.005	

N 15
Mean 0.417
SD 0.5109
CV% 122.66

Aquakem v. 7.2AQ1

Results from time period:

Mon May 16 15:12:22 2022

Mon May 16 15:58:15 2022

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.00PPM	A	Phenolics-	P	0	mg/l	5/16/2022 15:12:22	
0.05PPM	A	Phenolics-	P	0.0517	mg/l	5/16/2022 15:12:23	
0.1PPM	A	Phenolics-	P	0.1011	mg/l	5/16/2022 15:12:24	
0.25PPM	A	Phenolics-	P	0.2512	mg/l	5/16/2022 15:12:25	
0.50PPM	A	Phenolics-	P	0.4935	mg/l	5/16/2022 15:12:26	
1.0PPM	A	Phenolics-	P	1.0021	mg/l	5/16/2022 15:12:27	
2.0PPM	A	Phenolics-	P	2.0003	mg/l	5/16/2022 15:12:28	
ICV1	S	Phenolics-	P	0.9998	mg/l	5/16/2022 15:50:54	
ICB1	S	Phenolics-	P	0.0022	mg/l	5/16/2022 15:50:56	
CCV1	S	Phenolics-	P	0.981	mg/l	5/16/2022 15:50:58	
CCB1	S	Phenolics-	P	0.0007	mg/l	5/16/2022 15:51:00	
PB144832BL	S	Phenolics-	P	0.002	mg/l	5/16/2022 15:51:01	
PB144832BS	S	Phenolics-	P	1.019	mg/l	5/16/2022 15:51:03	
N2833-01	S	Phenolics-	P	-0.0001	mg/l	5/16/2022 15:58:07	
N2833-01DUP	S	Phenolics-	P	-0.0019	mg/l	5/16/2022 15:58:08	
N2833-01MS	S	Phenolics-	P	1.0698	mg/l	5/16/2022 15:58:09	
N2833-01MSD	S	Phenolics-	P	1.0668	mg/l	5/16/2022 15:58:10	
N2833-02	S	Phenolics-	P	0.1216	mg/l	5/16/2022 15:58:11	
N2852-03	S	Phenolics-	P	0.003	mg/l	5/16/2022 15:58:12	
N2852-04	S	Phenolics-	P	0.0036	mg/l	5/16/2022 15:58:13	
CCV2	S	Phenolics-	P	0.9798	mg/l	5/16/2022 15:58:14	
CCB2	S	Phenolics-	P	0.0012	mg/l	5/16/2022 15:58:15	

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092Reviewed by : R Instrument ID : Konelab5/16/2022 15:22


Test Phenolics-

Accepted 5/16/2022 15:22

Factor 9.822
Bias 0.005

Coeff. of det. 0.999983

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.00PPM	0.005	0.0000	0.0000	—
2	Phenol-2	0.010	0.0517	0.0500	3.4
3	Phenol-2	0.015	0.1011	0.1000	1.1
4	Phenol-2	0.031	0.2512	0.2500	0.5
5	Phenol-2	0.055	0.4935	0.5000	-1.3
6	Phenol-2	0.107	1.0021	1.0000	0.2
7	Phenol-2	0.209	2.0003	2.0000	0.0

12
5/16/22