


LB12022)

7470A

INSTRUMENT ID: CV2

 μ Abs.:

34337



Linear

A= 0.0000e+000

B= 2.8982e-004

C= 1.6953e-002

Rho= 0.9999759

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	6/10/11
0.0	0.000	0.015	0.015	-8	0.000	-8					
0.2	0.200	0.165	-0.035	510	0.0 %	510					-18
2.5	2.500	2.498	-0.002	8562	0.0 %	8562					0
5.0	5.000	5.032	0.032	17305	0.0 %	17305					1
7.5	7.500	7.521	0.021	25893	0.0 %	25893					0
10.0	10.000	9.969	-0.031	34337	0.0 %	34337					0

LB120221 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 18 May 2022 09:29:39

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-8	-	0.00007470A	PPB		18 May 2022 10:54:22	S	Std
0.2 - 1	50.2	510	-	0.20007470A	PPB		18 May 2022 10:56:39	S	Std
2.5 - 1	52.5	8562	-	2.50007470A	PPB		18 May 2022 10:58:56	S	Std
5.0 - 1	55	17305	-	5.00007470A	PPB		18 May 2022 11:01:13	S	Std
7.5 - 1	57.5	25893	-	7.50007470A	PPB		18 May 2022 11:03:29	S	Std
10.0 - 1	510	34337	-	10.00007470A	PPB		18 May 2022 11:05:45	S	Std
ICV26 - 1	ICV26	13209	3.8452	-7470A	PPB		18 May 2022 11:08:58	U	SMPL
ICB26 - 1	ICB26	84	0.0413	-7470A	PPB		18 May 2022 11:11:13	U	SMPL
CCV55 - 1	CCV55	16830	4.8947	-7470A	PPB		18 May 2022 11:13:27	U	SMPL
CCB55 - 1	CCB55	-100	-0.0120	-7470A	PPB		18 May 2022 11:15:42	U	SMPL
CRA - 1	CRA	551	0.1766	-7470A	PPB		18 May 2022 11:17:57	U	SMPL
HighStd - 1	HighStd	34361	9.9755	-7470A	PPB		18 May 2022 11:20:12	U	SMPL
ChkStd - 1	ChkStd	23543	6.8402	-7470A	PPB		18 May 2022 11:22:27	U	SMPL
PB144845BL - 1	PBW	-43	0.0045	-7470A	PPB		18 May 2022 11:24:43	U	SMPL
PB144845BS - 1	LCSW	12042	3.5070	-7470A	PPB		18 May 2022 11:27:00	U	SMPL
N2806-02 - 1	WC-819-1	16	0.0216	-7470A	PPB		18 May 2022 11:29:16	U	SMPL
N2809-02 - 1	WC-14A-TP-5R	37	0.0277	-7470A	PPB		18 May 2022 11:31:32	U	SMPL
N2832-13 - 1	SB01	14	0.0210	-7470A	PPB		18 May 2022 11:33:48	U	SMPL
N2832-13DUP - 1	SB01DUP	24	0.0239	-7470A	PPB		18 May 2022 11:36:05	U	SMPL
N2832-13MS - 1	SB01MS	14378	4.1840	-7470A	PPB		18 May 2022 11:38:20	U	SMPL
CCV56 - 1	CCV56	16864	4.9045	-7470A	PPB		18 May 2022 11:40:35	U	SMPL
CCB56 - 1	CCB56	-39	0.0056	-7470A	PPB		18 May 2022 11:42:51	U	SMPL
N2832-13MSD - 1	SB01MSD	12052	3.5099	-7470A	PPB		18 May 2022 11:45:06	U	SMPL
N2832-14 - 1	SB08	6	0.0187	-7470A	PPB		18 May 2022 11:47:21	U	SMPL
N2832-15 - 1	SB21	29	0.0254	-7470A	PPB		18 May 2022 11:49:36	U	SMPL
N2832-16 - 1	SB20	43	0.0294	-7470A	PPB		18 May 2022 11:51:51	U	SMPL
N2832-17 - 1	SB06	38	0.0280	-7470A	PPB		18 May 2022 11:54:07	U	SMPL
N2832-18 - 1	SB07	21	0.0230	-7470A	PPB		18 May 2022 11:56:24	U	SMPL
N2832-19 - 1	SB09	30	0.0256	-7470A	PPB		18 May 2022 11:58:40	U	SMPL
N2832-20 - 1	SB10	38	0.0280	-7470A	PPB		18 May 2022 12:00:57	U	SMPL
N2832-21 - 1	SB12	36	0.0274	-7470A	PPB		18 May 2022 12:03:13	U	SMPL
N2832-22 - 1	SB11	19	0.0225	-7470A	PPB		18 May 2022 12:05:29	U	SMPL
CCV57 - 1	CCV57	16751	4.8718	-7470A	PPB		18 May 2022 12:07:44	U	SMPL
CCB57 - 1	CCB57	-42	0.0048	-7470A	PPB		18 May 2022 12:10:00	U	SMPL
N2832-23 - 1	SB13	18	0.0222	-7470A	PPB		18 May 2022 12:12:15	U	SMPL
N2832-24 - 1	SB14	42	0.0291	-7470A	PPB		18 May 2022 12:14:31	U	SMPL
PB144846BL - 1	PBW	61	0.0346	-7470A	PPB		18 May 2022 12:16:46	U	SMPL
PB144846BS - 1	LCSW	11381	3.3154	-7470A	PPB		18 May 2022 12:24:29	U	SMPL
N2814-13 - 1	HEATER-WATER	14	0.0210	-7470A	PPB		18 May 2022 12:26:45	U	SMPL
N2814-13DUP - 1	HEATER-WATERDUP	-2	0.0164	-7470A	PPB		18 May 2022 12:29:00	U	SMPL
N2814-13MS - 1	HEATER-WATERMS	11332	3.3012	-7470A	PPB		18 May 2022 12:31:17	U	SMPL
N2814-13MSD - 1	HEATER-WATERMSD	11583	3.3740	-7470A	PPB		18 May 2022 12:53:19	U	SMPL
N2850-01 - 1	40014	0	0.0170	-7470A	PPB		18 May 2022 12:55:39	U	SMPL
PB144862BL - 1	PBW	410	0.1358	-7470A	PPB		18 May 2022 12:56:03	U	SMPL
CCV58 - 1	CCV58	17667	5.1372	-7470A	PPB		18 May 2022 13:06:17	U	SMPL
CCB58 - 1	CCB58	211	0.0781	-7470A	PPB		18 May 2022 13:11:30	U	SMPL
PB144862BS - 1	LCSW	15397	4.4793	-7470A	PPB		18 May 2022 13:13:51	U	SMPL
N2847-02 - 1	PSC-124019	230	0.0836	-7470A	PPB		18 May 2022 13:16:19	U	SMPL
N2847-04 - 1	REDBOX-164718	-328	-0.0781	-7470A	PPB		18 May 2022 13:19:03	U	SMPL
N2858-02 - 1	TP-38	-191	-0.0384	-7470A	PPB		18 May 2022 13:21:30	U	SMPL
N2858-02DUP - 1	TP-38DUP	241	0.0868	-7470A	PPB		18 May 2022 13:23:50	U	SMPL
N2858-02MS - 1	TP-38MS	11856	3.4531	-7470A	PPB		18 May 2022 13:26:31	U	SMPL
N2858-02MSD - 1	TP-38MSD	11639	3.3902	-7470A	PPB		18 May 2022 13:28:51	U	SMPL
N2824-11 - 1	COM-1	-448	-0.1129	-7470A	PPB		18 May 2022 13:31:10	U	SMPL
N2824-12 - 1	COM-2	-24	0.0100	-7470A	PPB		18 May 2022 13:33:28	U	SMPL
N2824-13 - 1	COM-3	279	0.0978	-7470A	PPB		18 May 2022 13:35:49	U	SMPL
CCV59 - 1	CCV59	16273	4.7332	-7470A	PPB		18 May 2022 13:38:07	U	SMPL
CCB59 - 1	CCB59	127	0.0538	-7470A	PPB		18 May 2022 13:40:25	U	SMPL
N2824-14 - 1	SAS-1	75	0.0387	-7470A	PPB		18 May 2022 13:42:49	U	SMPL
PB144819TB - 1	PB144819TB	-30	0.0083	-7470A	PPB		18 May 2022 13:47:38	U	SMPL
N2832-13LX5 - 1		213	0.0787	-7470A	PPB		18 May 2022 13:52:32	U	SMPL
N2832-13A - 1		11836	3.4473	-7470A	PPB		18 May 2022 13:54:49	U	SMPL
N2814-13LX5 - 1		227	0.0827	-7470A	PPB		18 May 2022 13:57:05	U	SMPL
N2814-13A - 1		12293	3.5797	-7470A	PPB		18 May 2022 13:59:26	U	SMPL
N2858-02LX5 - 1		88	0.0425	-7470A	PPB		18 May 2022 14:07:12	U	SMPL
N2858-02A - 1		13157	3.8301	-7470A	PPB		18 May 2022 14:09:58	U	SMPL
PB144780TB - 1	PB144780TB	146	0.0593	-7470A	PPB		18 May 2022 14:12:34	U	SMPL
PB144818TB - 1	PB144818TB	102	0.0465	-7470A	PPB		18 May 2022 14:17:09	U	SMPL
CCV60 - 1	CCV60	17189	4.9987	-7470A	PPB		18 May 2022 14:22:42	U	SMPL
CCB60 - 1	CCB60	-408	-0.1013	-7470A	PPB		18 May 2022 14:24:58	U	SMPL