



PERCENT SOLID

Supervisor: mohammad
Analyst: jignesh
Date: 5/19/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 17:00
In Date: 05/18/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:35
Out Date: 05/19/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB120222

Table with 9 columns: Lab ID, Client SampleID, Dish #, Dish Wt (g) (A), Sample Wt (g), Dish + Sample Wt (g) (B), Dish+Dry Sample Wt (g) (C), % Solid, Comments. It contains 28 rows of data.



PERCENT SOLID

Supervisor: mohammad
Analyst: jignesh
Date: 5/19/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 17:00
In Date: 05/18/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:35
Out Date: 05/19/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB120222

Table with 9 columns: Lab ID, Client SampleID, Dish #, Dish Wt (g) (A), Sample Wt (g), Dish + Sample Wt (g) (B), Dish+Dry Sample Wt (g) (C), % Solid, Comments. It contains 28 rows of data.



PERCENT SOLID

Supervisor: mohammad
Analyst: jignesh
Date: 5/19/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 17:00
In Date: 05/18/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:35
Out Date: 05/19/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB120222

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
N2895-15	P030-SS003-0006-01	57	0.93	8.57	9.5	8.24	85.3	
N2895-16	P030-SS003-0612-01	58	0.95	9.00	9.95	7.7	75.0	
N2896-01	P030-SS003-1218-01	59	0.93	9.03	9.96	7.79	76.0	
N2896-02	P030-SS003-1824-01	60	0.91	9.03	9.94	7.82	76.5	
N2896-03	P030-SS003-2430-01	61	0.97	8.95	9.92	7.38	71.6	
N2896-04	P038-SS001-0006-01	62	0.91	8.70	9.61	9.04	93.4	
N2896-05	N2896-05MS	63	0.91	8.70	9.61	9.04	93.4	
N2896-06	N2896-05MSD	64	0.91	8.70	9.61	9.04	93.4	
N2896-07	P038-SS001-0006-02	65	0.91	9.05	9.96	9.6	96.0	
N2896-08	P038-SS001-0612-01	66	0.93	8.91	9.84	8.99	90.5	
N2896-09	P038-SS001-1218-01	67	0.93	8.65	9.58	8.47	87.2	
N2896-10	P038-SS001-1824-01	68	0.93	8.74	9.67	8.16	82.7	
N2896-11	P038-SS001-2430-01	69	0.91	8.69	9.6	7.75	78.7	
N2896-12	P039-SS001-0006-01	70	0.94	8.66	9.6	9.12	94.5	
N2896-13	P039-SS001-0612-01	71	0.91	8.66	9.57	8.93	92.6	
N2896-14	P039-SS001-1218-01	72	0.93	8.81	9.74	8.54	86.4	
N2896-15	P039-SS001-1824-01	73	0.88	8.70	9.58	7.6	77.2	
N2896-16	P039-SS001-2430-01	74	0.94	9.02	9.96	8.08	79.2	
N2896-17	P040-SS001-0006-01	75	0.91	8.91	9.82	9.39	95.2	
N2896-18	P040-SS001-0612-01	76	0.97	8.85	9.82	9.00	90.7	
N2896-19	P040-SS001-1218-01	77	0.91	8.88	9.79	8.73	88.1	
N2896-20	P040-SS001-1824-01	78	0.95	9.00	9.95	8.92	88.6	
N2896-21	P040-SS001-2430-01	79	0.91	8.90	9.81	7.38	72.7	
N2897-01	P041-SS001-0006-01	81	0.9	8.67	9.57	8.53	88.0	
N2897-02	P041-SS001-0612-01	82	0.91	8.98	9.89	8.68	86.5	
N2897-03	P041-SS001-1218-01	83	0.95	9.02	9.97	9.04	89.7	
N2897-04	P041-SS001-1824-01	84	0.95	8.98	9.93	8.00	78.5	
N2897-05	P041-SS001-2430-01	85	0.93	9.01	9.94	7.68	74.9	



PERCENT SOLID

Supervisor: mohammad
Analyst: jignesh
Date: 5/19/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 17:00
In Date: 05/18/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:35
Out Date: 05/19/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB120222

Table with 9 columns: Lab ID, Client SampleID, Dish #, Dish Wt (g) (A), Sample Wt (g), Dish + Sample Wt (g) (B), Dish+Dry Sample Wt (g) (C), % Solid, Comments. It contains 25 rows of data.

% Solid = (C-A) * 100 / (B-A)