

LB120233


7471B

INSTRUMENT ID: CV2

Linear

 μ Abs.:

55318



A= 0.0000e+000

B= 1.8086e-004

C= 1.1954e-002

Rho= 0.9999223

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.036	0.036	131	0.000	131					-9
0.2	0.200	0.183	-0.017	943	0.0 %	943					-2
2.5	2.500	2.447	-0.053	13465	0.0 %	13465					1
5.0	5.000	5.072	0.072	27975	0.0 %	27975					-1
7.5	7.500	7.446	-0.054	41103	0.0 %	41103					0
10.0	10.000	10.017	0.017	55318	0.0 %	55318					

LB120233 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 18 May 2022 15:40:44

Sample ID	Extended ID	μ Abs.	Conc.	Std. Conc	Method	Units	Date	Type	Type
0.0 - 1	50	131	-	0.0000	7471B	PPB	18 May 2022 15:43:33	S	Std
0.2 - 1	50.2	943	-	0.2000	7471B	PPB	18 May 2022 15:45:50	S	Std
2.5 - 1	52.5	13465	-	2.5000	7471B	PPB	18 May 2022 15:48:07	S	Std
5.0 - 1	55	27975	-	5.0000	7471B	PPB	18 May 2022 15:50:23	S	Std
7.5 - 1	57.5	41103	-	7.5000	7471B	PPB	18 May 2022 15:52:40	S	Std
10.0 - 1	59	55318	-	10.0000	7471B	PPB	18 May 2022 15:54:57	S	Std
ICV27 - 1	ICV27	21606	3.9197	-	7471B	PPB	18 May 2022 15:57:44	U	SMPL
ICB27 - 1	ICB27	-69	-0.0005	-	7471B	PPB	18 May 2022 15:59:59	U	SMPL
CCV61 - 1	CCV61	28143	5.1020	-	7471B	PPB	18 May 2022 16:02:15	U	SMPL
CCB61 - 1	CCB61	-52	0.0025	-	7471B	PPB	18 May 2022 16:04:30	U	SMPL
CRA - 1	CRA	875	0.1702	-	7471B	PPB	18 May 2022 16:06:47	U	SMPL
HighStd - 1	HighStd	56685	10.2642	-	7471B	PPB	18 May 2022 16:09:02	U	SMPL
ChkStd - 1	ChkStd	39728	7.1973	-	7471B	PPB	18 May 2022 16:11:17	U	SMPL
PB144905BL - 1	PBS	-32	0.0062	-	7471B	PPB	18 May 2022 16:13:35	U	SMPL
PB144905BS - 1	LCSS	20633	3.7437	-	7471B	PPB	18 May 2022 16:15:53	U	SMPL
N2824-01 - 1	COM-1	1051	0.2020	-	7471B	PPB	18 May 2022 16:18:09	U	SMPL
N2824-02 - 1	COM-2	632	0.1263	-	7471B	PPB	18 May 2022 16:20:26	U	SMPL
N2824-03 - 1	COM-3	497	0.1018	-	7471B	PPB	18 May 2022 16:22:43	U	SMPL
N2824-04 - 1	SAS-1	3843	0.7070	-	7471B	PPB	18 May 2022 16:25:00	U	SMPL
N2863-01 - 1	P003-SS003-0006-01	4996	0.9155	-	7471B	PPB	18 May 2022 16:27:14	U	SMPL
CCV62 - 1	CCV62	27435	4.9739	-	7471B	PPB	18 May 2022 16:29:30	U	SMPL
CCB62 - 1	CCB62	30	0.0174	-	7471B	PPB	18 May 2022 16:31:44	U	SMPL
N2863-01DUP - 1	P003-SS003-0006-01DUP	4961	0.9092	-	7471B	PPB	18 May 2022 16:34:01	U	SMPL
N2863-01MS - 1	P003-SS003-0006-01MS	25262	4.5809	-	7471B	PPB	18 May 2022 16:36:17	U	SMPL
N2863-01MSD - 1	P003-SS003-0006-01MSD	26679	4.8372	-	7471B	PPB	18 May 2022 16:38:32	U	SMPL
N2863-02 - 1	P003-SS003-0612-01	1976	0.3693	-	7471B	PPB	18 May 2022 16:40:50	U	SMPL
N2863-03 - 1	P003-SS003-1218-01	1627	0.3062	-	7471B	PPB	18 May 2022 16:43:08	U	SMPL
N2863-04 - 1	P003-SS003-1824-01	1769	0.3319	-	7471B	PPB	18 May 2022 16:45:24	U	SMPL
N2863-05 - 1	P003-SS003-2430-01	1155	0.2209	-	7471B	PPB	18 May 2022 16:47:41	U	SMPL
N2863-06 - 1	P003-SS004-0006-01	4292	0.7882	-	7471B	PPB	18 May 2022 16:49:57	U	SMPL
N2863-11 - 1	P003-SS005-0006-01	9368	1.7063	-	7471B	PPB	18 May 2022 16:52:14	U	SMPL
N2863-16 - 1	P003-SS006-0006-01	6191	1.1317	-	7471B	PPB	18 May 2022 16:54:30	U	SMPL
CCV63 - 1	CCV63	27927	5.0629	-	7471B	PPB	18 May 2022 16:56:46	U	SMPL
CCB63 - 1	CCB63	20	0.0156	-	7471B	PPB	18 May 2022 16:59:02	U	SMPL
N2863-21 - 1	P003-SS007-0006-01	4217	0.7747	-	7471B	PPB	18 May 2022 17:01:18	U	SMPL
N2864-04 - 1	P011-SS003-0006-01	25306	4.5889	-	7471B	PPB	18 May 2022 17:03:34	U	SMPL
N2864-05 - 1	P011-SS003-0612-01	14074	2.5574	-	7471B	PPB	18 May 2022 17:05:49	U	SMPL
N2864-06 - 1	P011-SS003-1218-01	3683	0.6781	-	7471B	PPB	18 May 2022 17:08:06	U	SMPL
N2864-07 - 1	P011-SS003-1824-01	3786	0.6967	-	7471B	PPB	18 May 2022 17:10:22	U	SMPL
N2864-08 - 1	P011-SS003-2430-01	2959	0.5471	-	7471B	PPB	18 May 2022 17:12:38	U	SMPL
N2864-09 - 1	P011-SS004-0006-01	37876	6.8623	-	7471B	PPB	18 May 2022 17:14:54	U	SMPL
N2864-12 - 1	P011-SS005-0006-01	5148	0.9430	-	7471B	PPB	18 May 2022 17:17:11	U	SMPL
N2863-01LX5 - 1		933	0.1807	-	7471B	PPB	18 May 2022 17:19:30	U	SMPL
N2863-01A - 1		26914	4.8797	-	7471B	PPB	18 May 2022 17:21:45	U	SMPL
CCV64 - 1	CCV64	27952	5.0675	-	7471B	PPB	18 May 2022 17:24:00	U	SMPL
CCB64 - 1	CCB64	25	0.0165	-	7471B	PPB	18 May 2022 17:26:17	U	SMPL