



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: mohammad

Run Number: LB120234

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.3	05/18/2022	10:40
2	N2820-06		1	10	100.00	>212.0	>212.0	05/18/2022	11:10
3	N2820-06DUP		1	10	100.00	>212.0	>212.0	05/18/2022	11:35

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP51822 WorkList ID : 159747 Department : Wet-Chemistry Date : 05-18-2022 09:30:45

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/19/2022	Solid	N2820-06	Flash Point	Cool 4 deg C	PSEG03	K33	30585	05/12/2022	1010B

Date/Time 05/18/22 10:00
 Raw Sample Received by: KS (wg)
 Raw Sample Relinquished by: KS (wg)

Date/Time 05/18/22 11:40
 Raw Sample Received by: KS (wg)
 Raw Sample Relinquished by: KS (wg)

LB120234