

LB20246


7471B

INSTRUMENT ID: CV2

Linear

 μ Abs.:

58504



A= 0.0000e+000

B= 1.6902e-004

C= 3.7763e-002

Rho= 0.9995236

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.048	0.048	59	0.000	59					-
0.2	0.200	0.190	-0.010	901	0.0 %	901					-5
2.5	2.500	2.323	-0.177	13523	0.0 %	13523					-7
5.0	5.000	5.196	0.196	30520	0.0 %	30520					4
7.5	7.500	7.517	0.017	44250	0.0 %	44250					0
10.0	10.000	9.926	-0.074	58504	0.0 %	58504					-1

LB120240 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 19 May 2022 07:19:19				
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0.0 - 1	50	59	-	0.0000	7471B	PPB	19 May 2022 07:23:58	S Std
0.2 - 1	50.2	901	-	0.2000	7471B	PPB	19 May 2022 07:26:15	S Std
2.5 - 1	52.5	13523	-	2.5000	7471B	PPB	19 May 2022 07:28:32	S Std
5.0 - 1	55	30520	-	5.0000	7471B	PPB	19 May 2022 07:30:48	S Std
7.5 - 1	57.5	44250	-	7.5000	7471B	PPB	19 May 2022 07:37:54	S Std
10.0 - 1	510	58504	-	10.0000	7471B	PPB	19 May 2022 07:40:10	S Std
ICV28 - 1	ICV28	21750	3.7139	-	7471B	PPB	19 May 2022 07:42:55	U SMPL
ICB28 - 1	ICB28	-30	0.0327	-	7471B	PPB	19 May 2022 07:45:10	U SMPL
CCV65 - 1	CCV65	30702	5.2269	-	7471B	PPB	19 May 2022 07:47:25	U SMPL
CCB65 - 1	CCB65	17	0.0406	-	7471B	PPB	19 May 2022 07:49:41	U SMPL
CRA - 1	CRA	939	0.1965	-	7471B	PPB	19 May 2022 07:51:58	U SMPL
HighStd - 1	HighStd	60106	10.1967	-	7471B	PPB	19 May 2022 07:54:13	U SMPL
ChkStd - 1	ChkStd	43010	7.3072	-	7471B	PPB	19 May 2022 07:56:28	U SMPL
PB144906BL - 1	PBS	-23	0.0339	-	7471B	PPB	19 May 2022 07:58:45	U SMPL
PB144906BS - 1	LCSS	20131	3.4402	-	7471B	PPB	19 May 2022 08:01:03	U SMPL
N2858-01 - 1	TP-38	1438	0.2808	-	7471B	PPB	19 May 2022 08:03:19	U SMPL
N2865-01 - 1	P018-SS001-0006-01	6685	1.1676	-	7471B	PPB	19 May 2022 08:05:36	U SMPL
N2865-01DUP - 1	P018-SS001-0006-01DUP	6720	1.1736	-	7471B	PPB	19 May 2022 08:13:29	U SMPL
N2865-02 - 1	N2865-01MS	28973	4.9347	-	7471B	PPB	19 May 2022 08:15:45	U SMPL
N2865-03 - 1	N2865-01MSD	29811	5.0763	-	7471B	PPB	19 May 2022 08:18:00	U SMPL
CCV66 - 1	CCV66	30634	5.2154	-	7471B	PPB	19 May 2022 08:20:16	U SMPL
CCB66 - 1	CCB66	-85	0.0234	-	7471B	PPB	19 May 2022 08:22:33	U SMPL
N2865-04 - 1	P018-SS001-0006-02	8378	1.4538	-	7471B	PPB	19 May 2022 08:25:27	U SMPL
N2865-05 - 1	P018-SS001-0612-01	8900	1.5420	-	7471B	PPB	19 May 2022 08:27:42	U SMPL
N2865-06 - 1	P018-SS001-1218-01	3439	0.6190	-	7471B	PPB	19 May 2022 08:29:57	U SMPL
N2865-07 - 1	P018-SS001-1824-01	4127	0.7353	-	7471B	PPB	19 May 2022 08:32:13	U SMPL
N2865-08 - 1	P018-SS001-2430-01	2393	0.4422	-	7471B	PPB	19 May 2022 08:34:29	U SMPL
N2865-09 - 1	P019-SS001-0006-01	23005	3.9260	-	7471B	PPB	19 May 2022 08:36:45	U SMPL
N2865-10 - 1	P019-SS001-0612-01	22308	3.8082	-	7471B	PPB	19 May 2022 08:39:00	U SMPL
N2865-11 - 1	P019-SS001-1218-01	28987	4.9371	-	7471B	PPB	19 May 2022 08:41:18	U SMPL
N2865-12 - 1	P019-SS001-1824-01	28412	4.8399	-	7471B	PPB	19 May 2022 08:43:37	U SMPL
N2865-13 - 1	P019-SS001-2430-01	17604	3.0131	-	7471B	PPB	19 May 2022 08:45:54	U SMPL
CCV67 - 1	CCV67	30688	5.2246	-	7471B	PPB	19 May 2022 08:48:11	U SMPL
CCB67 - 1	CCB67	-24	0.0337	-	7471B	PPB	19 May 2022 08:50:28	U SMPL
N2865-14 - 1	P019-SS002-0006-01	2764	0.5049	-	7471B	PPB	19 May 2022 08:52:44	U SMPL
N2879-01 - 1	TP1-0516	3279	0.5920	-	7471B	PPB	19 May 2022 08:55:00	U SMPL
N2879-04 - 1	TP2-0516	4297	0.7640	-	7471B	PPB	19 May 2022 08:57:15	U SMPL
N2879-07 - 1	TP-3-0516	3915	0.6995	-	7471B	PPB	19 May 2022 08:59:31	U SMPL
N2879-10 - 1	TP-4-0516	4870	0.8609	-	7471B	PPB	19 May 2022 09:01:46	U SMPL
N2881-02 - 1	SLP-1	6215	1.0882	-	7471B	PPB	19 May 2022 09:04:02	U SMPL
N2882-02 - 1	SLP-2	6231	1.0909	-	7471B	PPB	19 May 2022 09:06:19	U SMPL
N2865-01LX5 - 1		1316	0.2602	-	7471B	PPB	19 May 2022 09:08:36	U SMPL
N2865-01A - 1		29873	5.0868	-	7471B	PPB	19 May 2022 09:10:53	U SMPL
CCV68 - 1	CCV68	30544	5.2002	-	7471B	PPB	19 May 2022 09:13:09	U SMPL
CCB68 - 1	CCB68	-59	0.0278	-	7471B	PPB	19 May 2022 09:15:25	U SMPL