



Analytical Summary Report

Analysis Method: 365.3

ANALYST: Iwona

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: jignesh

Run Number: LB120249

Reagent/Standard	Lot/Log #
calibration std. phosphate 1 ppm	WP96818
calibration std. phosphate 0.5 ppm	WP96817
calibration std. phosphate 0.3 ppm	WP96816
calibration std. phosphate 0.1 ppm	WP96815
calibration std. phosphate 0.05 ppm	WP96814
calibration std. 0 ppm	WP96813
phosphate CCV std.	WP96819
Combined reagent	WP96860
Phenolphthalein indicator	WP96861
Sodium hydroxide, 1N	WP95405
Phosphate ICV-LCS Std	WP96820

Intercept: -0.0014

Slope: 0.6505

Regression: 0.999956

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	%D	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	0.002		05/19/2022	13:30
2	CAL2	0.05	1	50	50	0.033	0.053	-6	05/19/2022	13:30
3	CAL3	0.10	1	50	50	0.064	0.101	-1	05/19/2022	13:31
4	CAL4	0.30	1	50	50	0.190	0.294	2	05/19/2022	13:31
5	CAL5	0.50	1	50	50	0.322	0.497	0.6	05/19/2022	13:32
6	CAL6	1.00	1	50	50	0.651	1.003	-0.3	05/19/2022	13:32



Analytical Summary Report

Analysis Method: 365.3

ANALYST: Iwona

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: jignesh

Run Number: LB120249

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.322	0.497	05/19/2022	13:33
2	ICB		1	50	50	0.001	0.004	05/19/2022	13:33
3	CCV1	0.50	1	50	50	0.328	0.506	05/19/2022	13:34
4	CCB1		1	50	50	0.002	0.005	05/19/2022	13:34
5	PB144927BL		1	50	50	0.001	0.004	05/19/2022	13:35
6	PB144927BS	0.50	1	50	50	0.320	0.494	05/19/2022	13:35
7	N2756-02		1	50	50	0.199	0.308	05/19/2022	13:36
8	N2756-02DUP		1	50	50	0.200	0.310	05/19/2022	13:36
9	N2756-02MS	0.50	1	50	50	0.478	0.737	05/19/2022	13:37
10	N2756-02MSD	0.50	1	50	50	0.479	0.739	05/19/2022	13:37
11	N2756-03		1	50	50	0.258	0.399	05/19/2022	13:38
12	N2757-02		1	50	50	0.581	0.895	05/19/2022	13:38
13	N2757-07		1	50	50	0.367	0.566	05/19/2022	13:39
14	N2786-01		1	50	50	0.718	1.106	05/19/2022	13:39
15	CCV2	0.50	1	50	50	0.326	0.503	05/19/2022	13:40
16	CCB2		1	50	50	0.003	0.007	05/19/2022	13:40
17	N2788-01		1	50	50	0.566	0.872	05/19/2022	13:41
18	N2793-01		1	50	50	0.001	0.004	05/19/2022	13:41
19	N2793-02		1	50	50	0.004	0.008	05/19/2022	13:42
20	N2793-03		1	50	50	0.135	0.210	05/19/2022	13:42
21	N2793-04		1	50	50	0.015	0.025	05/19/2022	13:43
22	N2903-01		1	50	50	1.201	1.848	05/19/2022	13:43
23	N2786-01		5	50	50	0.210	0.325	05/19/2022	13:44
24	N2903-01		10	50	50	0.274	0.423	05/19/2022	13:44
25	CCV3	0.50	1	50	50	0.322	0.497	05/19/2022	13:45
26	CCB3		1	50	50	0.002	0.005	05/19/2022	13:45