


LB 120250

245.1

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	1/1 D
0.0	0.000	-0.001	-0.001	250	0.000	250					30
0.05	0.050	0.065	0.015	515	0.0 %	515	0				4
0.2	0.200	0.209	0.009	1085	0.0 %	1085					4
2.5	2.500	2.395	-0.105	9751	0.0 %	9751					4
5.0	5.000	5.176	0.176	20777	0.0 %	20777					1
7.5	7.500	7.391	-0.109	29558	0.0 %	29558					0
10.0	10.000	10.014	0.014	39954	0.0 %	39954					

LB120250 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 19 May 2022 12:04:41

Sample ID	Extended ID	µ Abs.	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50	250	-	0.0000245.1	PPB	19 May 2022 12:08:48	S	Std
0.05 - 1	50.05	515	-	0.0500245.1	PPB	19 May 2022 12:11:05	S	Std
0.2 - 1	50.2	1085	-	0.2000245.1	PPB	19 May 2022 12:13:21	S	Std
2.5 - 1	52.5	9751	-	2.5000245.1	PPB	19 May 2022 12:15:38	S	Std
5.0 - 1	55	20777	-	5.0000245.1	PPB	19 May 2022 12:17:54	S	Std
7.5 - 1	57.5	29558	-	7.5000245.1	PPB	19 May 2022 12:20:10	S	Std
10.0 - 1	516	39954	-	10.0000245.1	PPB	19 May 2022 12:22:25	S	Std
ICV29 - 1	ICV29	15859	3.9359	-245.1	PPB	19 May 2022 12:27:50	U	SMPL
ICB29 - 1	ICB29	263	0.0019	-245.1	PPB	19 May 2022 12:30:04	U	SMPL
CCV69 - 1	CCV69	20701	5.1573	-245.1	PPB	19 May 2022 12:32:21	U	SMPL
CCB69 - 1	CCB69	268	0.0031	-245.1	PPB	19 May 2022 12:34:35	U	SMPL
CRA - 1	CRA	509	0.0639	-245.1	PPB	19 May 2022 12:36:50	U	SMPL
HighStd - 1	HighStd	39860	9.9900	-245.1	PPB	19 May 2022 12:39:05	U	SMPL
ChkStd - 1	ChkStd	26721	6.6758	-245.1	PPB	19 May 2022 12:41:20	U	SMPL
PB144947BL - 1	PBW	-19	-0.0693	-245.1	PPB	19 May 2022 12:43:37	U	SMPL
PB144947BS - 1	LCSW	14673	3.6367	-245.1	PPB	19 May 2022 12:45:53	U	SMPL
N2833-01 - 1	TWP02	1323	0.2693	-245.1	PPB	19 May 2022 12:48:09	U	SMPL
N2833-01DUP - 1	TWP02DUP	1325	0.2698	-245.1	PPB	19 May 2022 12:50:26	U	SMPL
N2833-01MS - 1	TWP02MS	17076	4.2429	-245.1	PPB	19 May 2022 12:52:43	U	SMPL
N2833-01MSD - 1	TWP02MSD	16976	4.2176	-245.1	PPB	19 May 2022 12:55:00	U	SMPL
N2833-02 - 1	TWP03	15573	3.8637	-245.1	PPB	19 May 2022 12:57:15	U	SMPL
CCV70 - 1	CCV70	21025	5.2390	-245.1	PPB	19 May 2022 12:59:30	U	SMPL
CCB70 - 1	CCB70	265	0.0024	-245.1	PPB	19 May 2022 13:01:46	U	SMPL
N2852-03 - 1	CBH-594	376	0.0304	-245.1	PPB	19 May 2022 13:04:01	U	SMPL
N2852-04 - 1	MH-725	51	-0.0516	-245.1	PPB	19 May 2022 13:06:17	U	SMPL
N2936-01 - 1	SEMI-ANNUAL-CONDENSATE	2842	0.6524	-245.1	PPB	19 May 2022 13:08:32	U	SMPL
N2833-01LX5 - 1		422	0.0420	-245.1	PPB	19 May 2022 13:10:48	U	SMPL
N2833-01A - 1		16977	4.2179	-245.1	PPB	19 May 2022 13:13:04	U	SMPL
CCV71 - 1	CCV71	20781	5.1774	-245.1	PPB	19 May 2022 13:15:20	U	SMPL
CCB71 - 1	CCB71	244	-0.0029	-245.1	PPB	19 May 2022 13:17:36	U	SMPL