


LB120268

7471B

INSTRUMENT ID: CV1

 μ Abs.:

54798



A= 0.0000e+000

B= 1.8295e-004

C= 2.4916e-002

Rho= 0.9998886

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	Yob
0.0	0.000	0.032	0.032	36	0.000	36					-
0.2	0.200	0.237	0.037	1159	0.0 %	1159					19
2.5	2.500	2.485	-0.015	13444	0.0 %	13444					-1
5.0	5.000	4.886	-0.114	26571	0.0 %	26571					-2
7.5	7.500	7.510	0.010	40915	0.0 %	40915					0
10.0	10.000	10.050	0.050	54798	0.0 %	54798					1

LB120268 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 20 May 2022 08:32:13

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	5.0	36	-	0.00007471B	PPB		20 May 2022 08:46:02	S	Std
0.2 - 1	5.2	1159	-	0.20007471B	PPB		20 May 2022 08:48:18	S	Std
2.5 - 1	5.25	13444	-	2.50007471B	PPB		20 May 2022 08:50:35	S	Std
5.0 - 1	5.5	26571	-	5.00007471B	PPB		20 May 2022 08:52:52	S	Std
7.5 - 1	5.75	40915	-	7.50007471B	PPB		20 May 2022 08:55:08	S	Std
10.0 - 1	6.0	54798	-	10.00007471B	PPB		20 May 2022 08:57:25	S	Std
ICV32 - 1	ICV32	21719	3.9985	-7471B	PPB		20 May 2022 09:02:48	U	SMPL
ICB32 - 1	ICB32	10	0.0267	-7471B	PPB		20 May 2022 09:05:03	U	SMPL
CCV79 - 1	CCV79	28794	5.2929	-7471B	PPB		20 May 2022 09:07:19	U	SMPL
CCB79 - 1	CCB79	12	0.0271	-7471B	PPB		20 May 2022 09:09:34	U	SMPL
CRA - 1	CRA	1137	0.2329	-7471B	PPB		20 May 2022 09:11:50	U	SMPL
HighStd - 1	HighStd	55117	10.1087	-7471B	PPB		20 May 2022 09:14:04	U	SMPL
ChkStd - 1	ChkStd	37968	6.9713	-7471B	PPB		20 May 2022 09:16:19	U	SMPL
PB144925BL - 1	PBS	-44	0.0169	-7471B	PPB		20 May 2022 09:18:37	U	SMPL
PB144925BS - 1	LCSS	23982	4.4125	-7471B	PPB		20 May 2022 09:20:54	U	SMPL
N2871-01 - 1	P017-SS001-0006-01	7652	1.4249	-7471B	PPB		20 May 2022 09:23:10	U	SMPL
N2871-02 - 1	P017-SS001-0612-01	3304	0.6294	-7471B	PPB		20 May 2022 09:25:27	U	SMPL
N2871-03 - 1	P017-SS001-1218-01	2702	0.5193	-7471B	PPB		20 May 2022 09:27:44	U	SMPL
N2871-04 - 1	P017-SS001-1824-01	2264	0.4391	-7471B	PPB		20 May 2022 09:30:01	U	SMPL
N2871-05 - 1	P017-SS001-2430-01	7146	1.3323	-7471B	PPB		20 May 2022 09:32:16	U	SMPL
CCV80 - 1	CCV80	27031	4.9703	-7471B	PPB		20 May 2022 09:34:31	U	SMPL
CCB80 - 1	CCB80	49	0.0339	-7471B	PPB		20 May 2022 09:36:46	U	SMPL
N2871-06 - 1	P017-SS002-0006-01	2396	0.4633	-7471B	PPB		20 May 2022 09:39:02	U	SMPL
N2871-11 - 1	P017-SS003-0006-01	3870	0.7329	-7471B	PPB		20 May 2022 09:41:17	U	SMPL
N2871-16 - 1	P017-SS004-0006-01	24907	4.5817	-7471B	PPB		20 May 2022 09:43:32	U	SMPL
N2871-24 - 1	P017-SS005-0006-01	6758	1.2613	-7471B	PPB		20 May 2022 09:45:48	U	SMPL
N2872-05 - 1	P017-SS006-0006-01	4533	0.8542	-7471B	PPB		20 May 2022 09:48:05	U	SMPL
N2872-10 - 1	P022-SS001-0006-01	3755	0.7119	-7471B	PPB		20 May 2022 09:50:21	U	SMPL
N2872-11 - 1	P022-SS001-0612-01	5004	0.9404	-7471B	PPB		20 May 2022 09:52:37	U	SMPL
N2872-12 - 1	P022-SS001-1218-01	3821	0.7240	-7471B	PPB		20 May 2022 09:54:54	U	SMPL
N2872-13 - 1	P022-SS001-1824-01	2752	0.5284	-7471B	PPB		20 May 2022 09:57:10	U	SMPL
N2872-14 - 1	P022-SS001-2430-01	2703	0.5194	-7471B	PPB		20 May 2022 09:59:26	U	SMPL
CCV81 - 1	CCV81	27366	5.0316	-7471B	PPB		20 May 2022 10:01:41	U	SMPL
CCB81 - 1	CCB81	32	0.0308	-7471B	PPB		20 May 2022 10:03:56	U	SMPL
N2872-15 - 1	P022-SS002-0006-01	4073	0.7701	-7471B	PPB		20 May 2022 10:06:14	U	SMPL
N2872-15DUP - 1	P022-SS002-0006-01DUP	4054	0.7666	-7471B	PPB		20 May 2022 10:08:29	U	SMPL
N2872-16 - 1	N2872-15MS	26329	4.8419	-7471B	PPB		20 May 2022 10:10:44	U	SMPL
N2872-17 - 1	N2872-15MSD	24120	4.4377	-7471B	PPB		20 May 2022 10:13:00	U	SMPL
N2872-18 - 1	P022-SS002-0006-02	4580	0.8628	-7471B	PPB		20 May 2022 10:15:17	U	SMPL
N2872-23 - 1	P029-SS001-0006-01	3987	0.7544	-7471B	PPB		20 May 2022 10:17:35	U	SMPL
PB144926BL - 1	PBS	13	0.0273	-7471B	PPB		20 May 2022 10:19:51	U	SMPL
PB144926BS - 1	LCSS	21341	3.9293	-7471B	PPB		20 May 2022 10:22:08	U	SMPL
N2873-01 - 1	P029-SS001-0612-01	2947	0.5641	-7471B	PPB		20 May 2022 10:24:24	U	SMPL
N2873-02 - 1	P029-SS001-1218-01	2659	0.5114	-7471B	PPB		20 May 2022 10:26:40	U	SMPL
CCV82 - 1	CCV82	27337	5.0263	-7471B	PPB		20 May 2022 10:28:55	U	SMPL
CCB82 - 1	CCB82	-60	0.0139	-7471B	PPB		20 May 2022 10:31:11	U	SMPL
N2873-03 - 1	P029-SS001-1824-01	2370	0.4585	-7471B	PPB		20 May 2022 10:33:28	U	SMPL
N2873-04 - 1	P029-SS001-2430-01	2139	0.4163	-7471B	PPB		20 May 2022 10:40:23	U	SMPL
N2873-05 - 1	P031-SS001-0006-01	2832	0.5430	-7471B	PPB		20 May 2022 10:42:40	U	SMPL
N2873-06 - 1	P031-SS001-0612-01	2254	0.4373	-7471B	PPB		20 May 2022 10:44:55	U	SMPL
N2873-07 - 1	P031-SS001-1218-01	1665	0.3295	-7471B	PPB		20 May 2022 10:47:12	U	SMPL
N2873-08 - 1	P031-SS001-1824-01	2094	0.4080	-7471B	PPB		20 May 2022 10:49:28	U	SMPL
N2873-09 - 1	P031-SS001-2430-01	3177	0.6062	-7471B	PPB		20 May 2022 10:51:44	U	SMPL
N2873-10 - 1	P032-SS001-0006-01	3089	0.5901	-7471B	PPB		20 May 2022 10:54:01	U	SMPL
N2873-11 - 1	P032-SS001-0612-01	1282	0.2595	-7471B	PPB		20 May 2022 10:56:18	U	SMPL
N2873-12 - 1	P032-SS001-1218-01	2919	0.5590	-7471B	PPB		20 May 2022 10:58:33	U	SMPL
CCV83 - 1	CCV83	27705	5.0936	-7471B	PPB		20 May 2022 11:00:48	U	SMPL
CCB83 - 1	CCB83	-69	0.0123	-7471B	PPB		20 May 2022 11:03:04	U	SMPL
N2873-13 - 1	P032-SS001-1824-01	2916	0.5584	-7471B	PPB		20 May 2022 11:05:21	U	SMPL
N2873-14 - 1	P032-SS001-2430-01	2411	0.4660	-7471B	PPB		20 May 2022 11:07:37	U	SMPL
N2873-14DUP - 1	P032-SS001-2430-01DUP	2402	0.4644	-7471B	PPB		20 May 2022 11:12:37	U	SMPL
N2873-14MS - 1	P032-SS001-2430-01MS	21895	4.0307	-7471B	PPB		20 May 2022 11:14:52	U	SMPL
N2873-14MSD - 1	P032-SS001-2430-01MSD	22327	4.1097	-7471B	PPB		20 May 2022 11:17:07	U	SMPL
N2875-01 - 1	P007-SS006-0006-01	1261	0.2556	-7471B	PPB		20 May 2022 11:19:24	U	SMPL
N2875-06 - 1	P007-SS007-0006-01	1620	0.3213	-7471B	PPB		20 May 2022 11:21:42	U	SMPL
N2875-11 - 1	P007-SS008-0006-01	264	0.0732	-7471B	PPB		20 May 2022 11:23:59	U	SMPL
N2875-16 - 1	P007-SS009-0006-01	7037	1.3124	-7471B	PPB		20 May 2022 11:26:15	U	SMPL
N2875-19 - 1	P007-SS010-0006-01	2444	0.4721	-7471B	PPB		20 May 2022 11:28:31	U	SMPL
CCV84 - 1	CCV84	27674	5.0880	-7471B	PPB		20 May 2022 11:30:47	U	SMPL
CCB84 - 1	CCB84	-72	0.0117	-7471B	PPB		20 May 2022 11:33:03	U	SMPL
PB144928BL - 1	PBS	42	0.0326	-7471B	PPB		20 May 2022 11:35:20	U	SMPL
PB144928BS - 1	LCSS	22199	4.0863	-7471B	PPB		20 May 2022 11:40:11	U	SMPL
N2874-01 - 1	P007-SS002-0006-01	3915	0.7412	-7471B	PPB		20 May 2022 11:42:27	U	SMPL
N2874-02 - 1	P007-SS002-0612-01	4741	0.8923	-7471B	PPB		20 May 2022 11:44:43	U	SMPL
N2874-03 - 1	P007-SS002-1218-01	4762	0.8961	-7471B	PPB		20 May 2022 11:46:59	U	SMPL
N2874-06 - 1	P007-SS002-1824-01	1684	0.3330	-7471B	PPB		20 May 2022 11:49:15	U	SMPL
N2874-06DUP - 1	P007-SS002-1824-01DUP	1531	0.3050	-7471B	PPB		20 May 2022 11:51:31	U	SMPL
N2874-07 - 1	N2874-06MS	23297	4.2872	-7471B	PPB		20 May 2022 11:53:48	U	SMPL
N2874-08 - 1	N2874-06MSD	22016	4.0528	-7471B	PPB		20 May 2022 11:56:05	U	SMPL
N2874-09 - 1	P007-SS002-1824-02	1420	0.2847	-7471B	PPB		20 May 2022 11:58:22	U	SMPL
CCV85 - 1	CCV85	28066	5.1597	-7471B	PPB		20 May 2022 12:00:39	U	SMPL

LB120268 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 20 May 2022 08:32:13

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
CCB85 - 1	CCB85	-79	0.0105	-7471B	PPB		20 May 2022 12:02:55	U	SMPL
N2874-10 - 1	P007-SS002-2430-01	1285	0.2600	-7471B	PPB		20 May 2022 12:05:13	U	SMPL
N2874-11 - 1	P007-SS003-0006-01	2332	0.4516	-7471B	PPB		20 May 2022 12:07:29	U	SMPL
N2874-16 - 1	P007-SS004-0006-01	8122	1.5109	-7471B	PPB		20 May 2022 12:09:45	U	SMPL
N2874-21 - 1	P007-SS005-0006-01	8130	1.5123	-7471B	PPB		20 May 2022 12:12:01	U	SMPL
N2877-02 - 1	P007-SS015-0006-01	2691	0.5172	-7471B	PPB		20 May 2022 12:14:17	U	SMPL
N2877-07 - 1	P007-SS016-0006-01	4455	0.8400	-7471B	PPB		20 May 2022 12:16:32	U	SMPL
N2877-12 - 1	P007-SS017-0006-01	13276	2.4538	-7471B	PPB		20 May 2022 12:18:49	U	SMPL
N2877-17 - 1	P007-SS018-0006-01	1044	0.2159	-7471B	PPB		20 May 2022 12:21:06	U	SMPL
N2878-01 - 1	P007-SS019-0006-01	3470	0.6598	-7471B	PPB		20 May 2022 12:23:22	U	SMPL
N2878-06 - 1	P007-SS020-0006-01	2590	0.4988	-7471B	PPB		20 May 2022 12:25:39	U	SMPL
CCV86 - 1	CCV86	28526	5.2438	-7471B	PPB		20 May 2022 12:27:54	U	SMPL
CCB86 - 1	CCB86	-78	0.0106	-7471B	PPB		20 May 2022 12:30:10	U	SMPL
N2878-11 - 1	P007-SS021-0006-01	1175	0.2399	-7471B	PPB		20 May 2022 12:32:28	U	SMPL
N2878-13 - 1	P007-SS022-0006-01	1981	0.3873	-7471B	PPB		20 May 2022 12:34:44	U	SMPL
PB144965BL - 1	PBS	-27	0.0200	-7471B	PPB		20 May 2022 12:36:59	U	SMPL
PB144965BS - 1	LCSS	19880	3.6620	-7471B	PPB		20 May 2022 12:39:15	U	SMPL
N2876-04 - 1	P007-SS011-0006-01	2468	0.4764	-7471B	PPB		20 May 2022 12:41:31	U	SMPL
N2876-09 - 1	P007-SS012-0006-01	1442	0.2887	-7471B	PPB		20 May 2022 12:43:48	U	SMPL
N2876-14 - 1	P007-SS013-0006-01	1287	0.2604	-7471B	PPB		20 May 2022 12:46:05	U	SMPL
N2876-21 - 1	P007-SS014-0006-01	1136	0.2328	-7471B	PPB		20 May 2022 12:48:21	U	SMPL
N2884-01 - 1	31490	129163	23.6557	-7471B	PPB		20 May 2022 12:50:38	U	SMPL
N2884-02 - 1	30887	316	0.0827	-7471B	PPB		20 May 2022 12:52:55	U	SMPL
CCV87 - 1	CCV87	28107	5.1672	-7471B	PPB		20 May 2022 12:55:15	U	SMPL
CCB87 - 1	CCB87	-70	0.0121	-7471B	PPB		20 May 2022 12:57:31	U	SMPL
N2900-01 - 1	SU-02-051722	1988	0.3886	-7471B	PPB		20 May 2022 12:59:48	U	SMPL
N2900-01DUP - 1	SU-02-051722DUP	1793	0.3530	-7471B	PPB		20 May 2022 13:02:05	U	SMPL
N2900-01MS - 1	SU-02-051722MS	22036	4.0565	-7471B	PPB		20 May 2022 13:04:22	U	SMPL
N2900-01MSD - 1	SU-02-051722MSD	24281	4.4672	-7471B	PPB		20 May 2022 13:06:38	U	SMPL
N2900-03 - 1	SU-04-051722	835	0.1777	-7471B	PPB		20 May 2022 13:08:55	U	SMPL
N2901-01 - 1	EO-02-051722	6449	1.2048	-7471B	PPB		20 May 2022 13:11:13	U	SMPL
N2935-01 - 1	OK-03-051822	2655	0.5107	-7471B	PPB		20 May 2022 13:13:29	U	SMPL
N2940-02 - 1	CL-1-051922	3261	0.6215	-7471B	PPB		20 May 2022 13:15:46	U	SMPL
N2959-02 - 1	284	-18	0.0216	-7471B	PPB		20 May 2022 13:18:03	U	SMPL
N2960-01 - 1	BIN-4-SP-1	1612	0.3198	-7471B	PPB		20 May 2022 13:20:19	U	SMPL
CCV88 - 1	CCV88	28090	5.1641	-7471B	PPB		20 May 2022 13:22:35	U	SMPL
CCB88 - 1	CCB88	-69	0.0123	-7471B	PPB		20 May 2022 13:24:52	U	SMPL
N2960-03 - 1	BIN-4-SP-2	1683	0.3328	-7471B	PPB		20 May 2022 13:27:10	U	SMPL
N2963-01 - 1	RBR-057	1176	0.2401	-7471B	PPB		20 May 2022 13:29:26	U	SMPL
N2963-03 - 1	VNJ-208	1213	0.2468	-7471B	PPB		20 May 2022 13:31:43	U	SMPL
N2964-01 - 1	EO-01-051922	653	0.1444	-7471B	PPB		20 May 2022 13:34:00	U	SMPL
N2965-02 - 1	16972	29	0.0302	-7471B	PPB		20 May 2022 13:36:16	U	SMPL
N2884-01DLX5 - 1		27395	5.0369	-7471B	PPB		20 May 2022 13:38:33	U	SMPL
CCV89 - 1	CCV89	28098	5.1655	-7471B	PPB		20 May 2022 13:40:49	U	SMPL
CCB89 - 1	CCB89	-46	0.0165	-7471B	PPB		20 May 2022 13:43:07	U	SMPL
N2872-16LX5 - 1		773	0.1663	-7471B	PPB		20 May 2022 13:45:26	U	SMPL
N2872-16A - 1		26399	4.8547	-7471B	PPB		20 May 2022 13:47:42	U	SMPL
N2873-14A - 1		24969	4.5931	-7471B	PPB		20 May 2022 13:52:17	U	SMPL
N2874-06LX5 - 1		284	0.0769	-7471B	PPB		20 May 2022 13:54:33	U	SMPL
N2900-01LX5 - 1		287	0.0774	-7471B	PPB		20 May 2022 13:59:08	U	SMPL
N2900-01A - 1		28213	5.1866	-7471B	PPB		20 May 2022 14:01:26	U	SMPL
N2873-14LX5 - 1		404	0.0988	-7471B	PPB		20 May 2022 14:10:20	U	SMPL
N2874-06A - 1		23141	4.2586	-7471B	PPB		20 May 2022 14:13:02	U	SMPL
CCV90 - 1	CCV90	29615	5.4431	-7471B	PPB		20 May 2022 14:21:28	U	SMPL
CCB90 - 1	CCB90	-96	0.0074	-7471B	PPB		20 May 2022 14:23:45	U	SMPL