



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : mohammad

Run Number: LB120270

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	05/20/2022	08:15
2	CAL2	1	Water	NA	NA	20.3	7.00	05/20/2022	08:16
3	CAL3	1	Water	NA	NA	20.3	10.03	05/20/2022	08:19
4	ICV	1	Water	NA	NA	20.5	6.99	05/20/2022	08:20
5	CCV1	1	Water	NA	NA	20.5	2.01	05/20/2022	08:22
6	N2942-01	1	Solid	20.03	20	22.1	8.66	05/20/2022	08:25
7	N2965-01	1	Solid	20.02	20	20.3	7.49	05/20/2022	08:30
8	N2965-01DUP	1	Solid	20.03	20	20.4	7.50	05/20/2022	08:31
9	CCV2	1	Water	NA	NA	20.5	12.03	05/20/2022	08:35

VB 120270

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph n2954 s WorkList ID : 159808 Department : Wet-Chemistry Date : 05-20-2022 08:10:05

Due Date	Matrix	Sample	Test	Raw Sample			Collect Date	Method
				Preservative	Customer	Storage Location		
05/18/2022	Solid	N2942-01	pH	Cool 4 deg C	NEWP02	M11	RATZER-RD-SP	05/18/2022 9045D
05/19/2022	Solid	N2965-01	pH	Cool 4 deg C	PSEG03	M11	30985	05/19/2022 9045D

Date/Time 05/10/22 08:12
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 05/20/22 11:00
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]