

LB120281


SFAM01.1

INSTRUMENT ID: CV2

Linear

μ Abs.:

52800



A= 0.0000e+000

B= 1.9329e-004 slope

C= -5.2109e-002 $\approx 10\%$ for cv

Rho= 0.9995369

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%b
0.00	0.000	0.031	0.031	429	0.000	429					-
0.05	0.050					185	0				-2
0.20	0.200	0.196	-0.004	1286	0.0 %	1286					4
2.50	2.500	2.589	0.089	13662	0.0 %	13662					-2
5.00	5.000	4.924	-0.076	25747	0.0 %	25747					-3
7.50	7.500	7.306	-0.194	38069	0.0 %	38069					2
10.0	10.000	10.153	0.153	52800	0.0 %	52800					

LB120281
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	429	-		0 SFAM01.1	PPB	5/20/2022 15:05	Std
	0.2 S01	1286	-		0.2 SFAM01.1	PPB	5/20/2022 15:09	Std
	2.5 S02	13662	-		2.5 SFAM01.1	PPB	5/20/2022 15:11	Std
	5 S03	25747	-		5 SFAM01.1	PPB	5/20/2022 15:16	Std
	7.5 S04	38069	-		7.5 SFAM01.1	PPB	5/20/2022 15:21	Std
	10 S05	52800	-		10 SFAM01.1	PPB	5/20/2022 15:26	Std
ICV033	ICV033	21657	4.1339	-	SFAM01.1	PPB	5/20/2022 15:31	SMPL
ICB033	ICB033	564	0.0569	-	SFAM01.1	PPB	5/20/2022 15:33	SMPL
CCV091	CCV091	25867	4.9477	-	SFAM01.1	PPB	5/20/2022 15:36	SMPL
CCB091	CCB091	556	0.0554	-	SFAM01.1	PPB	5/20/2022 15:38	SMPL
PB144976BL	PBS9766	31	-0.0461	-	SFAM01.1	PPB	5/20/2022 15:40	SMPL
N2818-01	YB6R2	74264	14.3022	-	SFAM01.1	PPB	5/20/2022 15:42	SMPL
N2818-02	YB6R3	60940	11.7269	-	SFAM01.1	PPB	5/20/2022 15:45	SMPL
N2818-03	YB6R4	3654	0.6542	-	SFAM01.1	PPB	5/20/2022 15:47	SMPL
N2818-04	YB6R5	3080	0.5432	-	SFAM01.1	PPB	5/20/2022 15:49	SMPL
N2818-05	YB6R6	16217	3.0824	-	SFAM01.1	PPB	5/20/2022 15:52	SMPL
N2818-06	YB6R7	59982	11.5417	-	SFAM01.1	PPB	5/20/2022 15:54	SMPL
N2818-07	YB6R8	7615	1.4198	-	SFAM01.1	PPB	5/20/2022 15:56	SMPL
N2818-08	YB6R9	1243	0.1881	-	SFAM01.1	PPB	5/20/2022 15:58	SMPL
N2818-09	YB6S0	2950	0.5181	-	SFAM01.1	PPB	5/20/2022 16:01	SMPL
N2818-10	YB6S1	1292	0.1976	-	SFAM01.1	PPB	5/20/2022 16:03	SMPL
N2818-11	YB6T2	1443	0.2268	-	SFAM01.1	PPB	5/20/2022 16:05	SMPL
N2818-12	YB6T3	180638	34.863	-	SFAM01.1	PPB	5/20/2022 16:07	SMPL
N2818-13	YB6T3D	183072	35.3335	-	SFAM01.1	PPB	5/20/2022 16:10	SMPL
N2818-14	YB6T3S	196156	37.8625	-	SFAM01.1	PPB	5/20/2022 16:12	SMPL
N2818-15	YB6T4	160050	30.8836	-	SFAM01.1	PPB	5/20/2022 16:14	SMPL
N2818-16	YB6T5	21976	4.1956	-	SFAM01.1	PPB	5/20/2022 16:17	SMPL
N2818-17	YB6T6	6946	1.2905	-	SFAM01.1	PPB	5/20/2022 16:19	SMPL
N2818-18	YB6T7	7339	1.3664	-	SFAM01.1	PPB	5/20/2022 16:21	SMPL
N2818-19	YB6T8	15389	2.9224	-	SFAM01.1	PPB	5/20/2022 16:23	SMPL
CCV092	CCV092	25371	4.8518	-	SFAM01.1	PPB	5/20/2022 16:26	SMPL
CCB092	CCB092	458	0.0364	-	SFAM01.1	PPB	5/20/2022 16:28	SMPL
N2818-20	YB6T9	4413	0.8009	-	SFAM01.1	PPB	5/20/2022 16:30	SMPL
N2818-21	YB6W0	25492	4.8752	-	SFAM01.1	PPB	5/20/2022 16:33	SMPL
N2818-22	YB6W1	7065	1.3135	-	SFAM01.1	PPB	5/20/2022 16:35	SMPL
PB144977BL	PBS977	0	-0.0521	-	SFAM01.1	PPB	5/20/2022 16:37	SMPL
N2819-01	YB6S2	2629	0.456	-	SFAM01.1	PPB	5/20/2022 16:39	SMPL
N2819-02	YB6S3	1601	0.2573	-	SFAM01.1	PPB	5/20/2022 16:42	SMPL
N2819-03	YB6S4	1725	0.2813	-	SFAM01.1	PPB	5/20/2022 16:44	SMPL
N2819-04	YB6S5	3669	0.6571	-	SFAM01.1	PPB	5/20/2022 16:46	SMPL
N2819-05	YB6S6	3991	0.7193	-	SFAM01.1	PPB	5/20/2022 16:48	SMPL
N2819-06	YB6S7	2681	0.4661	-	SFAM01.1	PPB	5/20/2022 16:51	SMPL
N2819-07	YB6S8	4279	0.775	-	SFAM01.1	PPB	5/20/2022 16:53	SMPL
N2819-08	YB6S9	3148	0.5564	-	SFAM01.1	PPB	5/20/2022 16:55	SMPL
N2819-09	YB6T0	1515	0.2407	-	SFAM01.1	PPB	5/20/2022 16:57	SMPL
N2819-10	YB6T1	2826	0.4941	-	SFAM01.1	PPB	5/20/2022 17:00	SMPL

LB120281
INSTRUMENT ID : CV1

N2819-11	YB6W2	3146	0.556 -	SFAM01.1	PPB	5/20/2022 17:02	SMPL
N2819-12	YB6W3	16100	3.0598 -	SFAM01.1	PPB	5/20/2022 17:04	SMPL
N2819-13	YB6W4	16769	3.1891 -	SFAM01.1	PPB	5/20/2022 17:06	SMPL
N2819-14	YB6W5	266920	51.5403 -	SFAM01.1	PPB	5/20/2022 17:09	SMPL
N2819-15	YB6W5D	265664	51.2975 -	SFAM01.1	PPB	5/20/2022 17:11	SMPL
N2819-16	YB6W5S	272211	52.563 -	SFAM01.1	PPB	5/20/2022 17:13	SMPL
CCV093	CCV093	24832	4.7476 -	SFAM01.1	PPB	5/20/2022 17:16	SMPL
CCB093	CCB093	261	-0.0017 -	SFAM01.1	PPB	5/20/2022 17:18	SMPL
N2819-17	YB6W6	6771	1.2566 -	SFAM01.1	PPB	5/20/2022 17:20	SMPL
N2819-18	YB6W7	309603	59.7904 -	SFAM01.1	PPB	5/20/2022 17:23	SMPL
N2819-19	YB6W8	301054	58.138 -	SFAM01.1	PPB	5/20/2022 17:25	SMPL
N2819-20	YB6W9	18203	3.4663 -	SFAM01.1	PPB	5/20/2022 17:27	SMPL
N2819-21	YB6X0	8841	1.6567 -	SFAM01.1	PPB	5/20/2022 17:30	SMPL
N2819-22	YB6X1	1162	0.1725 -	SFAM01.1	PPB	5/20/2022 17:32	SMPL
PB144978BL	PBS978	517	0.0478 -	SFAM01.1	PPB	5/20/2022 17:34	SMPL
N2840-01	YB6X2	62349	11.9992 -	SFAM01.1	PPB	5/20/2022 17:37	SMPL
N2840-02	YB6X3	13996	2.6531 -	SFAM01.1	PPB	5/20/2022 17:39	SMPL
N2840-03	YB6X4	354525	HIGH -	SFAM01.1	PPB	5/20/2022 17:41	
N2840-04	YB6X5	334782	HIGH -	SFAM01.1	PPB	5/20/2022 17:43	
N2840-05	YB6X6	4292	0.7775 -	SFAM01.1	PPB	5/20/2022 17:45	SMPL
N2840-06	YB6X7	1916	0.3182 -	SFAM01.1	PPB	5/20/2022 17:47	SMPL
N2840-07	YB6X8	11603	2.1906 -	SFAM01.1	PPB	5/20/2022 17:50	SMPL
N2840-08	YB6X9	11297	2.1315 -	SFAM01.1	PPB	5/20/2022 17:52	SMPL
N2840-09	YB6Y0	5861	1.0808 -	SFAM01.1	PPB	5/20/2022 17:54	SMPL
N2840-10	YB6Y1	5629	1.0359 -	SFAM01.1	PPB	5/20/2022 17:56	SMPL
N2840-11	YB6Y2	6829	1.2679 -	SFAM01.1	PPB	5/20/2022 17:59	SMPL
N2840-12	YB6Y3	8786	1.6461 -	SFAM01.1	PPB	5/20/2022 18:01	SMPL
N2840-13	YB6Y4	10282	1.9353 -	SFAM01.1	PPB	5/20/2022 18:03	SMPL
CCV094	CCV094	25053	4.7903 -	SFAM01.1	PPB	5/20/2022 18:05	SMPL
CCB094	CCB094	322	0.0101 -	SFAM01.1	PPB	5/20/2022 18:08	SMPL
N2840-14	YB6Y5	76966	14.8245 -	SFAM01.1	PPB	5/20/2022 18:10	SMPL
N2840-15	YB6Y6	87783	16.9153 -	SFAM01.1	PPB	5/20/2022 18:12	SMPL
N2840-16	YB6Y7	23413	4.4733 -	SFAM01.1	PPB	5/20/2022 18:15	SMPL
N2840-17	YB6Y8	12599	2.3831 -	SFAM01.1	PPB	5/20/2022 18:17	SMPL
N2840-18	YB6Y9	3360	0.5973 -	SFAM01.1	PPB	5/20/2022 18:19	SMPL
N2840-19	YB6Y9D	3340	0.5935 -	SFAM01.1	PPB	5/20/2022 18:21	SMPL
N2840-20	YB6Y9S	15363	2.9174 -	SFAM01.1	PPB	5/20/2022 18:24	SMPL
N2840-21	YB6Z0	8502	1.5912 -	SFAM01.1	PPB	5/20/2022 18:26	SMPL
N2840-22	YB6Z1	40142	7.7069 -	SFAM01.1	PPB	5/20/2022 18:28	SMPL
N2818-01DLX5	YB6R2	14158	2.6845 -	SFAM01.1	PPB	5/20/2022 18:30	SMPL
N2818-02DLX2	YB6R3	29360	5.6228 -	SFAM01.1	PPB	5/20/2022 18:33	SMPL
N2818-06DLX2	YB6R7	30050	5.7562 -	SFAM01.1	PPB	5/20/2022 18:35	SMPL
N2818-12DLX10	YB6T3	18054	3.4375 -	SFAM01.1	PPB	5/20/2022 18:37	SMPL
N2818-13DLX10	YB6T3D	18717	3.5657 -	SFAM01.1	PPB	5/20/2022 18:40	SMPL
N2818-14DLX10	YB6T3S	19626	3.7414 -	SFAM01.1	PPB	5/20/2022 18:42	SMPL
N2818-15DLX5	YB6T4	31378	6.0129 -	SFAM01.1	PPB	5/20/2022 18:44	SMPL
CCV095	CCV095	25428	4.8628 -	SFAM01.1	PPB	5/20/2022 18:46	SMPL

LB120281
INSTRUMENT ID : CV1

CCB095	CCB095	326	0.0109 -	SFAM01.1	PPB	5/20/2022 18:49	SMPL
--------	--------	-----	----------	----------	-----	-----------------	------