



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB120292

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	05/21/2022	09:25
2	N2910-06		1	10	100.00	>212.0	>212.0	05/21/2022	09:40
3	N2910-06DUP		1	10	100.00	>212.0	>212.0	05/21/2022	10:05
4	N2917-07		1	15	100.00	>212.0	>212.0	05/21/2022	10:35
5	N2917-08		1	12	100.00	>212.0	>212.0	05/21/2022	10:55
6	N2917-09		1	14	100.00	>212.0	>212.0	05/21/2022	11:20
7	N2965-03		1	13	100.00	>212.0	>212.0	05/21/2022	11:45
8	N2965-04		1	18	100.00	>212.0	>212.0	05/21/2022	12:05
9	N2965-05		1	20	100.00	>212.0	>212.0	05/21/2022	12:32
10	N2965-06		1	18	100.00	>212.0	>212.0	05/21/2022	12:50
11	N2965-07		1	15	100.00	>212.0	>212.0	05/21/2022	13:15
12	N2965-08		1	17	100.00	>212.0	>212.0	05/21/2022	13:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB120292

WorkList Name : FP51922 WorkList ID : 159800 Department : Wet-Chemistry Date : 05-19-2022 17:25:03

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/24/2022	Water	N2910-06	Flash Point	Cool 4 deg C	PSEG03	K61	T2435-WATER	05/17/2022	1010B
05/25/2022	Water	N2917-07	Flash Point	Cool 4 deg C	PSEG03	K51	COMP-1	05/18/2022	1010B
05/25/2022	Water	N2917-08	Flash Point	Cool 4 deg C	PSEG03	K51	COMP-2-WATER	05/18/2022	1010B
05/25/2022	Water	N2917-09	Flash Point	Cool 4 deg C	PSEG03	K51	COMP-3	05/18/2022	1010B
05/24/2022	Water	N2965-03	Flash Point	Cool 4 deg C	PSEG03	M11	30984	05/19/2022	1010B
05/24/2022	Water	N2965-04	Flash Point	Cool 4 deg C	PSEG03	M11	31508	05/19/2022	1010B
05/24/2022	Water	N2965-05	Flash Point	Cool 4 deg C	PSEG03	M11	31600	05/19/2022	1010B
05/24/2022	Water	N2965-06	Flash Point	Cool 4 deg C	PSEG03	M11	31599	05/19/2022	1010B
05/24/2022	Water	N2965-07	Flash Point	Cool 4 deg C	PSEG03	M11	31608	05/19/2022	1010B
05/24/2022	Water	N2965-08	Flash Point	Cool 4 deg C	PSEG03	M11	31619	05/19/2022	1010B

Date/Time 05/21/22 09:05
 Raw Sample Received by: RS(WC)
 Raw Sample Relinquished by: RJ (EXT-126)

Date/Time 05/21/22 13:45
 Raw Sample Received by: RJ (EXT-126)
 Raw Sample Relinquished by: RS(WC)