

LB 120294


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

44778



A= 0.0000e+000

B= 2.2283e-004 slope

C= -1.7313e-002 7=15 Intercept

Rho= 0.9999108

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	g/vd
0.00	0.000	-0.005	-0.005	57	0.000	57					1
0.05	0.050					185	0				-5
0.20	0.200	0.191	-0.009	933	0.0 %	933					-2
2.50	2.500	2.452	-0.048	11083	0.0 %	11083					2
5.00	5.000	5.103	0.103	22979	0.0 %	22979					0
7.50	7.500	7.498	-0.002	33727	0.0 %	33727					0
10.0	10.000	9.961	-0.039	44778	0.0 %	44778					0

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	57	-		0 SFAM01.1	PPB	5/21/2022 15:42	Std
	0.2 S01	933	-		0.2 SFAM01.1	PPB	5/21/2022 15:44	Std
	2.5 S02	11083	-		2.5 SFAM01.1	PPB	5/21/2022 15:46	Std
	5 S03	22979	-		5 SFAM01.1	PPB	5/21/2022 15:49	Std
	7.5 S04	33727	-		7.5 SFAM01.1	PPB	5/21/2022 15:51	Std
	10 S05	44778	-		10 SFAM01.1	PPB	5/21/2022 15:56	Std
ICV036	ICV036	17749	3.9377 -		SFAM01.1	PPB	5/21/2022 15:59	SMPL
ICB036	ICB036	-72	-0.0334 -		SFAM01.1	PPB	5/21/2022 16:01	SMPL
CCV007	CCV007	22935	5.0933 -		SFAM01.1	PPB	5/21/2022 16:03	SMPL
CCB007	CCB007	-93	-0.038 -		SFAM01.1	PPB	5/21/2022 16:05	SMPL
PB144763BL	PBS763	13	-0.0144 -		SFAM01.1	PPB	5/21/2022 16:08	SMPL
N2772-01DLX10	YB776	89343	19.891 -		SFAM01.1	PPB	5/21/2022 16:15	SMPL
N2772-02DLX10	YB777	56088	12.4808 -		SFAM01.1	PPB	5/21/2022 16:18	SMPL
N2772-03DLX10	YB778	149224	33.2342 -		SFAM01.1	PPB	5/21/2022 16:20	SMPL
N2772-04DLX10	YB779	59401	13.219 -		SFAM01.1	PPB	5/21/2022 16:22	SMPL
N2772-05DLX10	YB780	70332	15.6548 -		SFAM01.1	PPB	5/21/2022 16:25	SMPL
N2772-09DLX10	YB784	22640	5.0276 -		SFAM01.1	PPB	5/21/2022 16:34	SMPL
N2772-10DLX10	YB785	41671	9.2682 -		SFAM01.1	PPB	5/21/2022 16:36	SMPL
N2772-11DLX10	YB786	43860	9.756 -		SFAM01.1	PPB	5/21/2022 16:38	SMPL
CCV008	CCV008	23555	5.2314 -		SFAM01.1	PPB	5/21/2022 16:44	SMPL
CCB008	CCB008	-117	-0.0434 -		SFAM01.1	PPB	5/21/2022 16:46	SMPL
N2772-12DLX10	YB787	363992	HIGH -		SFAM01.1	PPB	5/21/2022 16:48	
N2772-14DLX10	YB789	302216	HIGH -		SFAM01.1	PPB	5/21/2022 16:57	
N2772-15DLX10	YB790	12679	2.8079 -		SFAM01.1	PPB	5/21/2022 16:59	SMPL
N2772-16DLX10	YB791	100617	22.4031 -		SFAM01.1	PPB	5/21/2022 17:01	SMPL
N2772-17DLX10	YB792	695388	HIGH -		SFAM01.1	PPB	5/21/2022 17:04	
N2772-18DLX10	YB793	316936	HIGH -		SFAM01.1	PPB	5/21/2022 17:05	
N2772-19DLX10	YB794	108962	24.2627 -		SFAM01.1	PPB	5/21/2022 17:07	SMPL
N2772-20DLX10	YB795	509977	HIGH -		SFAM01.1	PPB	5/21/2022 17:10	
N2772-21DLX10	YB795D	519789	HIGH -		SFAM01.1	PPB	5/21/2022 17:12	
N2772-22DLX10	YB795S	514033	HIGH -		SFAM01.1	PPB	5/21/2022 17:14	
CCV009	CCV009	21558	4.7865 -		SFAM01.1	PPB	5/21/2022 17:16	SMPL
CCB009	CCB009	-249	-0.0728 -		SFAM01.1	PPB	5/21/2022 17:18	SMPL
CCV010	CCV010	19427	4.3116 -		SFAM01.1	PPB	5/21/2022 18:06	SMPL
CCB010	CCB010	-249	-0.0728 -		SFAM01.1	PPB	5/21/2022 18:08	SMPL
N2772-06	YB781	51231	11.3985 -		SFAM01.1	PPB	5/21/2022 18:11	SMPL
N2772-07	YB782	4651	1.0191 -		SFAM01.1	PPB	5/21/2022 18:13	SMPL
N2772-08	YB783	3088	0.6708 -		SFAM01.1	PPB	5/21/2022 18:15	SMPL
N2772-13	YB788	6681	1.4714 -		SFAM01.1	PPB	5/21/2022 18:18	SMPL
CCV011	CCV011	21385	4.7479 -		SFAM01.1	PPB	5/21/2022 18:20	SMPL
CCB011	CCB011	-102	-0.04 -		SFAM01.1	PPB	5/21/2022 18:22	SMPL