

LB120295


SFAM01.1

INSTRUMENT ID: CY1

Linear

 μ Abs.:

47134



A= 0.0000e+000

B= 2.1461e-004 slope

C= 2.3542e-002 intercept

Rho= 0.9996472

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.025	0.025	6	0.000	6					6
0.05	0.050					185	0				2
0.20	0.200	0.212	0.012	877	0.0 %	877					-1
2.50	2.500	2.542	0.042	11736	0.0 %	11736					-2
5.00	5.000	4.969	-0.031	23043	0.0 %	23043					1
7.50	7.500	7.313	-0.187	33968	0.0 %	33968					
10.0	10.000	10.139	0.139	47134	0.0 %	47134					

LB120295
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	6	-		0 SFAM01.1	PPB	5/22/2022 10:38	Std
	0.2 S01	877	-		0.2 SFAM01.1	PPB	5/22/2022 10:43	Std
	2.5 S02	11736	-		2.5 SFAM01.1	PPB	5/22/2022 10:46	Std
	5 S03	23043	-		5 SFAM01.1	PPB	5/22/2022 10:50	Std
	7.5 S04	33968	-		7.5 SFAM01.1	PPB	5/22/2022 10:55	Std
	10 S05	47134	-		10 SFAM01.1	PPB	5/22/2022 10:59	Std
ICV037	ICV037	19447	4.1971	-	SFAM01.1	PPB	5/22/2022 11:07	SMPL
ICB037	ICB037	-74	0.0077	-	SFAM01.1	PPB	5/22/2022 11:10	SMPL
CCV012	CCV012	23106	4.9823	-	SFAM01.1	PPB	5/22/2022 11:12	SMPL
CCB012	CCB012	-83	0.0057	-	SFAM01.1	PPB	5/22/2022 11:14	SMPL
N2772-01DL2X50	YB776	18330	3.9573	-	SFAM01.1	PPB	5/22/2022 11:16	SMPL
N2772-02DL2X25	YB777	15730	3.3994	-	SFAM01.1	PPB	5/22/2022 11:19	SMPL
N2772-03DL2X100	YB778	13638	2.9504	-	SFAM01.1	PPB	5/22/2022 11:21	SMPL
N2772-04DL2X25	YB779	21534	4.645	-	SFAM01.1	PPB	5/22/2022 11:23	SMPL
N2772-05DL2X50	YB780	13588	2.9397	-	SFAM01.1	PPB	5/22/2022 11:26	SMPL
N2772-06DLX2	YB781	24466	5.2742	-	SFAM01.1	PPB	5/22/2022 11:28	SMPL
N2772-16DL2X50	YB791	17550	3.79	-	SFAM01.1	PPB	5/22/2022 11:35	SMPL
N2772-18DL2X100	YB793	27550	5.9361	-	SFAM01.1	PPB	5/22/2022 11:39	SMPL
CCV013	CCV013	21871	4.7173	-	SFAM01.1	PPB	5/22/2022 11:53	SMPL
CCB013	CCB013	-83	0.0057	-	SFAM01.1	PPB	5/22/2022 11:55	SMPL
PB144764BL	PBS764	13	0.0263	-	SFAM01.1	PPB	5/22/2022 11:57	SMPL
N2773-01	YB796	7930	1.7254	-	SFAM01.1	PPB	5/22/2022 11:59	SMPL
N2773-02	YB797	2867	0.6388	-	SFAM01.1	PPB	5/22/2022 12:02	SMPL
N2773-03DLX5	YB798	20097	4.3366	-	SFAM01.1	PPB	5/22/2022 12:04	SMPL
N2773-04DLX10	YB799	49448	10.6356	-	SFAM01.1	PPB	5/22/2022 12:06	SMPL
N2773-05	YB7A0	42869	9.2237	-	SFAM01.1	PPB	5/22/2022 12:09	SMPL
N2773-06	YB7A1	13024	2.8186	-	SFAM01.1	PPB	5/22/2022 12:11	SMPL
CCV014	CCV014	22541	4.8611	-	SFAM01.1	PPB	5/22/2022 12:13	SMPL
CCB014	CCB014	-93	0.0036	-	SFAM01.1	PPB	5/22/2022 12:15	SMPL
N2773-07DLX10	YB7A2	46268	9.9531	-	SFAM01.1	PPB	5/22/2022 12:18	SMPL
N2773-08DLX10	YB7A3	526330	HIGH	-	SFAM01.1	PPB	5/22/2022 12:20	
N2773-09DLX10	YB7A4	108042	23.2105	-	SFAM01.1	PPB	5/22/2022 12:22	SMPL
N2773-10DLX10	YB7A5	810861	HIGH	-	SFAM01.1	PPB	5/22/2022 12:24	
N2773-11	YB7A6	4959	1.0878	-	SFAM01.1	PPB	5/22/2022 12:35	SMPL
N2773-12	YB7A6D	4786	1.0507	-	SFAM01.1	PPB	5/22/2022 12:37	SMPL
N2773-13	YB7A6S	16888	3.6479	-	SFAM01.1	PPB	5/22/2022 12:42	SMPL
CCV014	CCV014	22253	4.7993	-	SFAM01.1	PPB	5/22/2022 12:45	SMPL
CCB014	CCB014	-42	0.0145	-	SFAM01.1	PPB	5/22/2022 12:47	SMPL
N2772-12DL2X200		16921	3.655	-	SFAM01.1	PPB	5/22/2022 12:49	SMPL
N2772-14DL2X200		14172	3.065	-	SFAM01.1	PPB	5/22/2022 12:51	SMPL
N2772-17DL2X500		17958	3.8775	-	SFAM01.1	PPB	5/22/2022 12:54	SMPL
N2772-19DL2X50		22491	4.8503	-	SFAM01.1	PPB	5/22/2022 12:56	SMPL
N2772-20DL2X500		12246	2.6517	-	SFAM01.1	PPB	5/22/2022 12:58	SMPL
N2772-21DL2X500		12509	2.7081	-	SFAM01.1	PPB	5/22/2022 13:00	SMPL
N2772-22DL2X500		11702	2.5349	-	SFAM01.1	PPB	5/22/2022 13:03	SMPL
N2773-04DL2X25		18846	4.0681	-	SFAM01.1	PPB	5/22/2022 13:05	SMPL

LB120295

INSTRUMENT ID : CV1

N2773-07DL2X20		21570	4.6527 -	SFAM01.1	PPB	5/22/2022 13:07 SMPL
N2773-08DL2X500		13459	2.912 -	SFAM01.1	PPB	5/22/2022 13:10 SMPL
N2773-09DLX50		20873	4.5031 -	SFAM01.1	PPB	5/22/2022 13:12 SMPL
N2773-10DL2X1000		12888	2.7894 -	SFAM01.1	PPB	5/22/2022 13:16 SMPL
CCV015	CCV015	21263	4.5868 -	SFAM01.1	PPB	5/22/2022 13:31 SMPL
CCB015	CCB015	-47	0.0135 -	SFAM01.1	PPB	5/22/2022 13:33 SMPL