

LB120296


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

54332



A= 0.0000e+000

B= 1.8660e-004

C= 1.5910e-002

Rho= 0.9996139

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% D
0.0	0.000	0.037	0.037	115	0.000	115					-
0.2	0.200	0.229	0.029	1143	0.0 %	1143					15
2.5	2.500	2.529	0.029	13465	0.0 %	13465					1
5.0	5.000	4.925	-0.075	26309	0.0 %	26309					-2
7.5	7.500	7.325	-0.175	39170	0.0 %	39170					-2
10.0	10.000	10.154	0.154	54332	0.0 %	54332					2

LB120296 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 22 May 2022 15:04:08

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	5.6	115	-	0.0000	7471B	PPB	22 May 2022 15:11:48	S	Std
0.2 - 1	5.2	1143	-	0.2000	7471B	PPB	22 May 2022 15:14:05	S	Std
2.5 - 1	5.2.5	13465	-	2.5000	7471B	PPB	22 May 2022 15:16:22	S	Std
5.0 - 1	5.5	26309	-	5.0000	7471B	PPB	22 May 2022 15:18:38	S	Std
7.5 - 1	5.7.5	39170	-	7.5000	7471B	PPB	22 May 2022 15:20:56	S	Std
10.0 - 1	5.10	54332	-	10.0000	7471B	PPB	22 May 2022 15:23:13	S	Std
ICV38 - 1	ICV38	21545	4.0363	-	7471B	PPB	22 May 2022 15:26:01	U	SMPL
ICB38 - 1	ICB38	-37	0.0090	-	7471B	PPB	22 May 2022 15:28:16	U	SMPL
CCV16 - 1	CCV16	25067	4.6935	-	7471B	PPB	22 May 2022 15:30:32	U	SMPL
CCB16 - 1	CCB16	-32	0.0099	-	7471B	PPB	22 May 2022 15:32:47	U	SMPL
CRA - 1	CRA	1121	0.2251	-	7471B	PPB	22 May 2022 15:35:04	U	SMPL
HighStd - 1	HighStd	53866	10.0675	-	7471B	PPB	22 May 2022 15:37:18	U	SMPL
ChkStd - 1	ChkStd	37148	6.9479	-	7471B	PPB	22 May 2022 15:39:33	U	SMPL
PB145027BL - 1	PB145027BL	-73	0.0023	-	7471B	PPB	22 May 2022 15:41:51	U	SMPL
PB145027BS - 1	PB145027BS	21085	3.9504	-	7471B	PPB	22 May 2022 15:44:08	U	SMPL
N2984-01 - 1	P014-SS002-0006-01	23926	4.4806	-	7471B	PPB	22 May 2022 15:46:24	U	SMPL
N2984-02 - 1	P014-SS002-0612-01	28772	5.3849	-	7471B	PPB	22 May 2022 15:48:41	U	SMPL
N2984-03 - 1	P014-SS002-1218-01	4810	0.9135	-	7471B	PPB	22 May 2022 15:51:00	U	SMPL
N2984-04 - 1	P014-SS002-1824-01	2229	0.4318	-	7471B	PPB	22 May 2022 15:53:19	U	SMPL
N2984-05 - 1	P014-SS002-2430-01	2276	0.4406	-	7471B	PPB	22 May 2022 15:55:33	U	SMPL
CCV17 - 1	CCV17	24643	4.6144	-	7471B	PPB	22 May 2022 15:57:49	U	SMPL
CCB17 - 1	CCB17	-64	0.0040	-	7471B	PPB	22 May 2022 16:00:04	U	SMPL
N2984-06 - 1	P014-SS003-0006-01	5721	1.0835	-	7471B	PPB	22 May 2022 16:02:21	U	SMPL
N2984-11 - 1	P020-SS001-0006-01	7861	1.4828	-	7471B	PPB	22 May 2022 16:04:37	U	SMPL
N2984-12 - 1	P020-SS001-0612-01	3908	0.7452	-	7471B	PPB	22 May 2022 16:06:52	U	SMPL
N2984-13 - 1	P020-SS001-1218-01	2533	0.4886	-	7471B	PPB	22 May 2022 16:09:08	U	SMPL
N2984-14 - 1	P020-SS001-1824-01	1724	0.3376	-	7471B	PPB	22 May 2022 16:11:24	U	SMPL
N2984-15 - 1	P020-SS001-2430-01	1443	0.2852	-	7471B	PPB	22 May 2022 16:13:41	U	SMPL
N2984-16 - 1	P021-SS001-0006-01	4160	0.7922	-	7471B	PPB	22 May 2022 16:15:57	U	SMPL
N2984-17 - 1	P021-SS001-0612-01	5526	1.0471	-	7471B	PPB	22 May 2022 16:18:14	U	SMPL
N2984-18 - 1	P021-SS001-1218-01	7632	1.4401	-	7471B	PPB	22 May 2022 16:20:31	U	SMPL
N2984-19 - 1	P021-SS002-0006-01	5007	0.9502	-	7471B	PPB	22 May 2022 16:22:46	U	SMPL
CCV18 - 1	CCV18	25047	4.6898	-	7471B	PPB	22 May 2022 16:25:02	U	SMPL
CCB18 - 1	CCB18	-22	0.0118	-	7471B	PPB	22 May 2022 16:27:18	U	SMPL
N2985-04 - 1	P021-SS003-0006-01	3736	0.7131	-	7471B	PPB	22 May 2022 16:29:35	U	SMPL
N2985-09 - 1	P021-SS004-0006-01	8326	1.5696	-	7471B	PPB	22 May 2022 16:31:50	U	SMPL
N2985-10 - 1	P021-SS005-0006-01	5224	0.9907	-	7471B	PPB	22 May 2022 16:34:06	U	SMPL
N2985-15 - 1	P021-SS006-0006-01	7115	1.3436	-	7471B	PPB	22 May 2022 16:36:21	U	SMPL
N2985-19 - 1	P021-SS007-0006-01	37057	6.9309	-	7471B	PPB	22 May 2022 16:38:36	U	SMPL
N2985-19DUP - 1	P021-SS007-0006-01DUP	37298	6.9759	-	7471B	PPB	22 May 2022 16:40:52	U	SMPL
N2985-19MS - 1	P021-SS007-0006-01MS	59339	11.0888	-	7471B	PPB	22 May 2022 16:43:10	U	SMPL
N2985-19MSD - 1	P021-SS007-0006-01MSD	64819	12.1114	-	7471B	PPB	22 May 2022 16:53:00	U	SMPL
PB145028BL - 1	PBS	-194	-0.0203	-	7471B	PPB	22 May 2022 16:55:17	U	SMPL
PB145028BS - 1	LCSS	22224	4.1630	-	7471B	PPB	22 May 2022 16:57:35	U	SMPL
CCV19 - 1	CCV19	26936	5.0423	-	7471B	PPB	22 May 2022 16:59:51	U	SMPL
CCB19 - 1	CCB19	-47	0.0071	-	7471B	PPB	22 May 2022 17:02:07	U	SMPL
N2986-04 - 1	P027-SS001-0006-01	12974	2.4369	-	7471B	PPB	22 May 2022 17:04:24	U	SMPL
N2986-04DUP - 1	P027-SS001-0006-01DUP	12546	2.3570	-	7471B	PPB	22 May 2022 17:06:39	U	SMPL
N2986-04MS - 1	P027-SS001-0006-01MS	32967	6.1677	-	7471B	PPB	22 May 2022 17:08:55	U	SMPL
N2986-04MSD - 1	P027-SS001-0006-01MSD	32823	6.1408	-	7471B	PPB	22 May 2022 17:11:12	U	SMPL
N2986-05 - 1	P027-SS001-0612-01	5250	0.9956	-	7471B	PPB	22 May 2022 17:13:29	U	SMPL
N2986-06 - 1	P027-SS001-1218-01	2595	0.5001	-	7471B	PPB	22 May 2022 17:15:48	U	SMPL
N2986-07 - 1	P027-SS001-1824-01	1446	0.2857	-	7471B	PPB	22 May 2022 17:18:04	U	SMPL
N2986-08 - 1	P027-SS001-2430-01	1410	0.2790	-	7471B	PPB	22 May 2022 17:20:21	U	SMPL
N2986-09 - 1	P027-SS002-0006-01	10836	2.0379	-	7471B	PPB	22 May 2022 17:22:38	U	SMPL
N2986-14 - 1	P027-SS003-0006-01	4658	0.8851	-	7471B	PPB	22 May 2022 17:24:53	U	SMPL
CCV20 - 1	CCV20	27112	5.0751	-	7471B	PPB	22 May 2022 17:27:09	U	SMPL
CCB20 - 1	CCB20	-37	0.0090	-	7471B	PPB	22 May 2022 17:29:24	U	SMPL
N2987-01 - 1	P033-SS001-0006-01	7430	1.4024	-	7471B	PPB	22 May 2022 17:31:41	U	SMPL
N2987-02 - 1	P033-SS001-0612-01	5556	1.0527	-	7471B	PPB	22 May 2022 17:33:57	U	SMPL
N2987-03 - 1	P033-SS001-1218-01	4149	0.7901	-	7471B	PPB	22 May 2022 17:36:12	U	SMPL
N2987-04 - 1	P033-SS001-1824-01	3300	0.6317	-	7471B	PPB	22 May 2022 17:38:27	U	SMPL
N2987-05 - 1	P033-SS001-2430-01	2446	0.4723	-	7471B	PPB	22 May 2022 17:40:43	U	SMPL
N2987-06 - 1	P034-SS001-0006-01	2575	0.4964	-	7471B	PPB	22 May 2022 17:42:59	U	SMPL
N2987-07 - 1	P034-SS001-0612-01	2472	0.4772	-	7471B	PPB	22 May 2022 17:45:15	U	SMPL
N2987-08 - 1	P034-SS001-1218-01	1930	0.3761	-	7471B	PPB	22 May 2022 17:47:32	U	SMPL
N2987-09 - 1	P034-SS001-1824-01	3168	0.6071	-	7471B	PPB	22 May 2022 17:49:48	U	SMPL
N2987-10 - 1	P034-SS001-2430-01	5024	0.9534	-	7471B	PPB	22 May 2022 17:52:04	U	SMPL
CCV21 - 1	CCV21	27381	5.1253	-	7471B	PPB	22 May 2022 17:54:20	U	SMPL
CCB21 - 1	CCB21	-1	0.0157	-	7471B	PPB	22 May 2022 17:56:36	U	SMPL
N2987-15 - 1	P027-SS004-0006-01	3588	0.6854	-	7471B	PPB	22 May 2022 17:58:53	U	SMPL
PB145029BL - 1	PBS	49	0.0251	-	7471B	PPB	22 May 2022 18:01:09	U	SMPL
PB145029BS - 1	LCSS	20531	3.8471	-	7471B	PPB	22 May 2022 18:03:25	U	SMPL
N2988-01 - 1	P015-SS002-0006-01	6561	1.2402	-	7471B	PPB	22 May 2022 18:05:41	U	SMPL
N2988-01DUP - 1	P015-SS002-0006-01DUP	5412	1.0258	-	7471B	PPB	22 May 2022 18:07:58	U	SMPL
N2988-02 - 1	N2988-01MS	24480	4.5840	-	7471B	PPB	22 May 2022 18:10:14	U	SMPL
N2988-03 - 1	N2988-01MSD	26561	4.9723	-	7471B	PPB	22 May 2022 18:12:30	U	SMPL
N2988-04 - 1	P015-SS002-0006-02	4788	0.9094	-	7471B	PPB	22 May 2022 18:14:48	U	SMPL
N2988-05 - 1	P015-SS002-0612-01	5041	0.9566	-	7471B	PPB	22 May 2022 18:17:06	U	SMPL
N2988-06 - 1	P015-SS002-1218-01	1529	0.3012	-	7471B	PPB	22 May 2022 18:19:22	U	SMPL
CCV22 - 1	CCV22	26732	5.0042	-	7471B	PPB	22 May 2022 18:21:38	U	SMPL
CCB22 - 1	CCB22	8	0.0174	-	7471B	PPB	22 May 2022 18:23:55	U	SMPL
N2988-07 - 1	P015-SS002-1824-01	1725	0.3378	-	7471B	PPB	22 May 2022 18:26:12	U	SMPL
N2988-08 - 1	P015-SS002-2430-01	2296	0.4444	-	7471B	PPB	22 May 2022 18:28:28	U	SMPL
N2988-09 - 1	P015-SS003-0006-01	2704	0.5205	-	7471B	PPB	22 May 2022 18:30:43	U	SMPL
N2988-14 - 1	P015-SS004-0006-01	2640	0.5085	-	7471B	PPB	22 May 2022 18:33:00	U	SMPL
N2988-19 - 1	P015-SS005-0006-01	3975	0.7577	-	7471B	PPB	22 May 2022 18:35:16	U	SMPL
N2988-24 - 1	P015-SS006-0006-01	2311	0.4472	-	7471B	PPB	22 May 2022 18:37:31	U	SMPL
N2989-05 - 1	P015-SS007-0006-01	1860	0.3630	-	7471B	PPB	22 May 2022 18:39:47	U	SMPL
N2989-10 - 1	P015-SS008-0006-01	4685	0.8901	-	7471B	PPB	22 May 2022 18:42:04	U	SMPL
N2989-15 - 1	P015-SS009-0006-01	3128	0.5996	-	7471B	PPB	22 May 2022 18:44:21	U	SMPL
N2989-20 - 1	P015-SS010-0006-01	2776	0.5339	-	7471B	PPB	22 May 2022 18:46:37	U	SMPL
CCV23 - 1	CCV23	25648	4.8019	-	7471B	PPB	22 May 2022 18:48:53	U	SMPL

LB120296 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 22 May 2022 15:04:08

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
CCB23 - 1	CCB23	-2	0.0155		-7471B	PPB	22 May 2022 18:51:09	U	SMPL
N2990-05 - 1	P028-SS001-0006-01	2676	0.5153		-7471B	PPB	22 May 2022 18:53:27	U	SMPL
N2990-06 - 1	P028-SS001-0612-01	2920	0.5608		-7471B	PPB	22 May 2022 18:55:43	U	SMPL
N2990-07 - 1	P028-SS001-0612-02	2673	0.5147		-7471B	PPB	22 May 2022 18:57:59	U	SMPL
N2990-08 - 1	P028-SS001-1218-01	2739	0.5270		-7471B	PPB	22 May 2022 19:00:15	U	SMPL
N2990-09 - 1	P028-SS002-0006-01	5086	0.9650		-7471B	PPB	22 May 2022 19:02:30	U	SMPL
N2990-21 - 1	N2990-06MS	25595	4.7920		-7471B	PPB	22 May 2022 19:04:46	U	SMPL
N2990-22 - 1	N2990-06MSD	23250	4.3544		-7471B	PPB	22 May 2022 19:07:02	U	SMPL
CCV24 - 1	CCV24	27204	5.0923		-7471B	PPB	22 May 2022 19:09:20	U	SMPL
CCB24 - 1	CCB24	25	0.0206		-7471B	PPB	22 May 2022 19:11:39	U	SMPL
N2985-19LX5 - 1		8596	1.6200		-7471B	PPB	22 May 2022 19:13:56	U	SMPL
N2985-19A - 1		64371	12.0278		-7471B	PPB	22 May 2022 19:16:13	U	SMPL
N2986-04LX5 - 1		2162	0.4193		-7471B	PPB	22 May 2022 19:18:29	U	SMPL
N2986-04A - 1		33272	6.2246		-7471B	PPB	22 May 2022 19:20:48	U	SMPL
N2988-01LX5 - 1		6303	1.1921		-7471B	PPB	22 May 2022 19:23:04	U	SMPL
N2988-01A - 1		25567	4.7868		-7471B	PPB	22 May 2022 19:25:22	U	SMPL
N2985-19DLX2 - 1		23773	4.4520		-7471B	PPB	22 May 2022 19:27:39	U	SMPL
N2985-19DUPDLX2 - 1		21344	3.9988		-7471B	PPB	22 May 2022 19:29:56	U	SMPL
N2985-19MSDLX2 - 1		33727	6.3095		-7471B	PPB	22 May 2022 19:32:13	U	SMPL
N2985-19MSDDLX2 - 1		31509	5.8956		-7471B	PPB	22 May 2022 19:34:31	U	SMPL
CCV25 - 1	CCV25	28365	5.3089		-7471B	PPB	22 May 2022 19:36:49	U	SMPL
CCB25 - 1	CCB25	-1	0.0157		-7471B	PPB	22 May 2022 19:39:08	U	SMPL