


LB120301

7471B

INSTRUMENT ID: CV2



Linear

A= 0.0000e+000

B= 1.7371e-004

C= 2.2774e-003

Rho= 0.9999644

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%ob
0.0	0.000	0.025	0.025	130	0.000	130					5
0.2	0.200	0.210	0.010	1194	0.0 %	1194					-2
2.5	2.500	2.440	-0.060	14034	0.0 %	14034					0
5.0	5.000	5.004	0.004	28796	0.0 %	28796					0
7.5	7.500	7.536	0.036	43368	0.0 %	43368					0
10.0	10.000	9.985	-0.015	57469	0.0 %	57469					0

LB120301 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 23 May 2022 12:09:15

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	130	-	0.0000	7471B	PPB	23 May 2022 12:14:57	S	Std
0.2 - 1	50.2	1194	-	0.2000	7471B	PPB	23 May 2022 12:17:14	S	Std
2.5 - 1	52.5	14034	-	2.5000	7471B	PPB	23 May 2022 12:19:31	S	Std
5.0 - 1	55	28796	-	5.0000	7471B	PPB	23 May 2022 12:21:48	S	Std
7.5 - 1	53.5	43368	-	7.5000	7471B	PPB	23 May 2022 12:24:05	S	Std
10.0 - 1	510	57469	-	10.0000	7471B	PPB	23 May 2022 12:31:45	S	Std
ICV39 - 1	ICV39	23854	4.1459	-	7471B	PPB	23 May 2022 12:35:06	U	SMPL
ICB39 - 1	ICB39	-13	0.0000	-	7471B	PPB	23 May 2022 12:37:20	U	SMPL
CCV26 - 1	CCV26	27319	4.7479	-	7471B	PPB	23 May 2022 12:39:37	U	SMPL
CCB26 - 1	CCB26	0	0.0023	-	7471B	PPB	23 May 2022 12:41:51	U	SMPL
CRA - 1	CRA	1199	0.2106	-	7471B	PPB	23 May 2022 12:44:08	U	SMPL
HighStd - 1	HighStd	57157	9.9310	-	7471B	PPB	23 May 2022 12:46:22	U	SMPL
ChkStd - 1	ChkStd	41307	7.1777	-	7471B	PPB	23 May 2022 12:48:37	U	SMPL
PB145032BL - 1	PBS	-185	-0.0299	-	7471B	PPB	23 May 2022 12:50:55	U	SMPL
PB145032BS - 1	LCSS	22800	3.9629	-	7471B	PPB	23 May 2022 12:53:13	U	SMPL
N2921-01 - 1	P013-SS002-0006-01	1605	0.2811	-	7471B	PPB	23 May 2022 12:55:29	U	SMPL
N2921-02 - 1	P013-SS002-0612-01	2678	0.4675	-	7471B	PPB	23 May 2022 12:57:47	U	SMPL
N2921-03 - 1	P013-SS002-1218-01	2578	0.4501	-	7471B	PPB	23 May 2022 13:00:04	U	SMPL
N2921-04 - 1	P013-SS002-1824-01	2345	0.4096	-	7471B	PPB	23 May 2022 13:02:20	U	SMPL
N2921-05 - 1	P013-SS002-2430-01	5221	0.9092	-	7471B	PPB	23 May 2022 13:04:35	U	SMPL
CCV27 - 1	CCV27	27902	4.8491	-	7471B	PPB	23 May 2022 13:06:50	U	SMPL
CCB27 - 1	CCB27	-2	0.0019	-	7471B	PPB	23 May 2022 13:09:05	U	SMPL
N2921-06 - 1	P013-SS003-0006-01	7247	1.2612	-	7471B	PPB	23 May 2022 13:11:21	U	SMPL
N2921-14 - 1	P013-SS004-0006-01	904	0.1593	-	7471B	PPB	23 May 2022 13:13:37	U	SMPL
N2921-19 - 1	P013-SS005-0006-01	935	0.1647	-	7471B	PPB	23 May 2022 13:15:52	U	SMPL
N2921-24 - 1	P013-SS006-0006-01	2272	0.3969	-	7471B	PPB	23 May 2022 13:18:08	U	SMPL
N2922-05 - 1	P024-SS001-0006-01	3715	0.6476	-	7471B	PPB	23 May 2022 13:20:23	U	SMPL
N2922-06 - 1	P024-SS001-0612-0	1671	0.2925	-	7471B	PPB	23 May 2022 13:22:39	U	SMPL
N2922-07 - 1	P024-SS001-1218-01	1734	0.3035	-	7471B	PPB	23 May 2022 13:24:56	U	SMPL
N2922-08 - 1	P024-SS001-1824-01	2209	0.3860	-	7471B	PPB	23 May 2022 13:27:13	U	SMPL
N2922-09 - 1	P024-SS001-2430-01	1468	0.2573	-	7471B	PPB	23 May 2022 13:29:29	U	SMPL
N2922-09DUP - 1	P024-SS001-2430-01DUP	1504	0.2635	-	7471B	PPB	23 May 2022 13:31:45	U	SMPL
CCV28 - 1	CCV28	27196	4.7265	-	7471B	PPB	23 May 2022 13:34:00	U	SMPL
CCB28 - 1	CCB28	-75	-0.0108	-	7471B	PPB	23 May 2022 13:36:16	U	SMPL
N2922-10 - 1	N2922-09MS	25021	4.3487	-	7471B	PPB	23 May 2022 13:38:32	U	SMPL
N2922-11 - 1	N2922-09MSD	24872	4.3228	-	7471B	PPB	23 May 2022 13:43:59	U	SMPL
N2922-12 - 1	P024-SS001-2430-02	1464	0.2566	-	7471B	PPB	23 May 2022 13:46:14	U	SMPL
N2922-13 - 1	P024-SS002-0006-01	3121	0.5444	-	7471B	PPB	23 May 2022 13:48:31	U	SMPL
N2922-21 - 1	P024-SS003-0006-01	19440	3.3792	-	7471B	PPB	23 May 2022 13:50:47	U	SMPL
N2922-25 - 1	P024-SS004-0006-01	3502	0.6106	-	7471B	PPB	23 May 2022 13:53:03	U	SMPL
PB145036BL - 1	PBS	-21	-0.0014	-	7471B	PPB	23 May 2022 13:55:21	U	SMPL
PB145036BS - 1	LCSS	20538	3.5699	-	7471B	PPB	23 May 2022 13:57:37	U	SMPL
N2893-01 - 1	P004-SS002-0006-01	1854	0.3243	-	7471B	PPB	23 May 2022 13:59:54	U	SMPL
N2893-02 - 1	P004-SS002-0612-01	1970	0.3445	-	7471B	PPB	23 May 2022 14:02:10	U	SMPL
CCV29 - 1	CCV29	28602	4.9707	-	7471B	PPB	23 May 2022 14:04:26	U	SMPL
CCB29 - 1	CCB29	-96	-0.0144	-	7471B	PPB	23 May 2022 14:06:41	U	SMPL
N2893-03 - 1	P004-SS002-1218-01	4235	0.7379	-	7471B	PPB	23 May 2022 14:08:58	U	SMPL
N2893-04 - 1	P004-SS002-1824-01	5246	0.9136	-	7471B	PPB	23 May 2022 14:11:14	U	SMPL
N2893-05 - 1	P004-SS002-2430-01	11433	1.9883	-	7471B	PPB	23 May 2022 14:13:29	U	SMPL
N2893-06 - 1	P004-SS003-0006-01	1544	0.2705	-	7471B	PPB	23 May 2022 14:15:45	U	SMPL
N2893-11 - 1	P004-SS004-0006-01	1973	0.3450	-	7471B	PPB	23 May 2022 14:18:01	U	SMPL
N2894-01 - 1	P004-SS005-0006-01	2906	0.5071	-	7471B	PPB	23 May 2022 14:20:17	U	SMPL
N2894-09 - 1	P004-SS006-0006-01	7584	1.3197	-	7471B	PPB	23 May 2022 14:22:34	U	SMPL
N2894-14 - 1	P008-SS002-0006-01	3926	0.6843	-	7471B	PPB	23 May 2022 14:24:50	U	SMPL
N2894-15 - 1	P008-SS002-0612-01	3216	0.5609	-	7471B	PPB	23 May 2022 14:27:07	U	SMPL
N2894-16 - 1	P008-SS002-1218-01	2715	0.4739	-	7471B	PPB	23 May 2022 14:29:23	U	SMPL
CCV30 - 1	CCV30	27459	4.7722	-	7471B	PPB	23 May 2022 14:31:39	U	SMPL
CCB30 - 1	CCB30	-77	-0.0111	-	7471B	PPB	23 May 2022 14:33:54	U	SMPL
N2894-17 - 1	P008-SS002-1824-01	3449	0.6014	-	7471B	PPB	23 May 2022 14:36:11	U	SMPL
N2894-18 - 1	P008-SS002-2430-01	2619	0.4572	-	7471B	PPB	23 May 2022 14:38:26	U	SMPL
N2894-19 - 1	P008-SS003-0006-01	3414	0.5953	-	7471B	PPB	23 May 2022 14:40:42	U	SMPL
N2894-24 - 1	P008-SS004-0006-01	2226	0.3890	-	7471B	PPB	23 May 2022 14:42:57	U	SMPL
N2894-24MS - 1	P008-SS004-0006-01MS	25248	4.3881	-	7471B	PPB	23 May 2022 14:47:29	U	SMPL
N2894-24MSD - 1	P008-SS004-0006-01MSD	25313	4.3994	-	7471B	PPB	23 May 2022 14:49:45	U	SMPL
N2894-24DUP - 1	P008-SS004-0006-01DUP	2204	0.3851	-	7471B	PPB	23 May 2022 14:52:52	U	SMPL
N2911-01 - 1	PASSAIC-AVE-SP	6167	1.0735	-	7471B	PPB	23 May 2022 14:55:29	U	SMPL
N2912-01 - 1	3/4-BLEND-RCA	5732	0.9980	-	7471B	PPB	23 May 2022 14:57:45	U	SMPL
PB145044BL - 1	PBS	-1	0.0021	-	7471B	PPB	23 May 2022 15:00:01	U	SMPL
CCV31 - 1	CCV31	28380	4.9322	-	7471B	PPB	23 May 2022 15:02:17	U	SMPL
CCB31 - 1	CCB31	-61	-0.0083	-	7471B	PPB	23 May 2022 15:04:32	U	SMPL
PB145044BS - 1	LCSS	23375	4.0627	-	7471B	PPB	23 May 2022 15:06:50	U	SMPL
N2895-05 - 1	P030-SS001-0006-01	5633	0.9808	-	7471B	PPB	23 May 2022 15:09:07	U	SMPL
N2895-06 - 1	P030-SS001-0612-01	8003	1.3925	-	7471B	PPB	23 May 2022 15:11:24	U	SMPL
N2895-07 - 1	P030-SS001-1218-01	7825	1.3616	-	7471B	PPB	23 May 2022 15:13:40	U	SMPL
N2895-08 - 1	P030-SS001-1824-01	8486	1.4764	-	7471B	PPB	23 May 2022 15:15:55	U	SMPL
N2895-09 - 1	P030-SS001-2430-01	2746	0.4793	-	7471B	PPB	23 May 2022 15:18:12	U	SMPL
N2895-10 - 1	P030-SS002-0006-01	3319	0.5788	-	7471B	PPB	23 May 2022 15:20:28	U	SMPL
N2895-15 - 1	P030-SS003-0006-01	7294	1.2693	-	7471B	PPB	23 May 2022 15:22:45	U	SMPL
N2895-15MS - 1	P030-SS003-0006-01MS	26854	4.6671	-	7471B	PPB	23 May 2022 15:27:18	U	SMPL
CCV32 - 1	CCV32	28045	4.8740	-	7471B	PPB	23 May 2022 15:29:34	U	SMPL
CCB32 - 1	CCB32	-99	-0.0149	-	7471B	PPB	23 May 2022 15:31:51	U	SMPL

LB120301 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 23 May 2022 12:09:15					
Sample ID	Extended ID	μ Abs.	Conc.	Std Cond	Method	Units	Date	Type	Type
N2895-15MSD - 1	P030-SS003-0006-01MSD	26245	4.5613	-7471B	PPB		23 May 2022 15:34:09	U	SMPL
N2897-01 - 1	P041-SS001-0006-01	5609	0.9766	-7471B	PPB		23 May 2022 15:36:25	U	SMPL
N2897-02 - 1	P041-SS001-0612-01	8008	1.3933	-7471B	PPB		23 May 2022 15:38:43	U	SMPL
N2897-03 - 1	P041-SS001-1218-01	6014	1.0470	-7471B	PPB		23 May 2022 15:40:59	U	SMPL
N2897-04 - 1	P041-SS001-1824-01	4010	0.6989	-7471B	PPB		23 May 2022 15:43:15	U	SMPL
N2897-05 - 1	P041-SS001-2430-01	10290	1.7897	-7471B	PPB		23 May 2022 15:45:31	U	SMPL
N2897-06 - 1	P042-SS001-0006-01	2382	0.4161	-7471B	PPB		23 May 2022 15:47:47	U	SMPL
N2897-07 - 1	P042-SS001-0612-01	4364	0.7603	-7471B	PPB		23 May 2022 15:50:04	U	SMPL
N2897-08 - 1	P042-SS001-1218-01	2670	0.4661	-7471B	PPB		23 May 2022 15:52:21	U	SMPL
N2897-09 - 1	P042-SS001-1824-01	2462	0.4300	-7471B	PPB		23 May 2022 15:54:37	U	SMPL
CCV33 - 1	CCV33	28085	4.8809	-7471B	PPB		23 May 2022 15:56:53	U	SMPL
CCB33 - 1	CCB33	-90	-0.0134	-7471B	PPB		23 May 2022 15:59:09	U	SMPL
N2897-10 - 1	P042-SS001-2430-01	2338	0.4084	-7471B	PPB		23 May 2022 16:01:26	U	SMPL
PB145045BL - 1	PBS	-33	-0.0035	-7471B	PPB		23 May 2022 16:03:42	U	SMPL
PB145045BS - 1	LCSS	17758	3.0870	-7471B	PPB		23 May 2022 16:05:58	U	SMPL
N2896-04 - 1	P038-SS001-0006-01	5409	0.9419	-7471B	PPB		23 May 2022 16:08:14	U	SMPL
N2896-04DUP - 1	P038-SS001-0006-01DUP	4791	0.8345	-7471B	PPB		23 May 2022 16:10:33	U	SMPL
N2896-05 - 1	N2896-04MS	25599	4.4491	-7471B	PPB		23 May 2022 16:12:49	U	SMPL
N2896-06 - 1	N2896-04MSD	25332	4.4027	-7471B	PPB		23 May 2022 16:15:05	U	SMPL
N2896-07 - 1	P038-SS001-0006-02	5292	0.9215	-7471B	PPB		23 May 2022 16:17:24	U	SMPL
N2896-08 - 1	P038-SS001-0612-01	3694	0.6440	-7471B	PPB		23 May 2022 16:19:42	U	SMPL
N2896-09 - 1	P038-SS001-1218-01	2993	0.5222	-7471B	PPB		23 May 2022 16:21:58	U	SMPL
CCV34 - 1	CCV34	28376	4.9315	-7471B	PPB		23 May 2022 16:24:14	U	SMPL
CCB34 - 1	CCB34	-94	-0.0141	-7471B	PPB		23 May 2022 16:26:31	U	SMPL
N2896-10 - 1	P038-SS001-1824-01	1745	0.3054	-7471B	PPB		23 May 2022 16:28:48	U	SMPL
N2896-11 - 1	P038-SS001-2430-01	1386	0.2430	-7471B	PPB		23 May 2022 16:31:04	U	SMPL
N2896-12 - 1	P039-SS001-0006-01	3614	0.6301	-7471B	PPB		23 May 2022 16:33:21	U	SMPL
N2896-13 - 1	P039-SS001-0612-01	3487	0.6080	-7471B	PPB		23 May 2022 16:35:37	U	SMPL
N2896-14 - 1	P039-SS001-1218-01	2738	0.4779	-7471B	PPB		23 May 2022 16:37:53	U	SMPL
N2896-15 - 1	P039-SS001-1824-01	1431	0.2509	-7471B	PPB		23 May 2022 16:40:10	U	SMPL
N2896-16 - 1	P039-SS001-2430-01	3611	0.6295	-7471B	PPB		23 May 2022 16:42:27	U	SMPL
N2896-17 - 1	P040-SS001-0006-01	2189	0.3825	-7471B	PPB		23 May 2022 16:44:43	U	SMPL
N2896-18 - 1	P040-SS001-0612-01	6867	1.1951	-7471B	PPB		23 May 2022 16:47:00	U	SMPL
N2896-19 - 1	P040-SS001-1218-01	2833	0.4944	-7471B	PPB		23 May 2022 16:49:17	U	SMPL
CCV35 - 1	CCV35	28376	4.9315	-7471B	PPB		23 May 2022 16:51:34	U	SMPL
CCB35 - 1	CCB35	-103	-0.0156	-7471B	PPB		23 May 2022 16:53:50	U	SMPL
N2896-20 - 1	P040-SS001-1824-01	3312	0.5776	-7471B	PPB		23 May 2022 16:56:09	U	SMPL
N2896-21 - 1	P040-SS001-2430-01	3774	0.6579	-7471B	PPB		23 May 2022 16:58:25	U	SMPL
N2922-09LX5 - 1		258	0.0471	-7471B	PPB		23 May 2022 17:00:42	U	SMPL
N2922-09A - 1		25019	4.3483	-7471B	PPB		23 May 2022 17:02:58	U	SMPL
N2894-24LX5 - 1		290	0.0527	-7471B	PPB		23 May 2022 17:05:15	U	SMPL
N2894-24A - 1		25135	4.3685	-7471B	PPB		23 May 2022 17:07:33	U	SMPL
N2895-15LX5 - 1		1314	0.2305	-7471B	PPB		23 May 2022 17:09:50	U	SMPL
N2895-15A - 1		27609	4.7982	-7471B	PPB		23 May 2022 17:12:08	U	SMPL
N2896-04LX5 - 1		982	0.1729	-7471B	PPB		23 May 2022 17:14:25	U	SMPL
N2896-04A - 1		25814	4.4864	-7471B	PPB		23 May 2022 17:16:43	U	SMPL
CCV36 - 1	CCV36	28381	4.9323	-7471B	PPB		23 May 2022 17:19:00	U	SMPL
CCB36 - 1	CCB36	-126	-0.0196	-7471B	PPB		23 May 2022 17:21:18	U	SMPL
N2895-15DUP - 1	P030-SS003-0006-01	7283	1.2674	-7471B	PPB		23 May 2022 17:23:36	U	SMPL
CCV37 - 1	CCV37	27802	4.8318	-7471B	PPB		23 May 2022 17:25:52	U	SMPL
CCB37 - 1	CCB37	-44	-0.0054	-7471B	PPB		23 May 2022 17:28:09	U	SMPL