


LB 120318

7471B

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 1.9368e-004

C= -2.9733e-002

Rho= 0.9997517

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.003	0.003	167	0.000	167					-10
0.2	0.200	0.181	-0.019	1087	0.0 %	1087					-2
2.5	2.500	2.446	-0.054	12783	0.0 %	12783					1
5.0	5.000	5.051	0.051	26233	0.0 %	26233					2
7.5	7.500	7.641	0.141	39607	0.0 %	39607					-1
10.0	10.000	9.878	-0.122	51159	0.0 %	51159					

LB120318 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 24 May 2022 08:25:47

Sample ID	Extended ID	μ Abs	Conc	SStd CondMethod	Units	Date	Type	Type
0.0 - 1	50	167	-	0.00007471B	PPB	24 May 2022 08:36:14	S	Std
0.2 - 1	50.2	1087	-	0.20007471B	PPB	24 May 2022 08:38:31	S	Std
2.5 - 1	52.5	12783	-	2.50007471B	PPB	24 May 2022 08:40:48	S	Std
5.0 - 1	55	26233	-	5.00007471B	PPB	24 May 2022 08:43:05	S	Std
7.5 - 1	57.5	39607	-	7.50007471B	PPB	24 May 2022 08:45:22	S	Std
10.0 - 1	51	51159	-	10.00007471B	PPB	24 May 2022 08:47:39	S	Std
ICV40 - 1	ICV40	20799	3.9985	7471B	PPB	24 May 2022 08:50:32	U	SMPL
ICB40 - 1	ICB40	-10	-0.0317	7471B	PPB	24 May 2022 08:52:46	U	SMPL
CCV38 - 1	CCV38	27012	5.2018	7471B	PPB	24 May 2022 08:55:02	U	SMPL
CCB38 - 1	CCB38	-7	-0.0311	7471B	PPB	24 May 2022 08:57:17	U	SMPL
CRA - 1	CRA	1088	0.1810	7471B	PPB	24 May 2022 08:59:33	U	SMPL
HighStd - 1	HighStd	51460	9.9368	7471B	PPB	24 May 2022 09:01:48	U	SMPL
ChkStd - 1	ChkStd	37128	7.1610	7471B	PPB	24 May 2022 09:04:03	U	SMPL
PB145045BL - 1	PBS	60	-0.0181	7471B	PPB	24 May 2022 09:09:13	U	SMPL
PB145045BS - 1	LCSS	21703	4.1738	7471B	PPB	24 May 2022 09:11:29	U	SMPL
N2896-04 - 1	P038-SS001-0006-01	5059	0.9501	7471B	PPB	24 May 2022 09:13:45	U	SMPL
N2896-04DUP - 1	P038-SS001-0006-01DUP	4696	0.8798	7471B	PPB	24 May 2022 09:16:02	U	SMPL
N2896-05 - 1	N2896-04MS	24401	4.6861	7471B	PPB	24 May 2022 09:18:19	U	SMPL
N2896-06 - 1	N2896-04MSD	24585	4.7318	7471B	PPB	24 May 2022 09:20:36	U	SMPL
N2896-07 - 1	P038-SS001-0006-02	5197	0.9768	7471B	PPB	24 May 2022 09:22:52	U	SMPL
CCV39 - 1	CCV39	27024	5.2041	7471B	PPB	24 May 2022 09:25:08	U	SMPL
CCB39 - 1	CCB39	7	-0.0284	7471B	PPB	24 May 2022 09:27:24	U	SMPL
N2896-08 - 1	P038-SS001-0612-01	3693	0.6855	7471B	PPB	24 May 2022 09:29:41	U	SMPL
N2896-09 - 1	P038-SS001-1218-01	2965	0.5445	7471B	PPB	24 May 2022 09:34:11	U	SMPL
N2896-10 - 1	P038-SS001-1824-01	1668	0.2935	7471B	PPB	24 May 2022 09:36:27	U	SMPL
N2896-11 - 1	P038-SS001-2430-01	1396	0.2406	7471B	PPB	24 May 2022 09:38:43	U	SMPL
N2896-12 - 1	P039-SS001-0006-01	3588	0.6652	7471B	PPB	24 May 2022 09:40:59	U	SMPL
N2896-13 - 1	P039-SS001-0612-01	3324	0.6140	7471B	PPB	24 May 2022 09:43:15	U	SMPL
N2896-14 - 1	P039-SS001-1218-01	2647	0.4829	7471B	PPB	24 May 2022 09:45:32	U	SMPL
N2896-15 - 1	P039-SS001-1824-01	1387	0.2389	7471B	PPB	24 May 2022 09:47:49	U	SMPL
N2896-16 - 1	P039-SS001-2430-01	3552	0.6582	7471B	PPB	24 May 2022 09:50:05	U	SMPL
N2896-17 - 1	P040-SS001-0006-01	2096	0.3762	7471B	PPB	24 May 2022 09:52:20	U	SMPL
CCV40 - 1	CCV40	27011	5.2016	7471B	PPB	24 May 2022 09:54:35	U	SMPL
CCB40 - 1	CCB40	27	-0.0245	7471B	PPB	24 May 2022 09:56:52	U	SMPL
N2896-18 - 1	P040-SS001-0612-01	6637	1.2557	7471B	PPB	24 May 2022 09:59:07	U	SMPL
N2896-19 - 1	P040-SS001-1218-01	2791	0.5108	7471B	PPB	24 May 2022 10:01:23	U	SMPL
N2896-20 - 1	P040-SS001-1824-01	3181	0.5863	7471B	PPB	24 May 2022 10:03:38	U	SMPL
N2896-21 - 1	P040-SS001-2430-01	3600	0.6675	7471B	PPB	24 May 2022 10:05:54	U	SMPL
N2896-04LX5 - 1		904	0.1453	7471B	PPB	24 May 2022 10:08:10	U	SMPL
N2896-04A - 1		25728	4.9531	7471B	PPB	24 May 2022 10:10:26	U	SMPL
N2985-19N2985-19N2985-19 - 1	PO21-SS007-0006-01	38089	7.3472	7471B	PPB	24 May 2022 10:12:44	U	SMPL
N2945-19BUP - 1 N2985-19 DUP	PO21-SS007-0006-01DUP	38345	7.3967	7471B	PPB	24 May 2022 10:15:03	U	SMPL
N2945-19MS - 1 N2985-19 MS	PO21-SS007-0006-01MS	54935	10.8098	7471B	PPB	24 May 2022 10:17:20	U	SMPL
N2945-19MSD - 1 N2985-19 MSD	PO21-SS007-0006-01MSD	54617	10.5482	7471B	PPB	24 May 2022 10:19:38	U	SMPL
CCV41 - 1	CCV41	26842	5.1689	7471B	PPB	24 May 2022 10:21:55	U	SMPL
CCB41 - 1	CCB41	103	-0.0098	7471B	PPB	24 May 2022 10:24:12	U	SMPL
N2945-19LX6 - 1 N2985-19 LX6		7473	1.4176	7471B	PPB	24 May 2022 10:26:27	U	SMPL
N2945-19A - 1 N2985-19 A		56709	10.9534	7471B	PPB	24 May 2022 10:28:43	U	SMPL
N2985-19DLX2N2985-19N2985-19 - 1	PO21-SS007-0006-01	21102	4.0572	7471B	PPB	24 May 2022 10:31:09	U	SMPL
N2945-19DUPDLX2 - 1 N2985-19 DUP DLX2	PO21-SS007-0006-01DUP	20263	3.8947	7471B	PPB	24 May 2022 10:33:24	U	SMPL
N2945-19MSDLX2 - 1 N2985-19 MS DLX2	PO21-SS007-0006-01MS	28563	5.5022	7471B	PPB	24 May 2022 10:35:42	U	SMPL
N2945-19MSDDLX2 - 1 N2985-19 MSD DLX2	PO21-SS007-0006-01MSD	29755	5.7331	7471B	PPB	24 May 2022 10:38:00	U	SMPL
N2945-19DLX10 - 1 N2985-19 DLX10		4146	0.7732	7471B	PPB	24 May 2022 10:40:17	U	SMPL
N2945-19ADLX2 - 1 N2985-19 ADLX2		29910	5.7631	7471B	PPB	24 May 2022 10:42:34	U	SMPL
CCV42 - 1	CCV42	27128	5.2243	7471B	PPB	24 May 2022 10:44:51	U	SMPL
CCB42 - 1	CCB42	15	-0.0268	7471B	PPB	24 May 2022 10:47:08	U	SMPL