



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : mohammad

Run Number: LB120319

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	05/24/2022	09:05
2	CAL2	1	Water	NA	NA	20.1	7.01	05/24/2022	09:06
3	CAL3	1	Water	NA	NA	20.1	10.03	05/24/2022	09:08
4	ICV	1	Water	NA	NA	20.2	6.98	05/24/2022	09:10
5	CCV1	1	Water	NA	NA	20.3	2.01	05/24/2022	09:13
6	N2965-02	1	Solid	20.02	20	20.3	6.19	05/24/2022	09:15
7	N3001-01	1	Solid	20.04	20	20.8	6.45	05/24/2022	09:25
8	N3001-01DUP	1	Solid	20.05	20	20.9	6.46	05/24/2022	09:26
9	CCV2	1	Water	NA	NA	20.4	12.03	05/24/2022	09:30

WORKLIST(Hardcopy Internal Chain)

W2120319

WorkList Name : ph n2965 WorkList ID : 159916 Department : Wet-Chemistry Date : 05-24-2022 08:53:35

Due Date	Matrix	Sample	Test	Raw Sample			Collect Date	Method
				Preservative	Customer	Raw Sample Storage Location		
05/19/2022	Solid	N2965-02	pH	Cool 4 deg C	PSEG03	M11 16972	05/19/2022	9045D
05/23/2022	Solid	N3001-01	pH	Cool 4 deg C	PSEG05	N11 BU-03-052322	05/23/2022	9045D

Date/Time 05/24/2022 09:11:00
 Raw Sample Received by: CP SR
 Raw Sample Relinquished by: CP SR

Date/Time 05/24/22 10:11:00
 Raw Sample Received by: CP SR
 Raw Sample Relinquished by: JP (WC)