

LB120323


SFAM01.1

INSTRUMENT ID: CV2

Linear

 μ Abs.:

40114



A= 0.0000e+000

B= 2.5090e-004 *slope*C= -6.5278e-002 *+ intercept*

Rho= 0.9999707

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>6/6/11</i>
0.00	0.000	0.038	0.038	412	0.000	412					-
0.05	0.050					185	0				-2
0.20	0.200	0.196	-0.004	1040	0.0 %	1040					-1
2.50	2.500	2.466	-0.034	10089	0.0 %	10089					-1
5.00	5.000	4.968	-0.032	20059	0.0 %	20059					0
7.50	7.500	7.533	0.033	30285	0.0 %	30285					0
10.0	10.000	9.999	-0.001	40114	0.0 %	40114					

LB120323
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	412	-		0 SFAM01.1	PPB	5/24/2022 11:49	Std
	0.2 S01	1040	-		0.2 SFAM01.1	PPB	5/24/2022 11:51	Std
	2.5 S02	10089	-		2.5 SFAM01.1	PPB	5/24/2022 11:54	Std
	5 S03	20059	-		5 SFAM01.1	PPB	5/24/2022 11:56	Std
	7.5 S04	30285	-		7.5 SFAM01.1	PPB	5/24/2022 11:58	Std
	10 S05	40114	-		10 SFAM01.1	PPB	5/24/2022 12:02	Std
ICV041	ICV041	16643	4.1105	-	SFAM01.1	PPB	5/24/2022 12:07	SMPL
ICB041	ICB041	408	0.0371	-	SFAM01.1	PPB	5/24/2022 12:09	SMPL
CCV043	CCV043	20825	5.1597	-	SFAM01.1	PPB	5/24/2022 12:11	SMPL
CCB043	CCB043	371	0.0278	-	SFAM01.1	PPB	5/24/2022 12:14	SMPL
PB144762BL	PBS762	52	-0.0522	-	SFAM01.1	PPB	5/24/2022 12:16	SMPL
N2715-01	EW2W4	3413	0.791	-	SFAM01.1	PPB	5/24/2022 12:18	SMPL
N2715-02	EW2X5	859	0.1502	-	SFAM01.1	PPB	5/24/2022 12:20	SMPL
N2715-03	EW2X5D	718	0.1149	-	SFAM01.1	PPB	5/24/2022 12:23	SMPL
N2715-04	EW2X5S	11533	2.8284	-	SFAM01.1	PPB	5/24/2022 12:25	SMPL
N2715-05	EW2X6	1155	0.2245	-	SFAM01.1	PPB	5/24/2022 12:27	SMPL
N2715-06	EW2X7	1110	0.2132	-	SFAM01.1	PPB	5/24/2022 12:30	SMPL
N2715-07	EW2X8	119	-0.0354	-	SFAM01.1	PPB	5/24/2022 12:32	SMPL
N2715-08	EW2X9	151	-0.0274	-	SFAM01.1	PPB	5/24/2022 12:34	SMPL
N2715-09	EW2Y0	981	0.1809	-	SFAM01.1	PPB	5/24/2022 12:36	SMPL
CCV044	CCV044	21302	5.2794	-	SFAM01.1	PPB	5/24/2022 12:39	SMPL
CCB044	CCB044	382	0.0306	-	SFAM01.1	PPB	5/24/2022 12:41	SMPL