


LB 120328

7471B



Linear

A= 0.0000e+000

B= 1.8972e-004

C= 7.0338e-003

Rho= 0.9999541

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% (1)
0.0	0.000	0.018	0.018	58	0.000	58					-
0.2	0.200	0.208	0.008	1057	0.0 %	1057					4
2.5	2.500	2.495	-0.005	13116	0.0 %	13116					0
5.0	5.000	4.932	-0.068	25962	0.0 %	25962					-1
7.5	7.500	7.550	0.050	39760	0.0 %	39760					1
10.0	10.000	9.996	-0.004	52654	0.0 %	52654					0

LB120328 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 24 May 2022 13:35:11

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	58	-	0.0000	7471B	PPB	24 May 2022 13:36:07	S	Std
0.2 - 1	50.2	1057	-	0.2000	7471B	PPB	24 May 2022 13:38:23	S	Std
2.5 - 1	52.5	13116	-	2.5000	7471B	PPB	24 May 2022 13:40:40	S	Std
5.0 - 1	55	25962	-	5.0000	7471B	PPB	24 May 2022 13:42:57	S	Std
7.5 - 1	57.5	39760	-	7.5000	7471B	PPB	24 May 2022 13:45:14	S	Std
10.0 - 1	510	52654	-	10.0000	7471B	PPB	24 May 2022 13:50:04	S	Std
ICV42 - 1	ICV42	20843	3.9613	-	7471B	PPB	24 May 2022 13:52:51	U	SMPL
ICB42 - 1	ICB42	-54	-0.0032	-	7471B	PPB	24 May 2022 13:55:06	U	SMPL
CCV45 - 1	CCV45	26267	4.9903	-	7471B	PPB	24 May 2022 13:57:22	U	SMPL
CCB45 - 1	CCB45	45	0.0156	-	7471B	PPB	24 May 2022 13:59:37	U	SMPL
CRA - 1	CRA	1150	0.2252	-	7471B	PPB	24 May 2022 14:01:54	U	SMPL
HighStd - 1	HighStd	53339	10.1263	-	7471B	PPB	24 May 2022 14:04:08	U	SMPL
ChkStd - 1	ChkStd	38429	7.2977	-	7471B	PPB	24 May 2022 14:06:23	U	SMPL
PB145058BL - 1	PBS	-99	-0.0117	-	7471B	PPB	24 May 2022 14:08:40	U	SMPL
PB145058BS - 1	LCS	21552	4.0958	-	7471B	PPB	24 May 2022 14:10:58	U	SMPL
N2923-03 - 1	P024-SS005-0006-01	2568	0.4942	-	7471B	PPB	24 May 2022 14:13:14	U	SMPL
N2923-08 - 1	P024-SS006-0006-01	1824	0.3531	-	7471B	PPB	24 May 2022 14:15:32	U	SMPL
N2923-13 - 1	P024-SS007-0006-01	1506	0.2927	-	7471B	PPB	24 May 2022 14:17:48	U	SMPL
N2923-20 - 1	P024-SS008-0006-01	3146	0.6039	-	7471B	PPB	24 May 2022 14:20:05	U	SMPL
N2927-01 - 1	P010-SS003-0006-01	3718	0.7124	-	7471B	PPB	24 May 2022 14:22:20	U	SMPL
CCV46 - 1	CCV46	26251	4.9873	-	7471B	PPB	24 May 2022 14:24:35	U	SMPL
CCB46 - 1	CCB46	64	0.0192	-	7471B	PPB	24 May 2022 14:26:50	U	SMPL
N2927-02 - 1	P010-SS003-0612-01	1826	0.3535	-	7471B	PPB	24 May 2022 14:29:06	U	SMPL
N2927-03 - 1	P010-SS003-1218-01	2411	0.4644	-	7471B	PPB	24 May 2022 14:31:21	U	SMPL
N2927-04 - 1	P010-SS003-1824-01	1676	0.3250	-	7471B	PPB	24 May 2022 14:33:37	U	SMPL
N2927-05 - 1	P010-SS003-2430-01	1543	0.2998	-	7471B	PPB	24 May 2022 14:35:52	U	SMPL
N2927-06 - 1	P010-SS004-0006-01	4258	0.8148	-	7471B	PPB	24 May 2022 14:38:08	U	SMPL
N2927-11 - 1	P010-SS005-0006-01	4166	0.7974	-	7471B	PPB	24 May 2022 14:40:26	U	SMPL
N2927-16 - 1	P010-SS006-0006-01	3736	0.7158	-	7471B	PPB	24 May 2022 14:42:42	U	SMPL
N2927-21 - 1	P010-SS007-0006-01	4926	0.9416	-	7471B	PPB	24 May 2022 14:44:59	U	SMPL
N2928-04 - 1	P026-SS001-0006-01	4789	0.9156	-	7471B	PPB	24 May 2022 14:47:16	U	SMPL
N2928-05 - 1	P026-SS001-0612-01	4918	0.9401	-	7471B	PPB	24 May 2022 14:49:31	U	SMPL
CCV47 - 1	CCV47	26327	5.0017	-	7471B	PPB	24 May 2022 14:51:46	U	SMPL
CCB47 - 1	CCB47	-5	0.0061	-	7471B	PPB	24 May 2022 14:54:01	U	SMPL
N2928-06 - 1	P026-SS001-1218-01	7156	1.3646	-	7471B	PPB	24 May 2022 14:56:19	U	SMPL
N2928-07 - 1	P026-SS001-1824-01	10014	1.9069	-	7471B	PPB	24 May 2022 14:58:34	U	SMPL
N2928-08 - 1	P026-SS001-2430-01	5061	0.9672	-	7471B	PPB	24 May 2022 15:00:50	U	SMPL
N2928-09 - 1	P026-SS002-0006-01	6355	1.2127	-	7471B	PPB	24 May 2022 15:03:05	U	SMPL
N2928-09DUP - 1	P026-SS002-0006-01DUP	5912	1.1286	-	7471B	PPB	24 May 2022 15:05:20	U	SMPL
N2928-09MS - 1	P026-SS002-0006-01MS	26784	5.0884	-	7471B	PPB	24 May 2022 15:07:36	U	SMPL
N2928-09MSD - 1	P026-SS002-0006-01MSD	27408	5.2068	-	7471B	PPB	24 May 2022 15:09:52	U	SMPL
N2979-01 - 1	NB-7-052022	2348	0.4525	-	7471B	PPB	24 May 2022 15:12:11	U	SMPL
N2928-09LX5 - 1		1052	0.2066	-	7471B	PPB	24 May 2022 15:14:30	U	SMPL
N2928-09A - 1		27267	5.1800	-	7471B	PPB	24 May 2022 15:16:45	U	SMPL
CCV48 - 1	CCV48	26445	5.0241	-	7471B	PPB	24 May 2022 15:19:00	U	SMPL
CCB48 - 1	CCB48	-4	0.0063	-	7471B	PPB	24 May 2022 15:21:17	U	SMPL