

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : 12 Instrument ID : Konelab

5/24/2022 13:36

Test: Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	93.209	0.0	0.079	
ICVB1 ICBI	-0.403	0.0	0.001	
CCV1	240.450	0.0	0.201	
CCB1	-0.429	0.0	0.001	
PB145054BL	-0.498	0.0	0.001	
N2923-03	3.901	0.0	0.005	
N2923-03DUP	4.347	0.0	0.005	
N2923-03MS	35.919	0.0	0.031	
N2923-03MSD	35.481	0.0	0.031	
N2923-08	2.062	0.0	0.003	
N2923-13	0.017	0.0	0.002	
N2923-20	1.588	0.0	0.003	
N2924-01	1.900	0.0	0.003	
CCV2	232.267	0.0	0.195	
CCB2	-0.505	0.0	0.001	
N2924-06	0.014	0.0	0.002	
N2928-04	-0.440	0.0	0.001	
N2928-05	4.375	0.0	0.005	
N2928-06	0.408	0.0	0.002	
N2928-07	-0.509	0.0	0.001	
N2928-08	-0.044	0.0	0.002	
N2928-09	-0.504	0.0	0.001	
PB145055BL	-0.447	0.0	0.001	
N2927-01	-0.343	0.0	0.001	
CCV3	233.528	0.0	0.196	
CCB3	-0.407	0.0	0.001	
N2927-01DUP	-0.423	0.0	0.001	
N2927-01MS	35.151	0.0	0.031	
N2927-01MSD	34.793	0.0	0.030	
N2927-02	-0.429	0.0	0.001	
N2927-03	-0.627	0.0	0.001	
N2927-04	0.148	0.0	0.002	
N2927-05	-0.911	0.0	0.001	
N2927-06	-0.788	0.0	0.001	
N2927-11	-0.780	0.0	0.001	
N2927-16	-0.664	0.0	0.001	
CCV4	243.316	0.0	0.204	
CCB4	-0.303	0.0	0.001	
N2927-21	0.404	0.0	0.002	
PB145054BS	97.785	0.0	0.083	
PB145055BS	97.334	0.0	0.082	
CCV5	250.027	0.0	0.209	
CCB5	-0.262	0.0	0.001	

N 43
Mean 38.109
SD 78.4302
CV% 205.80

Aquakem v. 7.2AQ1

Results from time period:

Tue May 24 09:50:18 2022

Tue May 24 13:31:01 2022

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Total CN	P	-0.8262	µg/l	5/24/2022 10:08:37	
5.0PPBCN	A	Total CN	P	4.2461	µg/l	5/24/2022 10:08:38	
10PPBCN	A	Total CN	P	9.3562	µg/l	5/24/2022 10:08:39	
50PPBCN	A	Total CN	P	50.2876	µg/l	5/24/2022 10:08:40	
100PPBCN	A	Total CN	P	97.9323	µg/l	5/24/2022 10:08:41	
250PPBCN	A	Total CN	P	257.1975	µg/l	5/24/2022 10:08:42	
500PPBCN	A	Total CN	P	496.8065	µg/l	5/24/2022 10:08:43	
ICV1	S	Total CN	P	93.2088	µg/l	5/24/2022 12:32:11	
ICB1	S	Total CN	P	-0.4026	µg/l	5/24/2022 12:32:14	
CCV1	S	Total CN	P	240.4501	µg/l	5/24/2022 12:32:15	
CCB1	S	Total CN	P	-0.4294	µg/l	5/24/2022 12:32:18	
PB145054BL	S	Total CN	P	-0.4983	µg/l	5/24/2022 12:32:19	
N2923-03	S	Total CN	P	3.901	µg/l	5/24/2022 12:39:47	
N2923-03DUP	S	Total CN	P	4.3471	µg/l	5/24/2022 12:39:48	
N2923-03MS	S	Total CN	P	35.9188	µg/l	5/24/2022 12:39:51	
N2923-03MSD	S	Total CN	P	35.4806	µg/l	5/24/2022 12:39:52	
N2923-08	S	Total CN	P	2.0624	µg/l	5/24/2022 12:39:53	
N2923-13	S	Total CN	P	0.0165	µg/l	5/24/2022 12:39:54	
N2923-20	S	Total CN	P	1.5876	µg/l	5/24/2022 12:39:55	
N2924-01	S	Total CN	P	1.8996	µg/l	5/24/2022 12:39:56	
CCV2	S	Total CN	P	232.2672	µg/l	5/24/2022 12:47:21	
CCB2	S	Total CN	P	-0.5046	µg/l	5/24/2022 12:47:22	
N2924-06	S	Total CN	P	0.0138	µg/l	5/24/2022 12:47:23	
N2928-04	S	Total CN	P	-0.44	µg/l	5/24/2022 12:47:24	
N2928-05	S	Total CN	P	4.3751	µg/l	5/24/2022 12:47:25	
N2928-06	S	Total CN	P	0.4079	µg/l	5/24/2022 12:47:26	
N2928-07	S	Total CN	P	-0.5091	µg/l	5/24/2022 12:47:27	
N2928-08	S	Total CN	P	-0.0445	µg/l	5/24/2022 12:47:28	
N2928-09	S	Total CN	P	-0.5037	µg/l	5/24/2022 12:47:29	
PB145055BL	S	Total CN	P	-0.4471	µg/l	5/24/2022 12:47:30	
N2927-01	S	Total CN	P	-0.3432	µg/l	5/24/2022 12:54:53	
CCV3	S	Total CN	P	233.5283	µg/l	5/24/2022 12:54:54	
CCB3	S	Total CN	P	-0.4066	µg/l	5/24/2022 12:54:55	
N2927-01DUP	S	Total CN	P	-0.4232	µg/l	5/24/2022 12:54:56	
N2927-01MS	S	Total CN	P	35.1506	µg/l	5/24/2022 12:54:59	
N2927-01MSD	S	Total CN	P	34.7927	µg/l	5/24/2022 12:55:00	
N2927-02	S	Total CN	P	-0.4295	µg/l	5/24/2022 12:55:01	
N2927-03	S	Total CN	P	-0.6268	µg/l	5/24/2022 12:55:02	
N2927-04	S	Total CN	P	0.1483	µg/l	5/24/2022 12:55:03	
N2927-05	S	Total CN	P	-0.911	µg/l	5/24/2022 13:01:32	
N2927-06	S	Total CN	P	-0.7875	µg/l	5/24/2022 13:01:33	
N2927-11	S	Total CN	P	-0.7805	µg/l	5/24/2022 13:01:34	

1 of 2

N2927-16	S	Total CN	P	-0.664 µg/l	5/24/2022 13:01:35
CCV4	S	Total CN	P	243.3161 µg/l	5/24/2022 13:01:36
CCB4	S	Total CN	P	-0.3029 µg/l	5/24/2022 13:01:37
N2927-21	S	Total CN	P	0.4038 µg/l	5/24/2022 13:01:38
PB145054BS	S	Total CN	P	97.7846 µg/l	5/24/2022 13:30:57
PB145055BS	S	Total CN	P	97.3342 µg/l	5/24/2022 13:30:59
CCV5	S	Total CN	P	250.0274 µg/l	5/24/2022 13:31:00
CCB5	S	Total CN	P	-0.262 µg/l	5/24/2022 13:31:01

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : 12 Instrument ID : Konelab

5/24/2022 10:10


Test Total CN

Accepted 5/24/2022 10:10

Factor 1204
Bias 0.002

Coeff. of det. 0.999669

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.001	-0.8262	0.0000	
2	5.0PPBCN	0.005	4.2461	5.0000	-15.1
3	10PPBCN	0.009	9.3562	10.0000	-6.4
4	50PPBCN	0.043	50.2876	50.0000	0.6
5	100PPBCN	0.083	97.9323	100.0000	-2.1
6	250PPBCN	0.215	257.1975	250.0000	2.8
7	500PPBCN	0.414	496.8065	500.0000	-0.6

12

5/24/22