

LB120331


7471B

INSTRUMENT ID: CV2

Linear

 μ Abs.:

53147



A= 0.0000e+000

B= 1.8627e-004

C= 3.9675e-003

Rho= 0.9996135

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.017	0.017	72	0.000	72					-
0.2	0.200	0.206	0.006	1085	0.0 %	1085					3
2.5	2.500	2.487	-0.013	13328	0.0 %	13328					-1
5.0	5.000	4.886	-0.114	26209	0.0 %	26209					-2
7.5	7.500	7.701	0.201	41321	0.0 %	41321					3
10.0	10.000	9.903	-0.097	53147	0.0 %	53147					-1

LB120331 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 24 May 2022 15:55:51

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	72	-	0.0000	7471B	PPB	24 May 2022 15:58:04	S	Std
0.2 - 1	50.2	1085	-	0.2000	7471B	PPB	24 May 2022 16:00:21	S	Std
2.5 - 1	52.5	13328	-	2.5000	7471B	PPB	24 May 2022 16:02:39	S	Std
5.0 - 1	55	26209	-	5.0000	7471B	PPB	24 May 2022 16:04:55	S	Std
7.5 - 1	57.5	41321	-	7.5000	7471B	PPB	24 May 2022 16:07:12	S	Std
10.0 - 1	510	53147	-	10.0000	7471B	PPB	24 May 2022 16:09:29	S	Std
ICV43 - 1	ICV43	21149	3.9433	-	7471B	PPB	24 May 2022 16:12:17	U	SMPL
ICB43 - 1	ICB43	7	0.0053	-	7471B	PPB	24 May 2022 16:14:31	U	SMPL
CCV49 - 1	CCV49	26606	4.9598	-	7471B	PPB	24 May 2022 16:16:48	U	SMPL
CCB49 - 1	CCB49	-7	0.0027	-	7471B	PPB	24 May 2022 16:19:02	U	SMPL
CRA - 1	CRA	1191	0.2258	-	7471B	PPB	24 May 2022 16:21:19	U	SMPL
HighStd - 1	HighStd	53825	10.0298	-	7471B	PPB	24 May 2022 16:23:34	U	SMPL
ChkStd - 1	ChkStd	39084	7.2840	-	7471B	PPB	24 May 2022 16:25:48	U	SMPL
PB145059BL - 1	PBS	-114	-0.0173	-	7471B	PPB	24 May 2022 16:28:06	U	SMPL
PB145059BS - 1	LCSS	21986	4.0992	-	7471B	PPB	24 May 2022 16:30:23	U	SMPL
N2931-01 - 1	P009-SS003-0006-01	11813	2.2043	-	7471B	PPB	24 May 2022 16:32:39	U	SMPL
N2931-02 - 1	P009-SS003-0612-01	6306	1.1786	-	7471B	PPB	24 May 2022 16:34:56	U	SMPL
N2931-03 - 1	P009-SS003-1218-01	4984	0.9323	-	7471B	PPB	24 May 2022 16:37:14	U	SMPL
N2931-04 - 1	P009-SS003-1824-01	4643	0.8688	-	7471B	PPB	24 May 2022 16:39:30	U	SMPL
N2931-05 - 1	P009-SS003-2430-01	2993	0.5615	-	7471B	PPB	24 May 2022 16:41:45	U	SMPL
CCV50 - 1	CCV50	26682	4.9739	-	7471B	PPB	24 May 2022 16:44:00	U	SMPL
CCB50 - 1	CCB50	27	0.0090	-	7471B	PPB	24 May 2022 16:46:15	U	SMPL
N2931-05DUP - 1	P009-SS003-2430-01DUP	3075	0.5767	-	7471B	PPB	24 May 2022 16:48:32	U	SMPL
N2931-06 - 1	N2931-05MS	23900	4.4557	-	7471B	PPB	24 May 2022 16:50:47	U	SMPL
N2931-07 - 1	N2931-05MSD	24221	4.5155	-	7471B	PPB	24 May 2022 16:53:02	U	SMPL
N2931-14 - 1	P016-SS001-0006-01	4514	0.8448	-	7471B	PPB	24 May 2022 16:55:20	U	SMPL
N2931-15 - 1	P016-SS001-0612-01	4195	0.7854	-	7471B	PPB	24 May 2022 16:57:37	U	SMPL
N2931-16 - 1	P016-SS001-1218-01	2631	0.4940	-	7471B	PPB	24 May 2022 16:59:53	U	SMPL
N2931-17 - 1	P016-SS002-0006-01	6226	1.1637	-	7471B	PPB	24 May 2022 17:02:09	U	SMPL
N2982-01 - 1	P-1	1721	0.3245	-	7471B	PPB	24 May 2022 17:04:26	U	SMPL
N2982-02 - 1	P-2	1502	0.2837	-	7471B	PPB	24 May 2022 17:06:45	U	SMPL
N2982-05 - 1	P-5	1117	0.2120	-	7471B	PPB	24 May 2022 17:09:00	U	SMPL
CCV51 - 1	CCV51	26916	5.0175	-	7471B	PPB	24 May 2022 17:11:15	U	SMPL
CCB51 - 1	CCB51	38	0.0110	-	7471B	PPB	24 May 2022 17:13:31	U	SMPL
N2982-07 - 1	P-7	2316	0.4354	-	7471B	PPB	24 May 2022 17:15:48	U	SMPL
N3001-01 - 1	BU-03-052322	1895	0.3569	-	7471B	PPB	24 May 2022 17:18:03	U	SMPL
PB145060BL - 1	PBS	62	0.0155	-	7471B	PPB	24 May 2022 17:20:18	U	SMPL
PB145060BS - 1	LCSS	22240	4.1465	-	7471B	PPB	24 May 2022 17:22:34	U	SMPL
N2981-01 - 1	SB-13(0.5-1)	43621	8.1291	-	7471B	PPB	24 May 2022 17:24:49	U	SMPL
N2981-02 - 1	SB-13(3-3.5)	75954	14.1517	-	7471B	PPB	24 May 2022 17:27:07	U	SMPL
N2981-03 - 1	SB-15(0.5-1)	28385	5.2912	-	7471B	PPB	24 May 2022 17:29:25	U	SMPL
N2981-04 - 1	SB-15(4.5-5)	10660	1.9896	-	7471B	PPB	24 May 2022 17:31:44	U	SMPL
N2981-05 - 1	SB-14(0.5-1)	37139	6.9217	-	7471B	PPB	24 May 2022 17:34:02	U	SMPL
N2981-05DUP - 1	SB-14(0.5-1)DUP	36342	6.7733	-	7471B	PPB	24 May 2022 17:36:18	U	SMPL
CCV52 - 1	CCV52	26891	5.0129	-	7471B	PPB	24 May 2022 17:38:35	U	SMPL
CCB52 - 1	CCB52	-41	-0.0037	-	7471B	PPB	24 May 2022 17:40:52	U	SMPL
N2981-05MS - 1	SB-14(0.5-1)MS	63436	11.8200	-	7471B	PPB	24 May 2022 17:43:09	U	SMPL
N2981-05MSD - 1	SB-14(0.5-1)MSD	58758	10.9486	-	7471B	PPB	24 May 2022 17:45:25	U	SMPL
N2981-06 - 1	SB-14(3.5-4)	26977	5.0289	-	7471B	PPB	24 May 2022 17:47:42	U	SMPL
N2981-07 - 1	SB-12(0.5-1)	113113	21.0732	-	7471B	PPB	24 May 2022 17:50:00	U	SMPL
N2981-08 - 1	SB-12(2.5-3)	23668	4.4125	-	7471B	PPB	24 May 2022 17:52:16	U	SMPL
N2981-09 - 1	SB-11(0.5-1)	8661	1.6172	-	7471B	PPB	24 May 2022 17:54:35	U	SMPL
N2981-10 - 1	SB-11(2-2.5)	26814	4.9985	-	7471B	PPB	24 May 2022 17:56:53	U	SMPL
N2981-11 - 1	SB-10(0.5-1)	21310	3.9733	-	7471B	PPB	24 May 2022 17:59:10	U	SMPL
N2981-12 - 1	SB-10(1.5-2)	38703	7.2131	-	7471B	PPB	24 May 2022 18:01:28	U	SMPL
N2981-13 - 1	SB-9(0.5-1)	33885	6.3156	-	7471B	PPB	24 May 2022 18:03:44	U	SMPL
CCV53 - 1	CCV53	26679	4.9734	-	7471B	PPB	24 May 2022 18:06:01	U	SMPL
CCB53 - 1	CCB53	18	0.0073	-	7471B	PPB	24 May 2022 18:08:18	U	SMPL
N2981-14 - 1	SB-9(4.5-5)	9680	1.8070	-	7471B	PPB	24 May 2022 18:10:35	U	SMPL
N2981-15 - 1	SB-8(0.5-1)	26485	4.9372	-	7471B	PPB	24 May 2022 18:12:50	U	SMPL
N2981-16 - 1	SB-8(4.4-5)	3041	0.5704	-	7471B	PPB	24 May 2022 18:15:06	U	SMPL
N2981-17 - 1	SB-7(0.5-1)	2623	0.4925	-	7471B	PPB	24 May 2022 18:17:23	U	SMPL
N2981-18 - 1	SB-7(3.5-4)	623	0.1200	-	7471B	PPB	24 May 2022 18:19:38	U	SMPL
N2981-19 - 1	SB-6(0.5-1)	11368	2.1214	-	7471B	PPB	24 May 2022 18:21:54	U	SMPL
N2981-20 - 1	SB-6(3-3.5)	47829	8.9129	-	7471B	PPB	24 May 2022 18:24:11	U	SMPL
N2981-02DLX5 - 1		14818	2.7641	-	7471B	PPB	24 May 2022 18:26:27	U	SMPL
N2981-07DLX5 - 1		23196	4.3246	-	7471B	PPB	24 May 2022 18:28:46	U	SMPL
CCV54 - 1	CCV54	26877	5.0103	-	7471B	PPB	24 May 2022 18:31:01	U	SMPL
CCB54 - 1	CCB54	-113	-0.0171	-	7471B	PPB	24 May 2022 18:33:18	U	SMPL
N2931-05LX5 - 1		534	0.1034	-	7471B	PPB	24 May 2022 18:35:36	U	SMPL
N2931-05A - 1		25156	4.6897	-	7471B	PPB	24 May 2022 18:37:52	U	SMPL
N2981-05LX5 - 1		7167	1.3389	-	7471B	PPB	24 May 2022 18:40:08	U	SMPL
N2981-05A - 1		58377	10.8777	-	7471B	PPB	24 May 2022 18:42:25	U	SMPL
N2981-05DLX2 - 1		17302	3.2268	-	7471B	PPB	24 May 2022 18:44:41	U	SMPL
N2981-05DUPDLX2 - 1		18045	3.3652	-	7471B	PPB	24 May 2022 18:46:59	U	SMPL
N2981-05MSDLX2 - 1		30936	5.7663	-	7471B	PPB	24 May 2022 18:49:17	U	SMPL
N2981-05MSDDLX2 - 1		30008	5.5935	-	7471B	PPB	24 May 2022 18:51:34	U	SMPL
N2981-05LDLX2 - 1		0	0.0040	-	7471B	PPB	24 May 2022 18:53:53	U	SMPL
N2981-05LDLX10 - 1		3703	0.6937	-	7471B	PPB	24 May 2022 18:55:46	U	SMPL
N2981-05ADLX2 - 1		29416	5.4832	-	7471B	PPB	24 May 2022 18:58:02	U	SMPL
CCV55 - 1	CCV55	26611	4.9607	-	7471B	PPB	24 May 2022 19:00:19	U	SMPL
CCB55 - 1	CCB55	-57	-0.0066	-	7471B	PPB	24 May 2022 19:02:36	U	SMPL