

Clear table

Instrument ID: IC-2 Analyst: IZ Method: 300.0 / 9056A

ident	concentra tion F-	concentratio n CL-	concentrati on NO2	concentrati on BR-	concentrati on NO3	concentrati on HPO4	concentrati on SO4	file name	date time	Initial wt/ Final Vol	Analyst
STD1	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2022-05-20	5/20/22 12:31		IZ/AP
STD2	0.4030	0.7070	0.6430	2.2040	0.5690	0.6390	3.4390	2022-05-20	5/20/22 13:01		IZ/AP
STD3	0.8690	1.2000	1.2220	4.0910	1.0200	1.9820	6.1420	2022-05-20	5/20/22 13:30		IZ/AP
STD4	0.9690	1.4990	1.5220	5.0820	1.2590	2.7440	7.5670	2022-05-20	5/20/22 14:00		IZ/AP
STD5	1.9600	2.8700	2.9000	9.5790	2.3880	5.2900	14.2330	2022-05-20	5/20/22 14:29		IZ/AP
STD6	3.9530	5.9280	5.9360	19.7010	4.9240	10.0110	29.5820	2022-05-20	5/20/22 14:59		IZ/AP
STD7	5.0460	7.5960	7.5770	25.3420	6.3400	12.3340	38.0360	2022-05-20	5/20/22 15:29		IZ/AP
ICB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2022-05-20	5/20/22 15:58		IZ/AP
STD2	0.4690	0.6510	0.6890	2.2190	0.5640	0.7920	3.5120	2022-05-20	5/20/22 16:29		IZ/AP
ICV	1.8800	2.8300	2.8580	9.5100	2.3570	5.1280	13.9840	2022-05-20	5/20/22 16:59		IZ/AP
ICB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2022-05-20	5/20/22 17:29		IZ/AP
CCV	1.8810	2.8850	2.7590	9.5840	2.3700	5.0290	14.2260	2022-05-24	5/24/22 9:51		IZ/AP
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2022-05-24	5/24/22 10:22		IZ/AP
LB120335BLW	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2022-05-24	5/24/22 10:52		IZ/AP
LB120335BSW	1.8290	2.8970	2.8740	9.6770	2.3980	5.1770	14.3340	2022-05-24	5/24/22 11:21		IZ/AP
N3023-01	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2022-05-24	5/24/22 11:54		IZ/AP
N3023-01MS	2.1090	3.1900	2.9000	10.0310	2.5520	5.8200	14.9560	2022-05-24	5/24/22 12:24		IZ/AP
N3023-01MSD	1.9660	3.0890	2.8690	9.7090	2.4790	5.7050	14.4720	2022-05-24	5/24/22 12:54		IZ/AP
N3026-01	0.6160	303.0160	0.0000	0.0000	0.0000	0.0000	5.2110	2022-05-24	5/24/22 16:27		IZ/AP
CCV	1.8270	2.8870	2.8280	9.5790	2.3720	5.4900	14.1440	2022-05-24	5/24/22 16:57		IZ/AP
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2022-05-24	5/24/22 17:26		IZ/AP

12
6/15/22

Clear table

Instrument ID: IC-2 Analyst: IZ Method: 300.0 / 9056A


ident	concentra tion F-	concentratio n CL-	concentratio on NO2	concentratio on BR-	concentratio on NO3	concentratio on HPO4	concentratio on SO4	file name	date time	Initial wt/ Final Vol	Analyst
STD1	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2022-05-20_	5/20/22 12:31		IZ/AP
STD2	0.4690	0.6510	0.6890	2.2190	0.5640	0.7920	3.5120	2022-05-20_	5/20/22 16:29		IZ/AP
STD3	0.8690	1.2000	1.2220	4.0910	1.0200	1.9820	6.1420	2022-05-20_	5/20/22 13:30		IZ/AP
STD4	0.9690	1.4990	1.5220	5.0820	1.2590	2.7440	7.5670	2022-05-20_	5/20/22 14:00		IZ/AP
STD5	1.9600	2.8700	2.9000	9.5790	2.3880	5.2900	14.2330	2022-05-20_	5/20/22 14:29		IZ/AP
STD6	3.9530	5.9280	5.9360	19.7010	4.9240	10.0110	29.5820	2022-05-20_	5/20/22 14:59		IZ/AP
STD7	5.0460	7.5960	7.5770	25.3420	6.3400	12.3340	38.0360	2022-05-20_	5/20/22 15:29		IZ/AP

ident	True Value F-	True Value CL-	True Value NO2	True Value BR-	True Value NO3	True Value HPO4	True Value SO4
STD1	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000
STD2	0.4000	0.6000	0.6000	2.0000	0.5000	1.0000	3.0000
STD3	0.8000	1.2000	1.2000	4.0000	1.0000	2.0000	6.0000
STD4	1.0000	1.5000	1.5000	5.0000	1.2500	2.5000	7.5000
STD5	2.0000	3.0000	3.0000	10.0000	2.5000	5.0000	15.0000
STD6	4.0000	6.0000	6.0000	20.0000	5.0000	10.0000	30.0000
STD7	5.0000	7.5000	7.5000	25.0000	6.2500	12.5000	37.0000

ident	Relative Error F-	Relative Error CL-	Relative Error NO2	Relative Error BR-	Relative Error NO3	Relative Error HPO4	Relative Error SO4
STD1							
STD2	17.2500	8.5000	14.8333	10.9500	12.8000	-20.8000	17.0667
STD3	8.6250	0.0000	1.8333	2.2750	2.0000	-0.9000	2.3667
STD4	-3.1000	-0.0667	1.4667	1.6400	0.7200	9.7600	0.8933
STD5	-2.0000	-4.3333	-3.3333	-4.2100	-4.4800	5.8000	-5.1133
STD6	-1.1750	-1.2000	-1.0667	-1.4950	-1.5200	0.1100	-1.3933
STD7	0.9200	1.2800	1.0267	1.3680	1.4400	-1.3280	2.8000

CALIBRATION OF COMPONENT F-

Method: AnionsIC2-052022.mtw
 Equation: $Q = 0.0634558 \cdot A + 2.30438$
 RSD: 2.467 %
 Correlation coefficient: 0.999668




K3 = 0 K2 = 0 K1 = 0.0634558 K0 = 2.30438
 Base: Area
 Ref.channel: ch1
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	4.953	90.6	0.4	20	5.653		
3	11.27	237.7	0.8	20	5.653		
4	14.61	269.1	1	20	5.653		
5	31.03	581.3	2	20	5.653		
6	70.42	1210	4	20	5.653		
7	92.27	1554	5	20	5.653		

CALIBRATION OF COMPONENT CL-

Method: AnionsIC2-052022.mtw
 Equation: $Q = 0.111786 \cdot A + 1.71724$
 RSD: 3.138 %
 Correlation coefficient: 0.999463




K3 = 0 K2 = 0 K1 = 0.111786 K0 = 1.71724
 Base: Area
 Ref.channel: ch1
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used File
1	0	0	0	0	0	0
2	7.34	111.2	0.6	20	8.643	
3	14.68	199.4	1.2	20	8.643	
4	18.7	252.9	1.5	20	8.643	
5	36.23	498	3	20	8.643	
6	78.49	1045	6	20	8.643	
7	102	1344	7.5	20	8.643	

CALIBRATION OF COMPONENT NO2

Method: AnionsIC2-052022.mtw
 Equation: $Q = 0.0485615 \cdot A + 0.176478$
 RSD: 2.292 %
 Correlation coefficient: 0.999714




K3 = 0 K2 = 0 K1 = 0.0485615 K0 = 0.176478
 Base: Area
 Ref.channel: ch1
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	13.77	261.1	0.6	20	10.45		
3	27.88	499.8	1.2	20	10.45		
4	35.21	623.2	1.5	20	10.45		
5	68.66	1191	3	20	10.45		
6	141.8	2441	6	20	10.45		
7	177.7	3117	7.5	20	10.45		

CALIBRATION OF COMPONENT BR-

Method: AnionsIC2-052022.mtw
 Equation: $Q = 0.250936 \cdot A + 8.0433$
 RSD: 3.016 %
 Correlation coefficient: 0.999504



K3 = 0 K2 = 0 K1 = 0.250936 K0 = 8.0433

Base: Area

Ref.channel: chl

ISTD:


Formula: Linear

Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	7.334	143.6	2	20	13.53		
3	15.16	294	4	20	13.53		
4	19.37	373	5	20	13.53		
5	38.08	731.4	10	20	13.53		
6	79.7	1538	20	20	13.53		
7	102	1988	25	20	13.53		

CALIBRATION OF COMPONENT NO3

Method: AnionsIC2-052022.mtw
 Equation: $Q = 0.0435814 \cdot A + 2.16289$
 RSD: 3.235 %
 Correlation coefficient: 0.999429



K3 = 0 K2 = 0 K1 = 0.0435814 K0 = 2.16289

Base: Area

Ref.channel: ch1

ISTD:

Formula: Linear


Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	8.768	211.6	0.5	20	15.58		
3	18.15	418.6	1	20	15.58		
4	23.25	527.9	1.25	20	15.58		
5	46.2	1046	2.5	20	15.58		
6	98.8	2210	5	20	15.58		
7	128.6	2860	6.25	20	15.58		

±□;□

CALIBRATION OF COMPONENT HPO4

Method: AnionsIC2-052022.mtw
 Equation: $Q = 0.0639104 \cdot A - 12.1493$
 RSD: 4.999 %
 Correlation coefficient: 0.998637



K3 = 0 K2 = 0 K1 = 0.0639104 K0 = -12.1493

Base: Area

Ref.channel: ch1

ISTD:


Formula: Linear

Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	8.923	389.9	1	20	20.18		
3	19.6	810.5	2	20	20.18		
4	26.09	1049	2.5	20	20.18		
5	48.92	1846	5	20	20.18		
6	94.23	3323	10	20	20.18		
7	117.8	4050	12.5	20	20.18		

CALIBRATION OF COMPONENT SO4

Method: AnionsIC2-052022.mtw
 Equation: $Q = 0.141832 \cdot A + 4.30727$
 RSD: 3.413 %
 Correlation coefficient: 0.999365



K3 = 0 K2 = 0 K1 = 0.141832 K0 = 4.30727
 Base: Area
 Ref.channel: ch1
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	2.653		85.4	0	20		23.09
2	14.87		454.6	3	20		23.09
3	28.56		835.8	6	20		23.09
4	35.91		1037	7.5	20		23.09
5	70.18		1977	15	20		23.09
6	149.8		4141	30	20		23.09
7	194.4		5333	37.5	20		23.09

Report date: 5/23/2022 10:34:11 AM
Printed by: wet


Ident: STD2
Analysis from: 5/20/2022 1:01:05 PM
File: _2022-05-20_

Last save: 5/20/2022 4:00:18 PM

Method: AnionsIC2-052022.mtw
Run operator: wet
Analysis number: 30442

Last save: 5/20/2022 3:26:

SAMPLE:
: IZ/AP
Vial number: 3
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.73	0.229	4.95	7.51	90.602	5.45
2	8.73	0.198	7.34	11.13	111.209	6.69
3	10.54	0.230	13.77	20.88	261.129	15.70
4	13.58	0.286	7.33	11.12	143.627	8.64
5	15.58	0.336	8.77	13.29	211.608	12.73
6	20.48	0.560	8.92	13.53	389.938	23.45
7	23.34	0.452	14.87	22.54	454.626	27.34
7	27.00	0.327	65.95	99.99	1662.739	100.00

This report has been created by IC Net
METROHM LTD

Report date: 5/23/2022 10:34:21 AM
Printed by: wet


Ident: STD3
Analysis from: 5/20/2022 1:30:43 PM
File: _2022-05-20_

Last save: 5/23/2022 10:16:03 AM

Method: AnionsIC2-052022.mtw
Run operator: wet
Analysis number: 30443

Last save: 5/20/2022 3:26:

SAMPLE:
: IZ/AP
Vial number: 4
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.73	0.218	11.27	8.33	237.688	7.21
2	8.70	0.192	14.68	10.85	199.360	6.05
3	10.48	0.228	27.88	20.61	499.846	15.17
4	13.49	0.283	15.16	11.20	293.972	8.92
5	15.44	0.330	18.15	13.42	418.550	12.70
6	20.38	0.530	19.60	14.49	810.466	24.59
7	23.23	0.435	28.56	21.11	835.764	25.36
7	27.00	0.316	135.29	100.00	3295.646	100.00

This report has been created by IC Net
METROHM LTD

Report date: 5/23/2022 10:34:28 AM
Printed by: wet


Ident: STD4
Analysis from: 5/20/2022 2:00:21 PM
File: _2022-05-20_

Last save: 5/23/2022 10:16:03 AM

Method: AnionsIC2-052022.mtw
Run operator: wet
Analysis number: 30444

Last save: 5/20/2022 3:26:

SAMPLE:
: IZ/AP
Vial number: 5
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.70	0.214	14.61	8.44	269.131	6.51
2	8.67	0.191	18.70	10.80	252.876	6.12
3	10.44	0.230	35.20	20.33	623.191	15.08
4	13.43	0.281	19.37	11.19	373.028	9.03
5	15.38	0.327	23.25	13.43	527.950	12.78
6	20.36	0.516	26.09	15.07	1048.797	25.38
7	23.22	0.430	35.91	20.74	1036.721	25.09
7	27.00	0.313	173.12	100.00	4131.695	100.00

This report has been created by IC Net
METROHM LTD

Report date: 5/23/2022 10:34:34 AM
Printed by: wet


Ident: STD5
Analysis from: 5/20/2022 2:29:59 PM
File: _2022-05-20_

Last save: 5/23/2022 10:16:03 AM

Method: AnionsIC2-052022.mtw
Run operator: wet
Analysis number: 30445

Last save: 5/20/2022 3:26:

SAMPLE:
: IZ/AP
Vial number: 6
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.72	0.224	31.03	9.15	581.328	7.39
2	8.68	0.195	36.23	10.68	498.046	6.33
3	10.44	0.234	68.65	20.23	1190.731	15.13
4	13.43	0.283	38.08	11.22	731.404	9.29
5	15.34	0.328	46.19	13.61	1046.128	13.29
6	20.38	0.490	48.92	14.42	1845.580	23.45
7	23.27	0.422	70.18	20.69	1976.656	25.12
7	27.00	0.311	339.28	100.00	7869.874	100.00

This report has been created by IC Net
METROHM LTD

Report date: 5/23/2022 10:34:41 AM
Printed by: wet


Ident: STD6
Analysis from: 5/20/2022 2:59:38 PM
File: _2022-05-20_

Last save: 5/23/2022 10:16:03 AM

Method: AnionsIC2-052022.mtw
Run operator: wet
Analysis number: 30446

Last save: 5/20/2022 3:26:

SAMPLE:
: IZ/AP
Vial number: 7
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.69	0.206	70.41	9.87	1209.516	7.60
2	8.63	0.187	78.49	11.01	1045.212	6.57
3	10.36	0.236	141.76	19.88	2440.985	15.34
4	13.34	0.284	79.70	11.18	1538.183	9.67
5	15.20	0.322	98.80	13.85	2210.244	13.89
6	20.25	0.462	94.23	13.21	3322.949	20.89
7	23.15	0.413	149.81	21.00	4140.989	26.03
7	27.00	0.301	713.21	100.00	15908.078	100.00

This report has been created by IC Net
METROHM LTD

Report date: 5/23/2022 10:34:48 AM
Printed by: wet


Ident: STD7
Analysis from: 5/20/2022 3:29:16 PM
File: _2022-05-20_

Last save: 5/23/2022 10:16:04 AM

Method: AnionsIC2-052022.mtw
Run operator: wet
Analysis number: 30447

Last save: 5/20/2022 3:26:

SAMPLE:
: IZ/AP
Vial number: 8
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.75	0.204	92.27	10.09	1554.222	7.68
2	8.68	0.184	101.98	11.15	1343.599	6.64
3	10.42	0.243	177.66	19.42	3116.928	15.40
4	13.39	0.287	101.99	11.15	1987.769	9.82
5	15.23	0.320	128.61	14.06	2859.787	14.13
6	20.30	0.453	117.80	12.88	4049.828	20.00
7	23.22	0.411	194.38	21.25	5333.182	26.34
7	27.00	0.300	914.70	100.00	20245.315	100.00

This report has been created by IC Net
METROHM LTD

Report date: 5/23/2022 10:35:04 AM
Printed by: wet


Ident: ICB
Analysis from: 5/20/2022 3:58:53 PM
File: _2022-05-20_

Last save: 5/23/2022 10:16:04 AM

Method: AnionsIC2-052022.mtw
Run operator: wet
Analysis number: 30448

Last save: 5/20/2022 3:26:

SAMPLE:
: IZ/AP
Vial number: 9
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 5/23/2022 10:35:12 AM
Printed by: wet


Ident: STD2
Analysis from: 5/20/2022 4:29:54 PM
File: _2022-05-20_

Last save: 5/23/2022 10:16:04 AM

Method: AnionsIC2-052022.mtw
Run operator: wet
Analysis number: 30449

Last save: 5/20/2022 3:26:

SAMPLE:
: IZ/AP
Vial number: 3
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.67	0.235	4.99	7.39	111.396	6.37
2	8.59	0.192	7.24	10.71	101.158	5.78
3	10.34	0.228	14.09	20.85	279.945	16.00
4	13.29	0.283	7.52	11.12	144.778	8.28
5	15.22	0.331	8.90	13.16	209.175	11.96
6	20.29	0.563	10.03	14.84	437.907	25.03
7	23.14	0.463	14.82	21.93	464.869	26.58
7	27.00	0.328	67.58	100.00	1749.227	100.00

This report has been created by IC Net
METROHM LTD

Report date: 5/23/2022 10:35:19 AM
Printed by: wet


Ident: ICV
Analysis from: 5/20/2022 4:59:34 PM
File: _2022-05-20_

Last save: 5/20/2022 5:26:34 PM

Method: AnionsIC2-052022.mtw
Run operator: wet
Analysis number: 30450

Last save: 5/20/2022 4:56:

SAMPLE:
: IZ/AP
Vial number: 9
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.68	0.220	29.19	9.03	560.386	7.25
2	8.60	0.192	34.48	10.66	489.038	6.33
3	10.34	0.233	65.12	20.14	1177.305	15.23
4	13.28	0.285	36.87	11.40	726.099	9.39
5	15.16	0.329	44.68	13.82	1031.580	13.34
6	20.29	0.501	46.95	14.52	1803.583	23.33
7	23.18	0.438	66.06	20.43	1943.530	25.14
7	27.00	0.314	323.36	100.00	7731.520	100.00

This report has been created by IC Net
METROHM LTD

Report date: 5/23/2022 10:35:27 AM
Printed by: wet


Ident: ICB
Analysis from: 5/20/2022 5:29:12 PM
File: _2022-05-20_

Last save: 5/23/2022 10:25:43 AM

Method: AnionsIC2-052022.mtw
Run operator: wet
Analysis number: 30451

Last save: 5/20/2022 4:56:

SAMPLE:
: IZ/AP
Vial number: 10
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 5/25/2022 1:00:13 PM
Printed by: wet


Ident: CCV
Analysis from: 5/24/2022 9:51:21 AM
File: _2022-05-24_

Last save: 5/24/2022 10:18:21 AM

Method: AnionsIC2-052022.mtw
Run operator: wet
Analysis number: 30493

Last save: 5/24/2022 9:17:

SAMPLE:
: IZ/AP
Vial number: 11
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.80	0.246	26.18	8.70	560.561	7.26
2	8.89	0.211	33.01	10.97	498.786	6.46
3	10.74	0.256	60.80	20.21	1136.531	14.73
4	13.95	0.314	33.90	11.27	732.066	9.49
5	16.02	0.373	40.09	13.33	1037.700	13.45
6	20.51	0.527	44.93	14.94	1772.893	22.98
7	23.58	0.480	61.91	20.58	1977.639	25.63
7	27.00	0.344	300.82	100.00	7716.176	100.00

This report has been created by IC Net
METROHM LTD

Report date: 5/25/2022 1:00:38 PM
Printed by: wet


Ident: CCB
Analysis from: 5/24/2022 10:22:29 AM
File: _2022-05-24_

Last save: 5/25/2022 1:00:33 PM

Method: AnionsIC2-052022.mtw
Run operator: wet
Analysis number: 30494

Last save: 5/24/2022 9:17:

SAMPLE:
: IZ/AP
Vial number: 12
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 5/25/2022 1:00:55 PM
Printed by: wet


Ident: LB120335BLW
Analysis from: 5/24/2022 10:52:06 AM
File: _2022-05-24_

Last save: 5/25/2022 1:00:52 PM

Method: AnionsIC2-052022.mtw
Run operator: wet
Analysis number: 30495

Last save: 5/24/2022 9:17:

SAMPLE:
: IZ/AP
Vial number: 13
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 5/25/2022 1:01:02 PM
Printed by: wet


Ident: LB120335BSW
Analysis from: 5/24/2022 11:21:44 AM
File: _2022-05-24_

Last save: 5/24/2022 4:38:42 PM

Method: AnionsIC2-052022.mtw
Run operator: wet
Analysis number: 30496

Last save: 5/24/2022 9:17:

SAMPLE:
:
Vial number: 14
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.75	0.242	26.15	8.57	544.377	6.95
2	8.79	0.209	33.09	10.84	501.027	6.40
3	10.60	0.253	61.83	20.25	1183.811	15.12
4	13.74	0.311	34.43	11.28	739.443	9.44
5	15.75	0.367	40.93	13.41	1050.304	13.41
6	20.39	0.522	46.45	15.21	1818.972	23.23
7	23.44	0.478	62.43	20.45	1992.919	25.45
7	27.00	0.340	305.29	100.00	7830.853	100.00

This report has been created by IC Net
METROHM LTD

Report date: 5/25/2022 1:01:12 PM
Printed by: wet


Ident: N3023-01
Analysis from: 5/24/2022 11:54:59 AM
File: _2022-05-24_

Last save: 5/24/2022 2:06:58 PM

Method: AnionsIC2-052022.mtw
Run operator: wet
Analysis number: 30497

Last save: 5/24/2022 9:17:

SAMPLE:
: IZ/AP
Vial number: 22
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 5/25/2022 1:01:20 PM
Printed by: wet


Ident: N3023-01MS
Analysis from: 5/24/2022 12:24:37 PM
File: _2022-05-24_

Last save: 5/24/2022 12:51:37 PM

Method: AnionsIC2-052022.mtw
Run operator: wet
Analysis number: 30498

Last save: 5/24/2022 9:17:

SAMPLE:
: IZ/AP
Vial number: 23
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.77	0.250	29.63	8.94	632.157	7.55
2	8.80	0.205	37.49	11.30	553.699	6.62
3	10.61	0.248	63.87	19.26	1194.372	14.27
4	13.74	0.302	36.91	11.13	767.680	9.17
5	15.74	0.356	45.05	13.58	1121.136	13.40
6	20.46	0.515	51.84	15.63	2018.813	24.12
7	23.52	0.465	66.84	20.15	2080.506	24.86
7	27.00	0.334	331.64	100.00	8368.364	100.00

This report has been created by IC Net
METROHM LTD

Report date: 5/25/2022 1:01:27 PM
Printed by: wet


Ident: N3023-01MSD
Analysis from: 5/24/2022 12:54:15 PM
File: _2022-05-24_

Last save: 5/24/2022 1:21:15 PM

Method: AnionsIC2-052022.mtw
Run operator: wet
Analysis number: 30499

Last save: 5/24/2022 9:17:

SAMPLE:
: IZ/AP
Vial number: 24
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.75	0.249	27.55	8.57	587.399	7.23
2	8.76	0.204	36.26	11.28	535.578	6.59
3	10.56	0.247	62.46	19.43	1181.619	14.54
4	13.68	0.301	35.68	11.10	742.004	9.13
5	15.66	0.355	43.57	13.56	1087.483	13.38
6	20.37	0.513	51.06	15.89	1982.862	24.39
7	23.43	0.464	64.78	20.16	2012.321	24.75
7	27.00	0.333	321.36	100.00	8129.266	100.00

This report has been created by IC Net
METROHM LTD

Report date: 5/25/2022 1:02:04 PM
Printed by: wet


Ident: N3026-01
Analysis from: 5/24/2022 4:27:28 PM
File: _2022-05-24_

Last save: 5/25/2022 1:02:02 PM

Method: AnionsIC2-052022.mtw
Run operator: wet
Analysis number: 30501

Last save: 5/24/2022 9:17:

SAMPLE:
: IZ/AP
Vial number: 25
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.81	0.415	4.73	0.29	164.414	0.30
2	9.17	0.543	1605.38	98.42	54386.713	98.42
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	0.00	0.000	0.00	0.00	0.000	0.00
6	0.00	0.000	0.00	0.00	0.000	0.00
7	23.68	0.488	20.94	1.28	707.646	1.28
7	27.00	0.207	1631.05	100.00	55258.773	100.00

This report has been created by IC Net
METROHM LTD

Report date: 5/25/2022 1:02:15 PM
Printed by: wet


Ident: CCV
Analysis from: 5/24/2022 4:57:12 PM
File: _2022-05-24_

Last save: 5/25/2022 9:12:56 AM

Method: AnionsIC2-052022.mtw
Run operator: wet
Analysis number: 30502

Last save: 5/24/2022 9:17:

SAMPLE:
: IZ/AP
Vial number: 11
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.76	0.239	26.08	8.40	543.747	6.92
2	8.81	0.205	33.32	10.73	499.263	6.35
3	10.64	0.248	61.88	19.92	1164.661	14.82
4	13.79	0.306	34.48	11.10	731.654	9.31
5	15.82	0.362	40.88	13.16	1038.436	13.21
6	20.40	0.510	50.06	16.11	1916.133	24.38
7	23.47	0.459	63.94	20.58	1966.086	25.01
7	27.00	0.333	310.64	100.00	7859.979	100.00

This report has been created by IC Net
METROHM LTD

Report date: 5/25/2022 1:03:27 PM
Printed by: wet


Ident: CCB
Analysis from: 5/24/2022 5:26:49 PM
File: _2022-05-24_

Last save: 5/25/2022 1:03:23 PM

Method: AnionsIC2-052022.mtw
Run operator: wet
Analysis number: 30503

Last save: 5/24/2022 9:17:

SAMPLE:
: IZ/AP
Vial number: 12
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD