



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB120337

Analyst By : jignesh

Supervisor Review By : mohammad

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	05/24/2022	16:10
2	CAL2	1	Water	NA	NA	20.6	7.00	05/24/2022	16:11
3	CAL3	1	Water	NA	NA	20.6	10.03	05/24/2022	16:12
4	ICV	1	Water	NA	NA	20.7	6.98	05/24/2022	16:15
5	CCV1	1	Water	NA	NA	20.7	2.01	05/24/2022	16:20
6	N3023-01	1	Water	NA	NA	20.1	5.64	05/24/2022	16:22
7	N3024-03	1	Water	NA	NA	20.1	8.40	05/24/2022	16:25
8	N3024-04	1	Water	NA	NA	20.7	6.10	05/24/2022	16:29
9	N3024-05	1	Water	NA	NA	20.7	6.11	05/24/2022	16:30
10	N3024-06	1	Water	NA	NA	20.4	6.47	05/24/2022	16:32
11	N3024-07	1	Water	NA	NA	20.6	7.80	05/24/2022	16:35
12	N3024-08	1	Water	NA	NA	20.1	7.43	05/24/2022	16:36
13	N3024-09	1	Water	NA	NA	20.1	7.92	05/24/2022	16:44
14	N3026-01	1	Water	NA	NA	20.1	7.23	05/24/2022	16:45
15	N3026-01DUP	1	Water	NA	NA	20.2	7.24	05/24/2022	16:46
16	CCV2	1	Water	NA	NA	20.5	12.03	05/24/2022	16:47

WORKLIST(Hardcopy Internal Chain)

VB 20337

WorkList Name: ph n3026 WorkList ID: 159942 Department: Wet-Chemistry Date: 05-24-2022 15:59:00

Due Date	Matrix	Sample	Test	Raw Sample			Customer Sample	Collect Date	Method
				Preservative	Customer	Storage Location			
04/23/2022	Water	N3023-01	pH	Cool 4 deg C	STER12	N11	A-B-CELL-POOL-WATER-2ND-Q	04/23/2022	9040C
05/24/2022	Water	N3024-03	pH	Cool 4 deg C	PSEG03	N11	31374	05/24/2022	9040C
05/24/2022	Water	N3024-04	pH	Cool 4 deg C	PSEG03	N11	30987	05/24/2022	9040C
05/24/2022	Water	N3024-05	pH	Cool 4 deg C	PSEG03	N11	31793	05/24/2022	9040C
05/24/2022	Water	N3024-06	pH	Cool 4 deg C	PSEG03	N11	31794	05/24/2022	9040C
05/24/2022	Water	N3024-07	pH	Cool 4 deg C	PSEG03	N11	31532AQ	05/24/2022	9040C
05/24/2022	Water	N3024-08	pH	Cool 4 deg C	PSEG03	N11	31533AQ	05/24/2022	9040C
05/24/2022	Water	N3024-09	pH	Cool 4 deg C	PSEG03	N11	31534AQ	05/24/2022	9040C
05/24/2022	Water	N3026-01	pH	Cool 4 deg C	PSEG03	N48967		05/24/2022	9040C

Date/Time 05/24/22 16:00
 Raw Sample Received by: JRC (SM)
 Raw Sample Relinquished by: JRC (SM)

Date/Time 05/24/22 17:30
 Raw Sample Received by: JRC (SM)
 Raw Sample Relinquished by: JRC (SM)