



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: mohammad

Run Number: LB120345

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.3	05/25/2022	11:40
2	N3024-03		1	15	100.00	>212.0	>212.0	05/25/2022	12:05
3	N3024-03DUP		1	15	100.00	>212.0	>212.0	05/25/2022	12:35
4	N3024-04		1	12	100.00	>212.0	>212.0	05/25/2022	12:55
5	N3024-10		1	18	100.00	>212.0	>212.0	05/25/2022	13:20
6	N3024-11		1	14	100.00	>212.0	>212.0	05/25/2022	13:55
7	N3026-01		1	18	100.00	>212.0	>212.0	05/25/2022	14:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB120345

WorkList Name : fp52522 WorkList ID : 159971 Department : Wet-Chemistry Date : 05-25-2022 11:23:44

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample		Collect Date	Method
						Storage Location	Customer Sample		
05/31/2022	Water	N3024-03	Flash Point	Cool 4 deg C	PSEG03	N11	31374	05/24/2022	1010B
05/31/2022	Water	N3024-04	Flash Point	Cool 4 deg C	PSEG03	N11	30987	05/24/2022	1010B
05/31/2022	Water	N3024-10	Flash Point	Cool 4 deg C	PSEG03	N11	31793-31794	05/24/2022	1010B
05/31/2022	Water	N3024-11	Flash Point	Cool 4 deg C	PSEG03	N11	31532-34AQ	05/24/2022	1010B
05/27/2022	Water	N3026-01	Flash Point	Cool 4 deg C	PSEG03		N48967	05/24/2022	1010B

Date/Time 05/25/22 11:25
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: 12 (WC)

Date/Time 05/25/22 14:20
 Raw Sample Received by: 12 (WC)
 Raw Sample Relinquished by: KS (WC)