



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: mohammad

Run Number: LB120346

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.3	05/25/2022	11:40
2	N3024-01		1	8	35.00	95.0	94.7	05/25/2022	14:40
3	N3024-01DUP		1	9	36.00	96.8	96.5	05/25/2022	15:05

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 120396

WorkList Name : fp52522-1 WorkList ID : 159972 Department : Wet-Chemistry Date : 05-25-2022 11:25:09

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/31/2022	Solid	N3024-01	Flash Point	Cool 4 deg C	PSEG03	N11	40730	05/24/2022	1010B

Date/Time 05/25/22 11:30
Raw Sample Received by: KS(wc)
Raw Sample Relinquished by: KS(wc)

Date/Time 05/25/22 15:10
Raw Sample Received by: KS(wc)
Raw Sample Relinquished by: KS(wc)