

LB120347


7471B

INSTRUMENT ID: CVZ

Linear

 μ Abs.:

55964



A= 0.0000e+000

B= 1.7877e-004

C= -5.0495e-003

Rho= 0.9999777

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	CV (%)
0.0	0.000	0.012	0.012	93	0.000	93					
0.2	0.200	0.179	-0.021	1030	0.0 %	1030					-11
2.5	2.500	2.490	-0.010	13958	0.0 %	13958					0
5.0	5.000	5.047	0.047	28260	0.0 %	28260					1
7.5	7.500	7.472	-0.028	41827	0.0 %	41827					0
10.0	10.000	10.000	-0.000	55964	0.0 %	55964					0

LB120347 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 25 May 2022 12:13:18

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	93	-	0.00007471B	PPB		25 May 2022 12:15:59	S	Std
0.2 - 1	50.2	1030	-	0.20007471B	PPB		25 May 2022 12:18:15	S	Std
2.5 - 1	52.5	13958	-	2.50007471B	PPB		25 May 2022 12:20:33	S	Std
5.0 - 1	55	28260	-	5.00007471B	PPB		25 May 2022 12:22:50	S	Std
7.5 - 1	57.5	41827	-	7.50007471B	PPB		25 May 2022 12:25:07	S	Std
10.0 - 1	57.5	55964	-	10.00007471B	PPB		25 May 2022 12:27:24	S	Std
ICV45 - 1	ICV45	21817	3.8952	-7471B	PPB		25 May 2022 12:30:05	U	SMPL
ICB45 - 1	ICB45	-39	-0.0120	-7471B	PPB		25 May 2022 12:32:21	U	SMPL
CCV60 - 1	CCV60	28460	5.0828	-7471B	PPB		25 May 2022 12:34:36	U	SMPL
CCB60 - 1	CCB60	-62	-0.0161	-7471B	PPB		25 May 2022 12:36:52	U	SMPL
CRA - 1	CRA	957	0.1660	-7471B	PPB		25 May 2022 12:39:08	U	SMPL
HighStd - 1	HighStd	56451	10.0867	-7471B	PPB		25 May 2022 12:41:22	U	SMPL
ChkStd - 1	ChkStd	39596	7.0736	-7471B	PPB		25 May 2022 12:43:37	U	SMPL
PB145083BL - 1	PBS	-81	-0.0195	-7471B	PPB		25 May 2022 12:45:55	U	SMPL
PB145083BS - 1	LCSS	20920	3.7348	-7471B	PPB		25 May 2022 12:48:13	U	SMPL
N2990-15 - 1	P028-SS003-0006-01	4325	0.7681	-7471B	PPB		25 May 2022 12:50:29	U	SMPL
N3000-01 - 1	TP-01-COMP	58899	10.5244	-7471B	PPB		25 May 2022 12:52:46	U	SMPL
N3000-02 - 1	TP-02-COMP	24655	4.4025	-7471B	PPB		25 May 2022 12:55:03	U	SMPL
N3004-02 - 1	220512087-01	385	0.0638	-7471B	PPB		25 May 2022 12:57:22	U	SMPL
N3005-02 - 1	220512029-03	3522	0.6246	-7471B	PPB		25 May 2022 12:59:39	U	SMPL
CCV61 - 1	CCV61	28537	5.0965	-7471B	PPB		25 May 2022 13:01:54	U	SMPL
CCB61 - 1	CCB61	-12	-0.0072	-7471B	PPB		25 May 2022 13:04:10	U	SMPL
N3007-02 - 1	220511038-01	3998	0.7097	-7471B	PPB		25 May 2022 13:06:27	U	SMPL
N3009-02 - 1	220511033-04	7417	1.3209	-7471B	PPB		25 May 2022 13:08:42	U	SMPL
N3010-02 - 1	220511032-04	2216	0.3911	-7471B	PPB		25 May 2022 13:10:57	U	SMPL
N3013-02 - 1	220511032-01	2415	0.4267	-7471B	PPB		25 May 2022 13:13:13	U	SMPL
N3014-02 - 1	220511025-01	832	0.1437	-7471B	PPB		25 May 2022 13:15:29	U	SMPL
N3015-02 - 1	220511006-04	11758	2.0969	-7471B	PPB		25 May 2022 13:17:45	U	SMPL
N3016-02 - 1	220511006-03	6556	1.1670	-7471B	PPB		25 May 2022 13:20:01	U	SMPL
N3024-02 - 1	31532-34-SOL	310446	55.4937	-7471B	PPB		25 May 2022 13:22:20	U	SMPL
N3027-01 - 1	OR-03-052422	4420	0.7851	-7471B	PPB		25 May 2022 13:29:43	U	SMPL
N3027-01DUP - 1	OR-03-052422DUP	4475	0.7950	-7471B	PPB		25 May 2022 13:31:58	U	SMPL
CCV62 - 1	CCV62	28881	5.1580	-7471B	PPB		25 May 2022 13:34:13	U	SMPL
CCB62 - 1	CCB62	-27	-0.0099	-7471B	PPB		25 May 2022 13:36:29	U	SMPL
N3027-01MS - 1	OR-03-052422MS	24680	4.4070	-7471B	PPB		25 May 2022 13:38:46	U	SMPL
N3027-01MSD - 1	OR-03-052422MSD	26308	4.6981	-7471B	PPB		25 May 2022 13:41:01	U	SMPL
N3028-01 - 1	TP-1-1-5FT	85898	15.3510	-7471B	PPB		25 May 2022 13:43:17	U	SMPL
N3027-01LX5 - 1		774	0.1333	-7471B	PPB		25 May 2022 13:45:34	U	SMPL
N3027-01A - 1		27762	4.9580	-7471B	PPB		25 May 2022 13:47:52	U	SMPL
N3000-01DLX2 - 1		30012	5.3602	-7471B	PPB		25 May 2022 13:50:09	U	SMPL
N3024-02DLX10 - 1		33899	6.0551	-7471B	PPB		25 May 2022 13:52:26	U	SMPL
N3028-01DLX5 - 1		17179	3.0661	-7471B	PPB		25 May 2022 13:54:44	U	SMPL
CCV63 - 1	CCV63	28998	5.1789	-7471B	PPB		25 May 2022 13:57:03	U	SMPL
CCB63 - 1	CCB63	-24	-0.0093	-7471B	PPB		25 May 2022 13:59:19	U	SMPL