

**PERCENT SOLID**

Supervisor: mohammad

Analyst: jignesh

Date: 5/26/2022

OVENTEMP IN Celsius(°C): 107

Time IN: 17:00

In Date: 05/25/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103

Time OUT: 08:30

Out Date: 05/26/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

BalanceID: M SC-4

Thermometer ID: %SOLIDS-OVEN

QC:LB120350

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N3025-01	NORTH-SW	1	0.91	8.85	9.76	9.11	92.7	
N3025-02	SOUTH-SW	2	0.96	9.01	9.97	9.47	94.5	
N3025-03	EAST-SW	3	0.92	9.01	9.93	9.27	92.7	
N3025-04	WEST-SW	4	0.91	9.02	9.93	9.14	91.2	
N3025-05	UST-BOT	5	0.92	8.72	9.64	8.86	91.1	
N3032-01	SU-406-ROLL-OFF	6	0.9	8.79	9.69	8.06	81.5	
N3032-02	SU-406-ROLL-OFF-EPH-2	7	0.91	8.97	9.88	8.26	81.9	
N3041-01	2147	18	0.91	8.79	9.7	8.91	91.0	
N3051-01	72-11944	19	0.93	9.03	9.96	8.19	80.4	
N3051-02	72-11944	20	0.9	8.75	9.65	8.42	85.9	
N3053-01	52322-OIL	21	1.00	1.00	2.00	2.00	100.0	OIL SAMPLE
N3054-01	OL-VC-80071-04	8	0.91	8.96	9.87	3.71	31.3	sediment sample
N3054-02	OL-VC-80071-410	9	0.9	9.03	9.93	4.82	43.4	sediment sample
N3054-03	OL-VC-80040-04	10	0.89	8.92	9.81	3.9	33.7	sediment sample
N3054-04	OL-VC-80040-410	11	0.93	9.04	9.97	4.75	42.3	sediment sample
N3054-05	OL-STA-80238-04	12	0.91	9.04	9.95	3.82	32.2	sediment sample
N3054-06	OL-STA-80238-410	13	0.91	9.05	9.96	5.86	54.7	sediment sample
N3054-07	ST-51-04	14	0.92	8.87	9.79	3.16	25.3	sediment sample
N3054-08	ST-51-410	15	0.88	9.04	9.92	2.88	22.1	sediment sample
N3054-09	OL-VC-80079-04	16	0.91	8.81	9.72	3.23	26.3	sediment sample
N3054-10	OL-VC-80079-410	17	0.91	9.06	9.97	3.5	28.6	sediment sample
N3058-01	0508	22	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
N3058-02	31669	23	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE
N3058-03	0507	24	1.00	1.00	2.00	2.00	100.0	WIPE SAMPLE

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$