



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB120358

Analyst By : jignesh

Supervisor Review By : mohammad

Slope : 98.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	05/26/2022	08:25
2	CAL2	1	Water	NA	NA	20.6	7.01	05/26/2022	08:26
3	CAL3	1	Water	NA	NA	20.6	10.03	05/26/2022	08:27
4	ICV	1	Water	NA	NA	20.5	7.00	05/26/2022	08:30
5	CCV1	1	Water	NA	NA	20.5	2.01	05/26/2022	08:32
6	N3038-01	1	Water	NA	NA	20.4	7.28	05/26/2022	08:35
7	N3042-08	1	Water	NA	NA	21.6	6.59	05/26/2022	08:40
8	N3044-01	1	Water	NA	NA	20.2	5.11	05/26/2022	08:42
9	N3053-03	1	Water	NA	NA	20.9	5.04	05/26/2022	08:45
10	N3053-04	1	Water	NA	NA	20.8	5.01	05/26/2022	08:46
11	N3053-05	1	Water	NA	NA	20.8	5.01	05/26/2022	08:47
12	N3053-06	1	Water	NA	NA	20.9	5.01	05/26/2022	08:50
13	N3057-03	1	Water	NA	NA	20.9	6.89	05/26/2022	08:52
14	N3059-01	1	Water	NA	NA	20.6	6.47	05/26/2022	09:00
15	N3059-01DUP	1	Water	NA	NA	20.7	6.48	05/26/2022	09:01
16	CCV2	1	Water	NA	NA	20.9	12.03	05/26/2022	09:03

120356

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph b3069

WorkList ID : 159997

Department : Wet-Chemistry

Date : 05-26-2022 08:17:10

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/24/2022	Water	N3038-01	pH C	Cool 4 deg C	TETR06	N11	TT-045-JP/WGW-20220524 JP	05/24/2022	9040C
05/25/2022	Water	N3042-08	pH A	Cool 4 deg C	PSEG03	N11	304 ID	05/25/2022	9040C
05/24/2022	Water	N3044-01	pH A	Cool 4 deg C	STER06	N11	POOL-WATER	05/24/2022	9040C
05/25/2022	Water	N3053-03	pH C	Cool 4 deg C	PSEG03	N11	52322-E	05/25/2022	9040C
05/25/2022	Water	N3053-04	pH C	Cool 4 deg C	PSEG03	N11	52322-F	05/25/2022	9040C
05/25/2022	Water	N3053-05	pH C	Cool 4 deg C	PSEG03	N11	52322-G	05/25/2022	9040C
05/25/2022	Water	N3053-06	pH C	Cool 4 deg C	PSEG03	N11	52322-H	05/25/2022	9040C
05/25/2022	Water	N3057-03	pH A	Cool 4 deg C	PSEG03	N11	WATER-DRUM	05/25/2022	9040C
05/25/2022	Water	N3059-01	pH A	Cool 4 deg C	PSEG03	N11	4922A	05/25/2022	9040C

Date/Time 05/26/22 08:20
 Raw Sample Received by: JP (GSC)
 Raw Sample Relinquished by: CA SM

Date/Time 05/26/22 11:30
 Raw Sample Received by: CA SM
 Raw Sample Relinquished by: JP (GSC)