



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : mohammad

Run Number: LB120359

Slope : 98.7

BalanceID: WC SC-6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.9	4.01	05/25/2022	17:05
2	CAL2	1	Water	NA	NA	20.9	7.00	05/25/2022	17:06
3	CAL3	1	Water	NA	NA	20.9	10.03	05/25/2022	17:08
4	ICV	1	Water	NA	NA	20.8	6.99	05/25/2022	17:10
5	CCV1	1	Water	NA	NA	20.8	2.01	05/25/2022	17:12
6	N3041-02	1	Solid	20.02	20	20.6	6.86	05/25/2022	17:15
7	N3041-02DUP	1	Solid	20.03	20	20.7	6.87	05/25/2022	17:16
8	N3042-07	1	Solid	20.01	20	20.6	6.62	05/25/2022	17:19
9	CCV2	1	Water	NA	NA	20.8	12.03	05/25/2022	17:20

WORKLIST(Hardcopy Internal Chain)

100 120359

WorkList Name : corrosivity n3028 WorkList ID : 159983 Department : Wet-Chemistry Date : 05-25-2022 13:32:57

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/25/2022	Solid	N3041-02	Corrosivity	Cool 4 deg C	PSEG03	N11	2147	05/25/2022	9045D
05/25/2022	Solid	N3042-07	Corrosivity	Cool 4 deg C	PSEG03	N11	302	05/25/2022	9045D

Date/Time 05/25/22 16:50
 Raw Sample Received by: DPG
 Raw Sample Relinquished by: CP SM

Date/Time 05/25/22 18:00
 Raw Sample Received by: CP SM
 Raw Sample Relinquished by: 2D(ACC)