

Analysis Method: SM2510 BANALYST: IwonaParameter: ConductanceSUPERVISOR REVIEW BY: jigneshRun Number: LB120365Meter ID: WC CONDUCTANCE METER-1Thermometer ID: WC PH METER-1

Reagent/Standard	Lot/Log #
conductivity standard	WP95377

Seq	Lab ID	True Value (mg/L)	Dilution	Temperature °C	Reading (umho/cm)	Result (umho/cm)	Anal Date	Anal Time
1	LB120365BL		1	21.2	0.78	0.84	05/26/2022	11:10
2	LB120365BS	1413	1	21.4	1326.00	1423.91	05/26/2022	11:13
3	N3044-01		1	20.4	1.41	1.55	05/26/2022	11:16
4	N3044-01DUP		1	20.2	1.40	1.54	05/26/2022	11:19

Note : Conductivity result is corrected to 25°C

Result = $\frac{\text{(Reading)}}{1 + 0.0191 (\text{Temperature} - 25)}$
--

WORKLIST(Hardcopy Internal Chain)

LIB 120365

WorkList Name : CONDUCTANCE-052522 WorkList ID : 159974 Department : Wet-Chemistry Date : 05-25-2022 11:58:42

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/07/2022	Water	N3044-01	Conductance	Cool 4 deg C	STER06	N11	POOL-WATER	05/24/2022	SM2510 B

Date/Time 5/26/22 11:40
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 5/26/22 11:30
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]