



PERCENT SOLID

Supervisor: mohammad

Analyst: jignesh

Date: 5/26/2022

OVENTEMP IN Celsius(°C): 107

Time IN: 14:10

In Date: 05/25/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103

Time OUT: 08:10

Out Date: 05/26/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

BalanceID: M SC-4

Thermometer ID: %SOLIDS-OVEN

QC:LB120367

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N2841-01	P002-SS003-0006-01	1	0.93	9.02	9.95	8.63	85.4	
N2841-02	P002-SS003-0612-01	2	0.91	8.77	9.68	8.25	83.7	
N2841-03	P002-SS003-1218-01	3	0.93	8.85	9.78	8.23	82.5	
N2841-04	P002-SS004-0006-01	4	0.91	8.66	9.57	8.12	83.3	
N2841-05	P005-SS003-0006-01	5	0.88	9.08	9.96	7.97	78.1	
N2841-06	P005-SS003-0612-01	6	0.91	8.77	9.68	8.22	83.4	
N2841-07	P005-SS003-1218-01	7	0.81	8.82	9.63	8.1	82.7	
N2841-08	P005-SS003-1824-01	8	0.84	8.88	9.72	8.32	84.2	
N2841-09	P005-SS003-2430-01	9	0.96	8.60	9.56	8.41	86.6	
N2841-10	P005-SS004-0006-01	10	0.93	8.74	9.67	8.47	86.3	
N2841-11	P006-SS002-0006-01	11	0.91	9.03	9.94	8.59	85.0	
N2841-12	P006-SS002-0612-01	12	0.88	9.08	9.96	8.5	83.9	
N2841-13	P006-SS002-1218-01	13	0.93	8.72	9.65	7.6	76.5	
N2841-14	P006-SS002-1824-01	14	0.9	8.60	9.5	7.91	81.5	
N2841-15	P006-SS002-2430-01	15	0.91	8.70	9.61	8.36	85.6	
N2841-16	P009-SS003-0006-01	16	0.9	9.01	9.91	8.26	81.7	
N2841-17	P009-SS003-0612-01	17	0.93	8.60	9.53	8.25	85.1	
N2841-18	P009-SS003-1218-01	18	0.89	8.66	9.55	8.42	87.0	
N2841-19	P009-SS003-1824-01	19	0.9	8.74	9.64	8.66	88.8	
N2841-20	P009-SS003-2430-01	20	0.96	8.99	9.95	9.26	92.3	
N2841-21	N2841-20MS	21	0.96	8.99	9.95	9.26	92.3	
N2841-22	N2841-20MSD	22	0.96	8.99	9.95	9.26	92.3	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$