



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : mohammad

Run Number: LB120375

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	05/26/2022	17:22
2	CAL2	1	Water	NA	NA	20.4	7.01	05/26/2022	17:23
3	CAL3	1	Water	NA	NA	20.4	10.03	05/26/2022	17:24
4	ICV	1	Water	NA	NA	20.3	6.98	05/26/2022	17:25
5	CCV1	1	Water	NA	NA	20.5	2.01	05/26/2022	17:28
6	N3067-01	1	Solid	20.03	20	21.0	8.84	05/26/2022	17:30
7	N3076-03	1	Solid	20.02	20	21.9	6.55	05/26/2022	17:33
8	N3080-02	1	Solid	20.03	20	21.9	10.08	05/26/2022	17:37
9	N3080-05	1	Solid	20.01	20	20.8	8.80	05/26/2022	17:40
10	N3080-05DUP	1	Solid	20.02	20	20.9	8.81	05/26/2022	17:41
11	CCV2	1	Water	NA	NA	20.6	12.03	05/26/2022	17:42

VP 120345

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity n3080 WorkList ID : 160026 Department : Wet-Chemistry Date : 05-26-2022 16:09:55

Due Date	Matrix	Sample	Test	Raw Sample			Collect Date	Method
				Preservative	Customer	Storage Location		
05/26/2022	Solid	N3067-01	Corrosivity	Cool 4 deg C	PSEG03	N11 33686	05/26/2022	9045D
05/26/2022	Solid	N3076-03	Corrosivity	Cool 4 deg C	PSEG03	N11 RBR-671	05/26/2022	9045D
05/26/2022	Solid	N3080-02	Corrosivity	Cool 4 deg C	PSEG03	N11 WC-4	05/26/2022	9045D
05/26/2022	Solid	N3080-05	Corrosivity	Cool 4 deg C	PSEG03	N11 WC-5	05/26/2022	9045D

Date/Time 05/26/22 16:35
Raw Sample Received by: DS (WC)
Raw Sample Relinquished by: DS (WC)

Date/Time 05/26/22 18:00
Raw Sample Received by: DS (WC)
Raw Sample Relinquished by: DS (WC)