



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB120376

Analyst By : jignesh

Supervisor Review By : mohammad

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	05/26/2022	17:22
2	CAL2	1	Water	NA	NA	20.4	7.01	05/26/2022	17:23
3	CAL3	1	Water	NA	NA	20.4	10.03	05/26/2022	17:24
4	ICV	1	Water	NA	NA	20.3	6.98	05/26/2022	17:25
5	CCV1	1	Water	NA	NA	20.5	2.01	05/26/2022	17:28
6	N3067-02	1	Water	NA	NA	20.7	6.89	05/26/2022	17:50
7	N3071-01	1	Water	NA	NA	20.1	4.53	05/26/2022	17:55
8	N3071-01DUP	1	Water	NA	NA	20.2	4.54	05/26/2022	17:56
9	CCV2	1	Water	NA	NA	20.6	12.03	05/26/2022	18:00

120376

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH N3054

WorkList ID : 160028

Department : Wet-Chemistry

Date : 05-26-2022 17:39:07

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/26/2022	Water	N3067-02	pH C	Cool 4 deg C	PSEG03	N11	SUM-WATER-0526	05/26/2022	9040C
05/25/2022	Water	N3071-01	pH A	Cool 4 deg C	STER05	N11	POOL-WATER	05/25/2022	9040C

Date/Time 05/26/22 17:40
Raw Sample Received by: 7D(WC)
Raw Sample Relinquished by: KS(WC)

Date/Time 05/26/22 18:20
Raw Sample Received by: KS(WC)
Raw Sample Relinquished by: 7D(WC)