


LB120379

7470A

INSTRUMENT ID: CV1

 μ Abs.:

42111



Linear

A= 0.0000e+000

B= 2.3623e-004

C= -1.0532e-002

Rho= 0.9998196

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	✓/0/1
0.0	0.000	-0.009	-0.009	7	0.000	7					-
0.2	0.200	0.178	-0.022	798	0.0 %	798					-1
2.5	2.500	2.563	0.063	10895	0.0 %	10895					3
5.0	5.000	4.914	-0.086	20845	0.0 %	20845					-2
7.5	7.500	7.616	0.116	32286	0.0 %	32286					2
10.0	10.000	9.937	-0.063	42111	0.0 %	42111					-1

LB120379 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 27 May 2022 10:52:02

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	7	-	0.0000	7470A	PPB	27 May 2022 10:52:12	S	Std
0.2 - 1	50.2	798	-	0.2000	7470A	PPB	27 May 2022 10:54:29	S	Std
2.5 - 1	52.5	10895	-	2.5000	7470A	PPB	27 May 2022 10:56:46	S	Std
5.0 - 1	55	20845	-	5.0000	7470A	PPB	27 May 2022 10:59:02	S	Std
7.5 - 1	57.5	32286	-	7.5000	7470A	PPB	27 May 2022 11:01:19	S	Std
10.0 - 1	514	42111	-	10.0000	7470A	PPB	27 May 2022 11:03:35	S	Std
ICV48 - 1	ICV48	16540	3.8968	-	7470A	PPB	27 May 2022 11:06:47	U	SMPL
ICB48 - 1	ICB48	90	0.0107	-	7470A	PPB	27 May 2022 11:11:41	U	SMPL
CCV74 - 1	CCV74	19321	4.5537	-	7470A	PPB	27 May 2022 11:19:33	U	SMPL
CCB74 - 1	CCB74	-79	-0.0292	-	7470A	PPB	27 May 2022 11:21:48	U	SMPL
CRA - 1	CRA	760	0.1690	-	7470A	PPB	27 May 2022 11:24:03	U	SMPL
HighStd - 1	HighStd	41943	9.8978	-	7470A	PPB	27 May 2022 11:26:18	U	SMPL
ChkStd - 1	ChkStd	28236	6.6597	-	7470A	PPB	27 May 2022 11:28:33	U	SMPL
N3000-03 - 1	TP-02-COMP	43	-0.0004	-	7470A	PPB	27 May 2022 11:38:21	U	SMPL
N3028-02 - 1	TP-1-1-5FT	-2	-0.0110	-	7470A	PPB	27 May 2022 11:40:37	U	SMPL
N3041-02 - 1	2147	33	-0.0027	-	7470A	PPB	27 May 2022 11:42:53	U	SMPL
N3041-02DUP - 1	2147DUP	48	0.0008	-	7470A	PPB	27 May 2022 11:45:11	U	SMPL
N3041-02MS - 1	2147MS	16649	3.9225	-	7470A	PPB	27 May 2022 11:50:03	U	SMPL
CCV75 - 1	CCV75	20742	4.8894	-	7470A	PPB	27 May 2022 11:52:20	U	SMPL
CCB75 - 1	CCB75	-11	-0.0131	-	7470A	PPB	27 May 2022 11:54:35	U	SMPL
N3041-02MSD - 1	2147MSD	16599	3.9107	-	7470A	PPB	27 May 2022 11:56:50	U	SMPL
N3042-07 - 1	302	2	-0.0101	-	7470A	PPB	27 May 2022 11:59:06	U	SMPL
N3041-02LX5 - 1		-21	-0.0155	-	7470A	PPB	27 May 2022 12:01:21	U	SMPL
N3041-02A - 1		16932	3.9894	-	7470A	PPB	27 May 2022 12:03:37	U	SMPL
PB145139BL - 1	PBW	44	-0.0001	-	7470A	PPB	27 May 2022 12:05:52	U	SMPL
PB145139BS - 1	LCSW	16751	3.9466	-	7470A	PPB	27 May 2022 12:08:07	U	SMPL
PB145105TB - 1	PB145105TB	24	-0.0049	-	7470A	PPB	27 May 2022 12:10:23	U	SMPL
CCV75 - 1	CCV75	21249	5.0092	-	7470A	PPB	27 May 2022 12:12:39	U	SMPL
CCB75 - 1	CCB75	-17	-0.0145	-	7470A	PPB	27 May 2022 12:14:55	U	SMPL