



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB120388

Analyst By : jignesh

Supervisor Review By : mohammad

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.00	05/27/2022	13:10
2	CAL2	1	Water	NA	NA	20.6	7.00	05/27/2022	13:11
3	CAL3	1	Water	NA	NA	20.7	10.03	05/27/2022	13:13
4	ICV	1	Water	NA	NA	20.8	6.98	05/27/2022	13:15
5	CCV1	1	Water	NA	NA	20.7	2.01	05/27/2022	13:20
6	N3068-01	1	Water	NA	NA	20.1	11.92	05/27/2022	13:22
7	N3068-01DUP	1	Water	NA	NA	20.2	11.93	05/27/2022	13:23
8	N3098-01	1	Water	NA	NA	21.0	4.59	05/27/2022	13:25
9	CCV2	1	Water	NA	NA	20.7	12.03	05/27/2022	13:28

B120388

WORKLIST(Hardcopy Internal Chain)

WorkList Name: PH N3068 WorkList ID: 160056 Department: Wet-Chemistry Date: 05-27-2022 12:58:51

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/26/2022	Water	N3068-01	pH	Cool 4 deg C	TULL02	N11	WATER-DRUM	05/26/2022	9040C
05/25/2022	Water	N3098-01	pH	Cool 4 deg C	STER08	N11	POOL-WATER	05/25/2022	9040C

Date/Time 05/27/22 12:59
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 05/27/22 14:00
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]